大型混合酸化物燃料加工施設保障措置機器 性能確認試験報告書 (令和6年度)

2025年3月

日本原子力研究開発機構 核燃料サイクル工学研究所 MOX 燃料技術開発部

本資料は、日本原子力研究開発機構が実施した受託業務の成果を取りまとめたものです。 したがって、その利用は限られた範囲としており、その取扱には十分な注意を払ってくだ さい。この資料の全部又は一部を複写・複製・転載あるいは引用する場合、特別な許可を 必要としますので、下記にお問い合わせください。

₹319-1194

茨城県那珂郡東海村村松 4 番地 33 国立研究開発法人 日本原子力研究開発機構 核燃料サイクル工学研究所 MOX 燃料技術開発部 プロジェクト管理課

大型混合酸化物燃料加工施設保障措置機器性能確認試験報告書 (令和 6 年度)

要旨

本報告は、原子力規制庁殿と国立研究開発法人日本原子力研究開発機構との間で令和6年4月に締結した受託業務契約「令和6年度軽水炉等改良技術確証試験等委託費(大型混合酸化物燃料加工施設保障措置機器性能確認試験)事業」に基づき実施した保障措置機器に対する性能確認試験の結果をとりまとめたものである。

本業務において性能確認試験の対象とした保障措置機器は、燃料集合体測定機器 (AFAS)、バルク工程内 MOX 粉末・ペレット測定機器 (AVIS) 及び MOX 原料粉末貯蔵容器測定機器 (IPCA2) である。

AFAS の性能確認試験のうち、「長期管理限界の妥当性確認」では、令和 6 年 4 月から令和 7 年 3 月までの期間において、連続的なバックグラウンド測定及び定期的な ²⁵²Cf 中性子線源測定を実施し、AFAS が正常な状態を維持し、長期間安定して動作していたことを確認した。

「パルス間隔分布の評価」では、リストモジュールを使用して ²⁵²Cf 中性子線源を測定し、得られたパルス間隔分布が正常であり、各アンプ、また、それぞれが接続されている ³He 比例計数管及びケーブルが健全に保たれていることを確認した。

「プルトニウム 240 実効質量解析手法 Known M 法の適用性評価」では、モンテカルロシミュレーション(MCNPX)による評価を実施し、Known α 法及び Known M 法を使用して解析した燃料集合体中のプルトニウム 240 実効質量を比較することで、Known M 法の AFAS 測定への適用性を確認した。

AVIS の性能確認試験として実施した、「パルス間隔分布の評価」では、AFAS と同様に、得られたパルス間隔分布が正常であり、各アンプ、また、それぞれが接続されている ³He 比例計数管及びケーブルが健全に保たれていることを確認した。

IPCA2 の性能確認試験のうち、「長期管理限界の妥当性確認」では、令和6年4月から令和7年3月までの期間において、連続的なバックグラウンド測定、定期的な線源測定及び分銅測定を実施し、各検出器及びロードセルが正常に長期間安定して動作していたことを確認した。

「ハードウェアの修理及び動作確認」では、IPCA2 に接続されている Power Supply Box の交換を行い、交換した機器が正常に動作することを確認した。

「測定パラメータの評価」では、測定パラメータの評価を行い、IPCA2の動作設定に用いられる標準的な測定パラメータと比較することで、IPCA2の性能が一定に保たれていることを確認した。

目次

1. 試験概要	1
1.1 試験対象機器	1
1.2 試験項目	3
1.3 試験スケジュール	4
1.4 試験実施場所	4
1.5 測定パラメータ	5
1.6 試験に使用した ²⁵² Cf 中性子線源	7
2. AFAS の性能確認試験	8
2.1 長期管理限界の妥当性確認	8
2.1.1 目的	8
2.1.2 方法	8
2.1.3 結果	11
2.1.4 考察	62
2.2 パルス間隔分布の評価	64
2.2.1 目的	64
2.2.2 方法	64
2.2.3 結果	67
2.2.4 考察	82
2.3 プルトニウム 240 実効質量解析手法 Known M 法の適用性評価	83
2.3.1 目的	83
2.3.2 方法	83
2.3.3 結果	85
2.3.4 考察	86
2.3.5 結果(再解析後)	88
2.3.6 考察(再解析後)	89
3. AVIS の性能確認試	89
3.1 パルス間隔分布の評価	90
3.1.1 目的	90
3.1.2 方法	90
3.1.3 結果	92
3.1.4 考察	101
4. IPCA2 の性能確認試験	102
4.1 長期管理限界の妥当性確認	102
4.1.1 目的	102
4.1.2 方法	102

4.1.3 結果	103
4.1.4 考察	103
4.2 ハードウェアの修理及び動作確認	104
4.2.1 目的	104
4.2.2 方法	104
4.2.3 結果	104
4.2.4 考察	104
4.3 測定パラメータの評価	105
4.3.1 目的	105
4.3.2 方法	105
4.3.3 結果	105
4.3.4 考察	105
5. 結論	106
【参考文献】	108

図目次

図 2.1.2-1	AFAS 機器接続図	9
図 2.1.3-1	バックグラウンドのシングル値(2024/4/5~2024/8/5)(AFAS-B Collar 検出器)	13
図 2.1.3-2	バックグラウンドのシングル値(2024/8/24~20 2 0/9)(AFAS-B Collar 検出器)	13
図 2.1.3-3	バックグラウンドのシングル値(2024/10/11~2024/10/21)(AFAS-B Collar 検出器)	13
図 2.1.3-4	バックグラウンドのシングル値(2024/10/24~2024/12/17)(AFAS-B Collar 検出器)	14
図 2.1.3-5	バックグラウンドのシングル値(2024/12/17~2025/3/11)(AFAS-B Collar 検出器)	14
図 2.1.3-6	バックグラウンドのダブル値(2024/4/5~2024/8/5)(AFAS-B Collar 検出器)	14
図 2.1.3-7	バックグラウンドのダブル値(2024/8/24~2024/10/9)(AFAS-B Collar 検出器)	15
図 2.1.3-8	バックグラウンドのダブル値(2024/10/11~2024/10/21)(AFAS-B Collar 検出器)	15
図 2.1.3-9	バックグラウンドのダブル値(2024/10/24~2024/12/17)(AFAS-B Collar 検出器)	15
図 2.1.3-10	バックグラウンドのダブル値(2024/12/17~2025/3/11)(AFAS-B Collar 検出器)	16
図 2.1.3-11	シングル値と温度の相関(2024/4/5~2024/8/5)(AFAS-B Collar 検出器)	16
図 2.1.3-12	シングル値と温度の相関(2024/8/24~2024/10/9)(AFAS-B Collar 検出器)	16
図 2.1.3-13	シングル値と温度の相関(2024/10/11~2024/10/21)(AFAS-B Collar 検出器)	17
図 2.1.3-14	シングル値と温度の相関(2024/10/24~2024/12/17)(AFAS-B Collar 検出器)	17
図 2.1.3-15	シングル値と温度の相関(2024/12/17~2025/3/11)(AFAS-B Collar 検出器)	17
図 2.1.3-16	ダブル値と温度の相関(2024/4/5~2024/8/5)(AFAS-B Collar 検出器)	18
図 2.1.3-17	ダブル値と温度の相関(2024/8/24~2024/10/9)(AFAS-B Collar 検出器)	18
図 2.1.3-18	ダブル値と温度の相関(2024/10/11~2024/10/21)(AFAS-B Collar 検出器)	18
図 2.1.3-19	ダブル値と温度の相関(2024/10/24~2024/12/17)(AFAS-B Collar 検出器)	19
図 2.1.3-20	ダブル値と温度の相関(2024/12/17~2025/3/11)(AFAS-B Collar 検出器)	19
図 2.1.3-21	シングル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-B Collar 検出器)	19
図 2.1.3-22	シングル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-B Collar 検出器)	20
図 2.1.3-23	シングル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-B Collar 検出器)	20
図 2.1.3-24	シングル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-B Collar 検出器)	20
図 2.1.3-25	シングル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-B Collar 検出器)	21
図 2.1.3-26	ダブル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-B Collar 検出器)	21
図 2.1.3-27	ダブル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-B Collar 検出器)	21
図 2.1.3-28	ダブル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-B Collar 検出器)	22
図 2.1.3-29	ダブル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-B Collar 検出器)	22
図 2.1.3-30	ダブル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-B Collar 検出器)	22
図 2.1.3-31	バックグラウンドのシングル値(2024/4/5~2024/8/5)(AFAS-P Collar 検出器)	23
図 2.1.3-32	バックグラウンドのシングル値(2024/8/24~2024/10/9)(AFAS-P Collar 検出器)	23
図 2.1.3-33	バックグラウンドのシングル値(2024/10/11~2024/10/21)(AFAS-P Collar 検出器)	24
図 213-34	バックグラウンドのシングル値(2024/10/24~2024/12/17)(AFAS-P Collar 検出器)	24

図 2.1.3-35	バックグラウンドのシングル値(2024/12/17~2025/3/11)(AFAS-P Collar 検出器)	24
図 2.1.3-36	バックグラウンドのダブル値(2024/4/5~2024/8/5)(AFAS-P Collar 検出器)	25
図 2.1.3-37	バックグラウンドのダブル値(2024/8/24~2024/10/9)(AFAS-P Collar 検出器)	25
図 2.1.3-38	バックグラウンドのダブル値(2024/10/11~2024/10/21)(AFAS-P Collar 検出器)	25
図 2.1.3-39	バックグラウンドのダブル値(2024/10/24~2024/12/17)(AFAS-P Collar 検出器)	26
図 2.1.3-40	バックグラウンドのダブル値(2024/12/17~2025/3/11)(AFAS-P Collar 検出器)	26
図 2.1.3-41	シングル値と温度の相関(2024/4/5~2024/8/5)(AFAS-P Collar 検出器)	26
図 2.1.3-42	シングル値と温度の相関(2024/8/24~2024/10/9)(AFAS-P Collar 検出器)	27
図 2.1.3-43	シングル値と温度の相関(2024/10/11~2024/10/21)(AFAS-P Collar 検出器)	27
図 2.1.3-44	シングル値と温度の相関(2024/10/24~2024/12/17)(AFAS-P Collar 検出器)	27
図 2.1.3-45	シングル値と温度の相関(2024/12/17~2025/3/11)(AFAS-P Collar 検出器)	28
図 2.1.3-46	ダブル値と温度の相関(2024/4/5~2024/8/5)(AFAS-P Collar 検出器)	28
図 2.1.3-47	ダブル値と温度の相関(2024/8/24~2024/10/9)(AFAS-P Collar 検出器)	28
図 2.1.3-48	ダブル値と温度の相関(2024/10/11~2024/10/21)(AFAS-P Collar 検出器)	29
図 2.1.3-49	ダブル値と温度の相関(2024/10/24~2024/12/17)(AFAS-P Collar 検出器)	29
図 2.1.3-50	ダブル値と温度の相関(2024/12/17~2025/3/11)(AFAS-P Collar 検出器)	29
図 2.1.3-51	シングル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-P Collar 検出器)	30
図 2.1.3-52	シングル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-P Collar 検出器)	30
図 2.1.3-53	シングル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-P Collar 検出器)	30
図 2.1.3-54	シングル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-P Collar 検出器)	31
図 2.1.3-55	シングル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-P Collar 検出器)	31
図 2.1.3-56	ダブル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-P Collar 検出器)	31
図 2.1.3-57	ダブル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-P Collar 検出器)	32
図 2.1.3-58	ダブル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-P Collar 検出器)	32
図 2.1.3-59	ダブル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-P Collar 検出器)	32
図 2.1.3-60	ダブル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-P Collar 検出器)	33
図 2.1.3-61	バックグラウンドのシングル値(2024/4/5~2024/8/5)(AFAS-B Top Fork 検出器)	33
図 2.1.3-62	バックグラウンドのシングル値(2024/8/24~2024/10/9)(AFAS-B Top Fork 検出器)	34
図 2.1.3-63	バックグラウンドのシングル値(2024/10/11~2024/10/21)(AFAS-B Top Fork 検出器)	34
図 2.1.3-64	バックグラウンドのシングル値(2024/10/24~2024/12/17)(AFAS-B Top Fork 検出器)	34
図 2.1.3-65	バックグラウンドのシングル値(2024/12/17~2025/3/11)(AFAS-B Top Fork 検出器)	35
図 2.1.3-66	シングル値と温度の相関(2024/4/5~2024/8/5)(AFAS-B Top Fork 検出器)	35
図 2.1.3-67	シングル値と温度の相関(2024/8/24~2024/10/9)(AFAS-B Top Fork 検出器)	35
図 2.1.3-68	シングル値と温度の相関(2024/10/11~2024/10/21)(AFAS-B Top Fork 検出器)	36
図 2.1.3-69	シングル値と温度の相関(2024/10/24~2024/12/17)(AFAS-B Top Fork 検出器)	36
図 2.1.3-70	シングル値と温度の相関(2024/12/17~2025/3/11)(AFAS-B Top Fork 検出器)	36
図 2.1.3-71	シングル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-B Top Fork 検出器)	37

図 2.1.3-72	シングル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-B Top Fork 検出器)	37
図 2.1.3-73	シングル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-B Top Fork 検出器)	37
図 2.1.3-74	シングル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-B Top Fork 検出器)	38
図 2.1.3-75	シングル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-B Top Fork 検出器)	38
図 2.1.3-76	バックグラウンドのシングル値(2024/4/5~2024/8/5)(AFAS-PTop Fork 検出器)	39
図 2.1.3-77	バックグラウンドのシングル値(2024/8/24~2024/10/9)(AFAS-PTop Fork 検出器)	39
図 2.1.3-78	バックグラウンドのシングル値(2024/10/11~2024/10/21)(AFAS-P Top Fork 検出器)	39
図 2.1.3-79	バックグラウンドのシングル値(2024/10/24~2024/12/17)(AFAS-P Top Fork 検出器)	40
図 2.1.3-80	バックグラウンドのシングル値(2024/12/17~2025/3/11)(AFAS-P Top Fork 検出器)	40
図 2.1.3-81	シングル値と温度の相関(2024/4/5~2024/8/5)(AFAS-P Top Fork 検出器)	40
図 2.1.3-82	シングル値と温度の相関(2024/8/24~2024/10/9)(AFAS-PTop Fork 検出器)	41
図 2.1.3-83	シングル値と温度の相関(2024/10/11~2024/10/21)(AFAS-PTop Fork 検出器)	41
図 2.1.3-84	シングル値と温度の相関(2024/10/24~2024/12/17)(AFAS-P Top Fork 検出器)	41
図 2.1.3-85	シングル値と温度の相関(2024/12/17~2025/3/11)(AFAS-P Top Fork 検出器)	42
図 2.1.3-86	シングル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-P Top Fork 検出器)	42
図 2.1.3-87	シングル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-P Top Fork 検出器)	42
図 2.1.3-88	シングル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-PTop Fork 検出器)	43
図 2.1.3-89	シングル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-PTop Fork 検出器)	43
図 2.1.3-90	シングル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-P Top Fork 検出器)	43
図 2.1.3-91	バックグラウンドのシングル値(2024/4/5~2024/8/5)(AFAS-B Bottom Fork 検出器)	44
図 2.1.3-92	バックグラウンドのシングル値(2024/8/24~2024/10/9)(AFAS-B Bottom Fork 検出器)	44
図 2.1.3-93	バックグラウンドのシングル値(2024/10/11~2024/10/21)(AFAS-B Bottom Fork 検出器)	45
図 2.1.3-94	バックグラウンドのシングル値(2024/10/24~2024/12/17)(AFAS-B Bottom Fork 検出器)	45
図 2.1.3-95	バックグラウンドのシングル値(2024/12/17~2025/3/11)(AFAS-B Bottom Fork 検出器)	45
図 2.1.3-96	シングル値と温度の相関(2024/4/5~2024/8/5)(AFAS-B Bottom Fork 検出器)	46
図 2.1.3-97	シングル値と温度の相関(2024/8/24~2024/10/9)(AFAS-B Bottom Fork 検出器)	46
図 2.1.3-98	シングル値と温度の相関(2024/10/11~2024/10/21)(AFAS-B Bottom Fork 検出器)	46
図 2.1.3-99	シングル値と温度の相関(2024/10/24~2024/12/17)(AFAS-B Bottom Fork 検出器)	47
図 2.1.3-100	シングル値と温度の相関(2024/12/17~2025/3/11)(AFAS-B Bottom Fork 検出器)	47
図 2.1.3-101	シングル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-B Bottom Fork 検出器)	47
図 2.1.3-102	シングル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-B Bottom Fork 検出器)	48
図 2.1.3-103	シングル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-B Bottom Fork 検出器)	48
図 2.1.3-104	シングル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-B Bottom Fork 検出器)	48
図 2.1.3-105	シングル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-B Bottom Fork 検出器)	49
図 2.1.3-106	バックグラウンドのシングル値(2024/4/5~2024/8/5)(AFAS-P Bottom Fork 検出器)	49
図 2.1.3-107	バックグラウンドのシングル値(2024/8/24~2024/10/9)(AFAS-P Bottom Fork 検出器)	50
図 2.1.3-108	バックグラウンドのシングル値(2024/10/11~2024/10/21)(AFAS-P Bottom Fork 検出器)	50

図 2.1.3-109) バックグラウンドのシングル値(2024/10/24~2024/12/17)(AFAS-P Bottom Fork 検出器)	50
図 2.1.3-110	バックグラウンドのシングル値(2024/12/17~2025/3/11)(AFAS-P Bottom Fork 検出器)	51
図 2.1.3-111	シングル値と温度の相関(2024/4/5~2024/8/5)(AFAS-P Bottom Fork 検出器)	.51
図 2.1.3-112	シングル値と温度の相関(2024/8/24~2024/10/9)(AFAS-P Bottom Fork 検出器)	51
図 2.1.3-113	シングル値と温度の相関(2024/10/11~2024/10/21)(AFAS-P Bottom Fork 検出器)	52
図 2.1.3-114	シングル値と温度の相関(2024/10/24~2024/12/17)(AFAS-P Bottom Fork 検出器)	52
図 2.1.3-115	シングル値と温度の相関(2024/12/17~2025/3/11)(AFAS-P Bottom Fork 検出器)	52
図 2.1.3-116	シングル値と湿度の相関(2024/4/5~2024/8/5)(AFAS-P Bottom Fork 検出器)	53
図 2.1.3-117	シングル値と湿度の相関(2024/8/24~2024/10/9)(AFAS-P Bottom Fork 検出器)	53
図 2.1.3-118	シングル値と湿度の相関(2024/10/11~2024/10/21)(AFAS-P Bottom Fork 検出器)	53
図 2.1.3-119	シングル値と湿度の相関(2024/10/24~2024/12/17)(AFAS-P Bottom Fork 検出器)	54
図 2.1.3-120	シングル値と湿度の相関(2024/12/17~2025/3/11)(AFAS-P Bottom Fork 検出器)	.54
図 2.1.3-121	月ごとの%Difference の評価結果(2024年4月~2025年3月)(AFAS-B Top Fork 検出器)	55
図 2.1.3-122	2 月ごとの%Difference の評価結果(2024年4月~2025年3月)(AFAS-B Collar 検出器)	56
図 2.1.3-123	月ごとの%Difference の評価結果(2024年4月~2025年3月)(AFAS-B Bottom Fork 検出器)	56
図 2.1.3-124	月ごとの%Difference の評価結果(2024年4月~2025年3月)(AFAS-PTop Fork 検出器)	56
図 2.1.3-125	5 月ごとの%Difference の評価結果(2024年4月~2025年3月)(AFAS-P Collar 検出器)	.57
図 2.1.3-126	5 月ごとの%Difference の評価結果(2024年4月~2025年3月)(AFAS-P Bottom Fork 検出器)	.57
図 2.1.3-127	7 %Difference と温度の相関(2024年4月~2025年3月)(AFAS-B Top Fork 検出器)	57
図 2.1.3-128	8 %Difference と温度の相関(2024年4月~2025年3月)(AFAS-B Collar 検出器)	58
図 2.1.3-129	%Difference と温度の相関(2024 年 4 月~2025 年 3 月)(AFAS-B Bottom Fork 検出器)	58
図 2.1.3-130	%Difference と温度の相関(2024年4月~2025年3月)(AFAS-PTop Fork 検出器)	58
図 2.1.3-131	%Difference と温度の相関(2024年4月~2025年3月)(AFAS-P Collar 検出器)	. 59
図 2.1.3-132	2 %Difference と温度の相関(2024年4月~2025年3月)(AFAS-P Bottom Fork 検出器)	. 59
図 2.1.3-133	%Difference と湿度の相関(2024年4月~2025年3月)(AFAS-B Top Fork 検出器)	. 59
図 2.1.3-134	** %Difference と湿度の相関(2024年4月~2025年3月)(AFAS-B Collar 検出器)	60
図 2.1.3-135	。 %Difference と湿度の相関(2024 年 4 月~2025 年 3 月)(AFAS-B Bottom Fork 検出器)	60
図 2.1.3-136	5 %Difference と湿度の相関(2024年4月~2025年3月)(AFAS-PTop Fork 検出器)	60
図 2.1.3-137	~ %Difference と湿度の相関(2024 年 4 月~2025 年 3 月)(AFAS-P Collar 検出器)	.61
図 2.1.3-138	8 %Difference と湿度の相関(2024年4月~2025年3月)(AFAS-P Bottom Fork 検出器)	.61
図 2.1.4-1	シングル値と温度及び湿度の相関(2024/4/5~2024/8/5)(AFAS-B Bottom Fork 検出器)	63
図 2.2.2-1	アンプ A のパルス間隔分布の例(AFAS-B Collar 検出器)	64
図 2.2.2-2	AFAS-B Collar 検出器のアンプ配置図	65
図 2.2.2-3	AFAS-B Top Fork 及び Bottom Fork 検出器のアンプ配置図	65
図 2.2.2-4	AFAS-P Collar 検出器のアンプ配置図	65
図 2.2.2-5	AFAS-P Top Fork 及び Bottom Fork 検出器のアンプ配置図	65
図 2.2.3-1	AFAS-B Collar 検出器 アンプ A のパルス間隔分布の比較 (0~1000μs)	67

义	2.2.3-2	AFAS-B Collar 検出器 アンプ A のパルス間隔分布の比較 (0~100μs)	.67
図	2.2.3-3	AFAS-B Collar 検出器 アンプ B のパルス間隔分布の比較 (0~1000μs)	.68
図	2.2.3-4	AFAS-B Collar 検出器 アンプ B のパルス間隔分布の比較 (0~100μs)	.68
図	2.2.3-5	AFAS-B Collar 検出器 アンプ C のパルス間隔分布の比較 (0~1000μs)	.68
図	2.2.3-6	AFAS-B Collar 検出器 アンプ C のパルス間隔分布の比較 (0~100μs)	.69
図	2.2.3-7	AFAS-B Collar 検出器 アンプ D のパルス間隔分布の比較 (0~1000μs)	.69
図	2.2.3-8	AFAS-B Collar 検出器 アンプ D のパルス間隔分布の比較 (0~100μs)	.69
図	2.2.3-9	AFAS-B Collar 検出器 アンプ E のパルス間隔分布の比較 (0~1000μs)	.70
図	2.2.3-10	AFAS-B Collar 検出器 アンプ E のパルス間隔分布の比較 (0~100μs)	.70
図	2.2.3-11	AFAS-B Collar 検出器 アンプ F のパルス間隔分布の比較 (0~1000μs)	.70
図	2.2.3-12	AFAS-B Collar 検出器 アンプ F のパルス間隔分布の比較 (0~100μs)	.71
図	2.2.3-13	AFAS-B Top Fork 検出器 パルス間隔分布の比較 (0~1000μs)	.71
図	2.2.3-14	AFAS-B Top Fork 検出器 パルス間隔分布の比較(0~100µs)	.72
义	2.2.3-15	AFAS-B Bottom Fork 検出器のパルス間隔分布の比較(0~1000µs)	.72
义	2.2.3-16	AFAS-B Bottom Fork 検出器のパルス間隔分布の比較(0~100μs)	.73
义	2.2.3-17	AFAS-P Collar 検出器 アンプ A のパルス間隔分布の比較 (0~1000μs)	.73
図	2.2.3-18	AFAS-P Collar 検出器 アンプ A のパルス間隔分布の比較 (0~100μs)	.74
図	2.2.3-19	AFAS-P Collar 検出器 アンプ B のパルス間隔分布の比較 (0~1000μs)	.74
図	2.2.3-20	AFAS-P Collar 検出器 アンプ B のパルス間隔分布の比較 (0~100μs)	.74
図	2.2.3-21	AFAS-P Collar 検出器 アンプ C のパルス間隔分布の比較 (0~1000μs)	.75
図	2.2.3-22	AFAS-P Collar 検出器 アンプ C のパルス間隔分布の比較 (0~100μs)	.75
図	2.2.3-23	AFAS-P Collar 検出器 アンプ D のパルス間隔分布の比較 (0~1000μs)	.75
図	2.2.3-24	AFAS-P Collar 検出器 アンプ D のパルス間隔分布の比較 (0~100μs)	.76
図	2.2.3-25	AFAS-P Collar 検出器 アンプ E のパルス間隔分布の比較 (0~1000μs)	.76
図	2.2.3-26	AFAS-P Collar 検出器 アンプ E のパルス間隔分布の比較 (0~100μs)	.76
図	2.2.3-27	AFAS-P Collar 検出器 アンプ F のパルス間隔分布の比較 (0~1000μs)	.77
図	2.2.3-28	AFAS-P Collar 検出器 アンプ F のパルス間隔分布の比較 (0~100μs)	.77
図	2.2.3-29	AFAS-P Collar 検出器 アンプ G のパルス間隔分布の比較 (0~1000μs)	.77
図	2.2.3-30	AFAS-P Collar 検出器 アンプ G のパルス間隔分布の比較 (0~100μs)	.78
図	2.2.3-31	AFAS-P Collar 検出器 アンプ H のパルス間隔分布の比較 (0~1000μs)	.78
図	2.2.3-32	AFAS-P Collar 検出器 アンプ H のパルス間隔分布の比較 (0~100μs)	.78
図	2.2.3-33	AFAS-P Collar 検出器 アンプ I のパルス間隔分布の比較(0~1000μs)	.79
図	2.2.3-34	AFAS-P Collar 検出器 アンプ I のパルス間隔分布の比較 (0~100μs)	.79
図	2.2.3-35	AFAS-P Collar 検出器 アンプ J のパルス間隔分布の比較 (0~1000μs)	.79
図	2.2.3-36	AFAS-P Collar 検出器 アンプ J のパルス間隔分布の比較(0~100µs)	.80
図	2.2.3-37	AFAS-P Top Fork 検出器のパルス間隔分布の比較(0~1000μs)	.80
図	2.2.3-38	AFAS-P Top Fork 検出器のパルス間隔分布の比較(0~100μs)	.81

図	2.2.3-39	AFAS-P Bottom Fork 検出器のパルス間隔分布の比較(0~1000μs)	81
図	2.2.3-40	AFAS-P Bottom Fork 検出器のパルス間隔分布の比較(0~100µs)	82
図	2.3.3-1	BWR 燃料集合体の ²⁴⁰ Pu 実効質量の%Difference の比較	85
図	2.3.3-2	PWR 燃料集合体の ²⁴⁰ Pu 実効質量の%Difference の比較	86
図	2.3.5-1	BWR 燃料集合体の ²⁴⁰ Pu 実効質量の%Difference の比較	88
図	2.3.5-2	PWR 燃料集合体の ²⁴⁰ Pu 実効質量の%Difference の比較	88
図	3.1.2-1	AVIS の中性子検出器 アンプ A のパルス間隔分布の例	90
図	3.1.2-2	AVIS のアンプ配置図	91
図	3.1.3-1	アンプAのパルス間隔分布の比較 (0~1000µs)	92
図	3.1.3-2	アンプ A のパルス間隔分布の比較 (0~100µs)	92
図	3.1.3-3	アンプBのパルス間隔分布の比較 (0~1000µs)	93
図	3.1.3-4	アンプBのパルス間隔分布の比較 (0~100µs)	93
図	3.1.3-5	アンプ C のパルス間隔分布の比較 (0~1000µs)	93
図	3.1.3-6	アンプ C のパルス間隔分布の比較 (0~100µs)	94
図	3.1.3-7	アンプ D のパルス間隔分布の比較 (0~1000µs)	94
図	3.1.3-8	アンプ D のパルス間隔分布の比較 (0~100µs)	94
図	3.1.3-9	アンプEのパルス間隔分布の比較 (0~1000µs)	95
図	3.1.3-10	アンプEのパルス間隔分布の比較 (0~100µs)	95
図	3.1.3-11	アンプ F のパルス間隔分布の比較 (0~1000µs)	95
図	3.1.3-12	アンプ F のパルス間隔分布の比較(0~100µs)	96
図	3.1.3-13	アンプ G のパルス間隔分布の比較(0~1000µs)	96
図	3.1.3-14	アンプ G のパルス間隔分布の比較(0~100µs)	96
図	3.1.3-15	アンプ H のパルス間隔分布の比較(0~1000μs)	97
図	3.1.3-16	アンプ H のパルス間隔分布の比較($0\sim100\mu s$)	97
図	3.1.3-17	アンプ I のパルス間隔分布の比較(0~1000µs)	97
図	3.1.3-18	アンプ I のパルス間隔分布の比較(0~100µs)	98
図	3.1.3-19	アンプ J のパルス間隔分布の比較(0~1000µs)	98
図	3.1.3-20	アンプJのパルス間隔分布の比較 (0~100µs)	98
図	3.1.3-21	アンプ K のパルス間隔分布の比較(0~1000µs)	99
図	3.1.3-22	アンプ K のパルス間隔分布の比較(0~100µs)	99
図	3.1.3-23	アンプLのパルス間隔分布の比較 (0~1000μs)	99
図	3.1.3-24	アンプLのパルス間隔分布の比較 (0~100µs)	100
図	3.1.3-25	アンプ M のパルス間隔分布の比較(0~1000µs)	100
図	3.1.3-26	アンプ M のパルス間隔分布の比較(0~100µs)	100
図	3.1.3-27	アンプ N のパルス間隔分布の比較($0\sim1000\mu s$)	101
図	3 1 3-28	アンプNのパルス間隔分布の比較 (0~100us)	101

表目次

表	1.2-1	AFAS 性能確認試験項目	3
表	1.2-2	AVIS 性能確認試験項目	3
表	1.2-3	IPCA2 性能確認試験項目	3
表	1.3-1	性能確認試験スケジュール	4
表	1.5-1	AFAS-B の中性子検出器の測定パラメータ	5
表	1.5-2	AFAS-P の中性子検出器の測定パラメータ	5
表	1.5-3	AVIS の中性子検出器の測定パラメータ	6
表	1.6-1	試験に使用した ²⁵² Cf 中性子線源	7
表	2.1.3-1	Collar 検出器の各期間におけるバックグラウンドのシングル値の平均値	12
表	2.1.3-2	Fork 検出器の各期間におけるバックグラウンドのシングル値の平均値	12
表	2.1.4-1	各計数装置のクロック周波数	63
表	2.2.2-1	各検出器におけるアンプ数	64
表	2.3.2-1	BWR 燃料集合体の同位体組成比	84
表	2.3.2-2	PWR 燃料集合体の同位体組成比	84
表	2.3.2-3	Known M 法による解析に使用したキャリブレーションパラメータ	85
表	2.3.4-1	シミュレーションから求めた各検出効率	87
表	3.1.2-1	AVIS におけるアンプ数	90

写真目次

写真	1.1-1	AFAS の外観写真	1
写真	1.1-2	AVIS の外観写真	2
写真	1.1-3	IPCA2 の外観写真	2
写真	2.1.4-1	検出器内の湿度確認用インジケーター(指示値が 40%の場合)	63
写真	2.2.2-1	AFAS のアンプ(AMPTEK A111)	66

添付資料目次

【絲	源情	報】	
線源	証明	書	添付-1
-	-	生能確認試験】 長期管理限界の妥当性確認・・・・・・・・・・・・・・・・・・・・・・・・・・・・・・・・・・・・	添付-6
[IP	CA2	性能確認試験】	
(1)	4.1	長期管理限界の妥当性確認	添付-61
(2)	4.2	ハードウェアの修理及び動作確認	添付-61
(3)	4.3	測定パラメータの評価	添付-61

略語集

略語	正式名	日本語名
AFAS	Advanced Fuel assembly Assay System	燃料集合体測定機器
AMSR	Advanced Multiplicity Shift Register	シフトレジスター
AVIS	Advanced Verification for Inventory Samples system	バルク工程内 MOX 粉末・ペ
		レット測定機器
INCC	IAEA Neutron Coincidence Counting	測定ソフトウェア
IPCA2	Improved Plutonium Canister Assay system 2	MOX 原料粉末貯蔵容器測定
		機器
JAEA	Japan Atomic Energy Agency	国立研究開発法人日本原子力
		研究開発機構
J-MOX	JNFL's MOX fuel fabrication plant	大型混合酸化物燃料加工施設
JNFL	Japan Nuclear Fuel Limited	日本原燃株式会社
JSR-12	Model JSR-12 Neutron Coincidence Analyzer	シフトレジスター
JSR-15	JSR-15 Handheld Multiplicity Register	シフトレジスター
JTC	Joint Technical Committee	共同技術会合
LANL	Los Alamos National Laboratory	米国ロスアラモス国立研究所
MIC	Multi-Instrument Collect	測定ソフトウェア
UDL-1	Unattended Data Logger	シフトレジスター

1. 試験概要

国立研究開発法人日本原子力研究開発機構(以下、「JAEA」という)は、原子力規制 庁殿と令和6年4月に締結した受託業務契約「令和6年度軽水炉等改良技術確証試験等 委託費(大型混合酸化物燃料加工施設保障措置機器性能確認試験)事業」に基づき、大 型混合酸化物燃料加工施設(以下、「J-MOX」という)に設置される保障措置機器であ る燃料集合体測定機器(以下、「AFAS」という)、バルク工程内 MOX 粉末・ペレット測 定機器(以下、「AVIS」という)及び MOX 原料粉末貯蔵容器測定機器(以下、「IPCA2」 という)の性能確認試験を実施した。

1.1 試験対象機器

(1) AFAS

AFAS は、軽水炉用 MOX 燃料集合体用の非破壊測定装置であり、PWR 用の AFAS-P(以下、「AFAS-PWR」または「AFAS-P」という)及び BWR 用の AFAS-B(以下、「AFAS-BWR」または「AFAS-B」という) の2式から構成される。1式の AFAS は、単位長さあたりの Pu 量測定用の中性子検出器 1 台(Collar 検出器)と燃料集合体の有効長(充填されている MOX ペレットのスタック長)測定用の中性子検出器 2 台(Top Fork 検出器及び Bottom Fork 検出器)により構成される。装置の外観写真を写真 1.1-1 に示す。




写真 1.1-1 AFAS の外観写真

(2) AVIS

AVIS は、工程内から収去した粉末及びペレット用の非破壊測定装置であり、Pu 量測 定用の中性子検出器及び Pu 同位体組成比測定用の高純度ゲルマニウム (HPGe) 半導体 検出器 (ガンマ線検出器) により構成される。装置の外観写真を写真 1.1-2 に示す。

ガンマ線検出器



中性子検出器

写真 1.1-2 AVIS の外観写真

(3) IPCA2

IPCA2 は、MOX 原料粉末用の非破壊測定装置であり、Pu 量測定用の中性子検出器、Pu 同位体組成比測定用の電気冷却式高純度ゲルマニウム(HPGe)半導体検出器(ガンマ線検出器)3 台及びロードセル(荷重変換器)により構成される。装置の外観写真を写真 1.1-3 に示す。




写真 1.1-3 IPCA2 の外観写真

(公益財団法人 核物質管理センター "H25 年度 大型 MOX 燃料加工施設保障措置試験"報告書より引用)

1.2 試験項目

(1) AFAS の性能確認試験

AFAS の性能確認試験項目を表 1.2-1 に示す。各試験の詳細については 2.項に示す。

表 1.2-1 AFAS 性能確認試験項目

機器名	試験項目	参照項
	長期管理限界の妥当性確認	2.1
AFAS	パルス間隔分布の評価	2.2
	プルトニウム 240 実効質量解析手法 Known M 法の適用性評価	2.3

(2) AVIS の性能確認試験

AVIS の性能確認試験項目を表 1.2-2 に示す。各試験の詳細な内容については 3.項に示す。

表 1.2-2 AVIS 性能確認試験項目

機器名	試験項目	参照項
AVIS	パルス間隔分布の評価	3.1

(3) IPCA2 の性能確認試験

IPCA2 の性能確認試験項目を表 1.2-3 に示す。各試験の詳細な内容については 4.項に示す。

表 1.2-3 IPCA2 性能確認試験項目

機器名	試験項目	参照項
	長期管理限界の妥当性確認	4.1
IPCA2	ハードウェアの修理及び動作確認	4.2
	測定パラメータの評価	4.3

1.3 試験スケジュール

AFAS、AVIS 及び IPCA2 の性能確認試験スケジュールを表 1.3-1 に示す。

令和6年度 内容 5 9 3 4 6 8 10 11 12 2 性能確認試験 報告書作成 JTC^(*1)会合 $(JNFL^{(*2)})$ JTC 会合 (リモート会議) 試験に係る会合等 (開催場所) LANL(*3)との技術会合 (リモート会議)

表 1.3-1 性能確認試験スケジュール

- *1: JTC (共同技術会合 (Joint Technical Committee))
- *2: JNFL (日本原燃株式会社 (Japan Nuclear Fuel Limited))
- *3: LANL(米国ロスアラモス国立研究所(Los Alamos National Laboratory))

1.4 試験実施場所

- (1) AFAS 性能確認試験
 - ・プルトニウム燃料第二開発室 試験検査室 (C) (C-137)
- (2) AVIS 性能確認試験
 - ・プルトニウム燃料第一開発室 燃料要素組立室 (R-120)
- (3) IPCA2 性能確認試験
 - ・米国ロスアラモス国立研究所

1.5 測定パラメータ

(1) AFAS 用測定パラメータ

本年度の試験に使用した AFAS-B 及び AFAS-P の各検出器の測定パラメータを表 1.5-1 及び表 1.5-2 に示す。

表 1.5-1 AFAS-B の中性子検出器の測定パラメータ

項目	Collar	Top Fork	Bottom Fork	
シフトレジスターの種類	JSR-12, JSR-15,	JSR-12, JSR-15,	JSR-12, JSR-15,	
	UDL-1	UDL-1	UDL-1	
プレディレイ (μs)	1.50	1.50	1.50	
ゲート幅 (μs)	64.00	64.00	64.00	
印加電圧 (V)	1,720	1,720	1,720	
ダイアウェイタイム (μs)	50.0000	50.0000	50.0000	
中性子検出効率(252Cf)	0.1970	0.0080	0.0060	
Multiplicity	160.0000×10 ⁻⁹	0.0000	0.0000	
デッドタイム補正係数	160.0000×10	0.0000	0.0000	
デッドタイム補正係数 a	0.6419×10 ⁻⁶	0.0000	0.0000	
デッドタイム補正係数 b	0.1030×10 ⁻¹²	0.0000	0.0000	

表 1.5-2 AFAS-P の中性子検出器の測定パラメータ

項目	Collar	Top Fork	Bottom Fork	
シフトレジスターの種類	JSR-12, JSR-15,	JSR-12, JSR-15,	JSR-12, JSR-15,	
シノトレンハグーの種類	UDL-1	UDL-1	UDL-1	
プレディレイ (μs)	1.50	1.50	1.50	
ゲート幅 (μs)	64.00	64.00	64.00	
印加電圧 (V)	1,720	1,720	1,720	
ダイアウェイタイム (μs)	50.0000	50.0000	50.0000	
中性子検出効率(252Cf)	0.1620	0.0126	0.0127	
Multiplicity	86.5000×10 ⁻⁹	0.0000	0.0000	
デッドタイム補正係数	80.3000^10	0.0000	0.0000	
デッドタイム補正係数 a	0.3458×10 ⁻⁶	0.0000	0.0000	
デッドタイム補正係数 b	0.02989×10 ⁻¹²	0.0000	0.0000	

(2) AVIS 用測定パラメータ

本年度の試験に使用した AVIS の中性子検出器の測定パラメータを表 1.5-3 に示す。

表 1.5-3 AVIS の中性子検出器の測定パラメータ

項目	パラメータ
シフトレジスターの種類	JSR-15
プリディレイ(μs)	1.50
ゲート幅(μs)	64.00
印加電圧(V)	1,740
ダイアウェイタイム(μs)	30.00
中性子検出効率	0.6750
Multiplicity デッドタイム補正係数	72.6000×10 ⁻⁹
デッドタイム補正係数 a	0.2904×10 ⁻⁶
デッドタイム補正係数 b	0.0211×10 ⁻¹²
ダブルゲートフラクション	0.7930
トリプルゲートフラクション	0.6225
スロープ b	3.240449×10 ²
Alpha weight	1.000
ρο	4.707231×10 ⁻¹
K	2.166

(3) IPCA2 用測定パラメータ

本年度の試験に使用した IPCA2 の中性子検出器の測定パラメータについては、LANL報告書 "Technical Specification for the Improved Plutonium Canister Assay System (IPCA2)" (参考文献 1)を参照のこと。

1.6 試験に使用した ²⁵²Cf 中性子線源

本年度の試験に使用した ²⁵²Cf 中性子線源の線源番号及び中性子放出率を表 1.6-1 に示す。表中の中性子放出率は、各線源の線源証明書に記載されている値を、評価日に合わせて崩壊計算することにより求めた値である。なお、LANLにおいて使用された中性子線源については、LANL報告書"Annual IPCA2 Performance Report for JFY24(添付資料)を参照のこと。

表 1.6-1 試験に使用した ²⁵²Cf 中性子線源

線源番号	中性子放出率(n/s)	評価日	
H4-694	1.3×10 ⁴	2024/4/1	
T1-349	1.6×10 ⁵	2024/4/1	

- 2. AFAS の性能確認試験
- 2.1 長期管理限界の妥当性確認

2.1.1 目的

J-MOX において査察官非立会測定システムとして使用される AFAS の長期安定性を確認する。また、AFAS の各検出器から得られる計数率の長期管理限界(*4)の妥当性確認 (長期管理限界内に収まっているか否かの確認) を定期的に行う。

*4: 長期管理限界:長期間の使用を考慮した管理限界(統計的に有効な一定数の測定値をグラフ化した時の、測定値がかなり高い確率で存在する範囲の境界)

2.1.2 方法

連続的なバックグラウンド測定を実施し、AFAS の各検出器の長期安定性を評価するとともに、本測定期間における測定環境の温度及び湿度を測定し、計数率との相関の有無を確認する。

また、定期的な ²⁵²Cf 中性子線源測定を実施し、長期にわたり取得した各計数率及び 誤差より、各検出器の長期管理限界を評価するとともに、測定環境の温度及び湿度と計 数率及びその期待値の相対差(以下、「%Difference」という)の相関の有無を確認する。 なお、本測定では、1.6 項の表 1.6-1 に示した ²⁵²Cf 中性子線源(H4-694)を使用する。

本確認は、AFAS-B 及び AFAS-P の各検出器について行い、Multi-Instrument Collect (以下、「MIC」という)ソフトウェア、IAEA Neutron Coincidence Counting (以下、「INCC」という)ソフトウェア及び JAEA 所掌の Model JSR-12 Neutron Coincidence Analyzer (以下、「JSR-12」という)を使用して測定及び評価を行う。また、AFAS-P の各検出器の測定については、IAEA から提供された Unattended Data Logger (以下、「UDL-1」という)を使用した長期管理限界の妥当性確認を合わせて実施する。JSR-12 及び UDL-1 を使用した場合の機器接続図を図 2.1.2-1 に示す。




図 2.1.2-1 AFAS 機器接続図

詳細な試験手順を以下に示す。

- (1) バックグラウンド測定
- ① AFAS の各検出器及び温湿度計を試験用架台に設置する。
- ② MIC ソフトウェア及び UDL-1 制御用アプリケーションを使用して、連続的なバックグラウンド測定を実施し、INCC ソフトウェアを使用して各計数率を評価する。測定条件を以下に示す。また、測定期間における温湿度の測定も同時に実施する。
 - · 測定時間: 60 秒×∞回
 - ·測定期間: 2024年4月~2025年3月(*5)
 - 取得する計数率: Collar 検出器: シングル値(*6)、ダブル値(*7)

Fork 検出器:シングル値

- ③ ②にて取得した各計数率における異常値の有無を確認し、各検出器の長期安定性を評価する。また、測定環境の温湿度と計数率の相関の有無を評価する。
- *5: AFAS の他の試験期間及び維持管理時、また核燃料物質を近傍で取り扱う査察期間等及び電源が使用不可となる施設保守期間等を除く。
- *6: 検出器で検出された全ての中性子の単位時間あたりの数 (cps)
- *7: 検出器で検出された核分裂事象に起因する中性子の単位時間あたりの数 (cps)
- (2) ²⁵²Cf 中性子線源測定
- ① AFAS の各検出器及び温湿度計を試験用架台に設置する。
- ② 線源固定治具を使用して、²⁵²Cf 中性子線源を任意の検出器にセットする。
- ③ MIC ソフトウェア及び UDL-1 制御用アプリケーションを使用して、²⁵²Cf 中性子線源測定を月 1 回の頻度で実施し、INCC ソフトウェアを使用して各計数率を評価する。測定条件を以下に示す。また、測定時における温湿度の測定も同時に実施する。
 - · 測定時間:600 秒
 - ・取得する計数率: Collar 検出器: ダブル値

Fork 検出器:シングル値

- ④ 全ての検出器について、②~③を実施する。
- ⑤ ②~④により得られた各計数率及び誤差より、各検出器の長期管理限界を評価する。また、測定環境の温湿度と各計数率の%Difference の相関の有無を評価する。

2.1.3 結果

(1) バックグラウンド測定

連続的なバックグラウンド測定の結果及びバックグラウンド計数率と測定環境の温度及び湿度の相関の評価結果を検出器毎に次頁以降の①~⑥に示す。評価結果については、測定期間中の環境変化(AFAS 検出器近傍の設備の移動、隣接する部屋内の放射線源の増減)により、バックグラウンドの計数率が変動した期間ごとに分けて示した。バックグラウンド測定結果を示すグラフ中の赤色の実線は各計数率の平均値を示し、破線は平均値の $\pm 3\sigma$ (3σ は標準偏差を 3 倍した値)を示す。また、Collar 検出器及び Fork 検出器の各期間におけるバックグラウンドのシングル値の平均値を表 2.1.3-1 及び表 2.1.3-1 とここを動作させているコンピュータの調整作業を実施したため、作業終了までの 1 か月間については当該計数装置によるバックグラウンド測定は停止した。

AFAS-B の Collar 検出器におけるシングル値は図 $2.1.3-1\sim5$ に示すとおりであり、ほぼ全ての計数率は平均値の $\pm3\sigma$ の範囲内に収まった。ダブル値は図 $2.1.3-6\sim10$ に示すとおりであり、平均値は全期間において約 0.005cps であった。

AFAS-P の Collar 検出器におけるシングル値は図 $2.1.3-31\sim35$ に示すとおりであり、ほぼ全ての計数率は平均値の $\pm3\sigma$ の範囲内に収まった。ダブル値は図 $2.1.3-36\sim40$ に示すとおりであり、平均値は全期間において約 0.190cps であった(異常値を含む)。なお、シングル値及びダブル値において、全期間に単発的な高い計数率(シングル値:約 19cps ~37 cps、ダブル値:約 43cps ~3766 cps)がシングル値及びダブル値それぞれに 45 回程度確認された(図 $2.1.3-31\sim40$ に示す)。

Fork 検出器における各検出器のシングル値の平均値は表 2.1.3-2 に示すとおりであり、ほぼ全ての計数率は平均値の $\pm 3\sigma$ の範囲内に収まった。なお、AFAS-B の Bottom Fork 検出器において、2024 年 6 月から 8 月にかけてシングル値が徐々に増加する事象が確認された(図 2.1.3-91 に示す)。

測定期間中、測定環境の温度は約 16℃~28℃、湿度は約 38%~77%の範囲で変動した。図 2.1.3-11~30、図 2.1.3-41~60、図 2.1.3-66~75、図 2.1.3-81~90、図 2.1.3-97~100、図 2.1.3-102~105 及び図 2.1.3-111~120 に示す期間では、各検出器の計数率と温度及び湿度に相関関係はみられなかったが、図 2.1.3-96、図 2.1.3-101 及び図 2.1.3-102 に示す期間では、AFAS-Bの Bottom Fork 検出器のシングル値と温度及び湿度に相関係数 0.5~0.6 程度の正の相関がみられた。

表 2.1.3-1 Collar 検出器の各期間におけるバックグラウンドのシングル値の平均値

	バックグラウンドのシングル値の平均値(cps)				
 検出器	2024/4/5	8/24	10/11	10/24	12/17
(火口(4)	5	5	5	5	5
	8/5	10/9	10/21	12/17	2025/3/11
AFAS-B Collar	9.971	9.298	9.680	8.554	6.730
AFAS-P Collar	14.846	13.700	14.298	12.042	9.462

表 2.1.3-2 Fork 検出器の各期間におけるバックグラウンドのシングル値の平均値

27,127					
	バックグラウンドのシングル値の平均値(cps)				
 検出器	2024/4/5	8/24	10/11	10/24	12/17
7人口40	5	5	5	5	5
	8/5	10/9	10/21	12/17	2025/3/11
AFAS-B Top Fork	1.331	1.243	1.293	1.175	0.884
AFAS-P Top Fork	2.220	2.092	2.188	1.878	1.426
AFAS-B Bottom Fork	1.268	1.113	1.021	0.913	0.869
AFAS-P Bottom Fork	1.650	1.510	1.585	1.343	1.056

① AFAS-B Collar 検出器

・バックグラウンドのシングル値:図2.1.3-1~5

・バックグラウンドのダブル値:図2.1.3-6~10

・シングル値と温度の相関:図2.1.3-11~15

・ダブル値と温度の相関:図2.1.3-16~20

・シングル値と湿度の相関:図2.1.3-21~25

・ダブル値と湿度の相関:図2.1.3-26~30




図 2.1.3-1 バックグラウンドのシングル値(2024/4/5~2024/8/5) (AFAS-B Collar 検出器)




図 2.1.3-2 バックグラウンドのシングル値(2024/8/24~20 2 0/9) (AFAS-B Collar 検出器)




図 2.1.3-3 バックグラウンドのシングル値(2024/10/11~2024/10/21) (AFAS-B Collar 検出器)




図 2.1.3-4 バックグラウンドのシングル値(2024/10/24~2024/12/17) (AFAS-B Collar 検出器)



図 2.1.3-5 バックグラウンドのシングル値(2024/12/17~2025/3/11) (AFAS-B Collar 検出器)




図 2.1.3-6 バックグラウンドのダブル値(2024/4/5~2024/8/5) (AFAS-B Collar 検出器)




図 2.1.3-7 バックグラウンドのダブル値(2024/8/24~2024/10/9) (AFAS-B Collar 検出器)




図 2.1.3-8 バックグラウンドのダブル値(2024/10/11~2024/10/21) (AFAS-B Collar 検出器)




図 2.1.3-9 バックグラウンドのダブル値(2024/10/24~2024/12/17) (AFAS-B Collar 検出器)




図 2.1.3-10 バックグラウンドのダブル値(2024/12/17~2025/3/11) (AFAS-B Collar 検出器)




図 2.1.3-11 シングル値と温度の相関(2024/4/5~2024/8/5) (AFAS-B Collar 検出器)




図 2.1.3-12 シングル値と温度の相関(2024/8/24~2024/10/9) (AFAS-B Collar 検出器)




図 2.1.3-13 シングル値と温度の相関(2024/10/11~2024/10/21) (AFAS-B Collar 検出器)




図 2.1.3-14 シングル値と温度の相関(2024/10/24~2024/12/17) (AFAS-B Collar 検出器)




図 2.1.3-15 シングル値と温度の相関(2024/12/17~2025/3/11) (AFAS-B Collar 検出器)




図 2.1.3-16 ダブル値と温度の相関(2024/4/5~2024/8/5) (AFAS-B Collar 検出器)




図 2.1.3-17 ダブル値と温度の相関(2024/8/24~2024/10/9) (AFAS-B Collar 検出器)




図 2.1.3-18 ダブル値と温度の相関(2024/10/11~2024/10/21) (AFAS-B Collar 検出器)




図 2.1.3-19 ダブル値と温度の相関(2024/10/24~2024/12/17) (AFAS-B Collar 検出器)




図 2.1.3-20 ダブル値と温度の相関(2024/12/17~2025/3/11) (AFAS-B Collar 検出器)




図 2.1.3-21 シングル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-B Collar 検出器)




図 2.1.3-22 シングル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-B Collar 検出器)




図 2.1.3-23 シングル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-B Collar 検出器)




図 2.1.3-24 シングル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-B Collar 検出器)




図 2.1.3-25 シングル値と湿度の相関(2024/12/17~2025/3/11) (AFAS-B Collar 検出器)




図 2.1.3-26 ダブル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-B Collar 検出器)




図 2.1.3-27 ダブル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-B Collar 検出器)




図 2.1.3-28 ダブル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-B Collar 検出器)




図 2.1.3-29 ダブル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-B Collar 検出器)




図 2.1.3-30 ダブル値と湿度の相関(2024/12/17~2025/3/11) (AFAS-B Collar 検出器)

② AFAS-P Collar 検出器

・バックグラウンドのシングル値:図2.1.3-31~35

・バックグラウンドのダブル値:図2.1.3-36~40

・シングル値と温度の相関:図2.1.3-41~45

・ダブル値と温度の相関:図2.1.3-46~50

・シングル値と湿度の相関:図2.1.3-51~55

・ダブル値と湿度の相関:図2.1.3-56~60




図 2.1.3-31 バックグラウンドのシングル値(2024/4/5~2024/8/5) (AFAS-P Collar 検出器)




図 2.1.3-32 バックグラウンドのシングル値(2024/8/24~2024/10/9) (AFAS-P Collar 検出器)




図 2.1.3-33 バックグラウンドのシングル値(2024/10/11~2024/10/21) (AFAS-P Collar 検出器)




図 2.1.3-34 バックグラウンドのシングル値(2024/10/24~2024/12/17) (AFAS-P Collar 検出器)




図 2.1.3-35 バックグラウンドのシングル値(2024/12/17~2025/3/11) (AFAS-P Collar 検出器)




図 2.1.3-36 バックグラウンドのダブル値(2024/4/5~2024/8/5) (AFAS-P Collar 検出器)




図 2.1.3-37 バックグラウンドのダブル値(2024/8/24~2024/10/9) (AFAS-P Collar 検出器)




図 2.1.3-38 バックグラウンドのダブル値(2024/10/11~2024/10/21) (AFAS-P Collar 検出器)




図 2.1.3-39 バックグラウンドのダブル値(2024/10/24~2024/12/17) (AFAS-P Collar 検出器)




図 2.1.3-40 バックグラウンドのダブル値(2024/12/17~2025/3/11) (AFAS-P Collar 検出器)




図 2.1.3-41 シングル値と温度の相関(2024/4/5~2024/8/5) (AFAS-P Collar 検出器)




図 2.1.3-42 シングル値と温度の相関(2024/8/24~2024/10/9) (AFAS-P Collar 検出器)




図 2.1.3-43 シングル値と温度の相関(2024/10/11~2024/10/21) (AFAS-P Collar 検出器)




図 2.1.3-44 シングル値と温度の相関(2024/10/24~2024/12/17) (AFAS-P Collar 検出器)




図 2.1.3-45 シングル値と温度の相関(2024/12/17~2025/3/11) (AFAS-P Collar 検出器)




図 2.1.3-46 ダブル値と温度の相関(2024/4/5~2024/8/5) (AFAS-P Collar 検出器)




図 2.1.3-47 ダブル値と温度の相関(2024/8/24~2024/10/9) (AFAS-P Collar 検出器)




図 2.1.3-48 ダブル値と温度の相関(2024/10/11~2024/10/21) (AFAS-P Collar 検出器)




図 2.1.3-49 ダブル値と温度の相関(2024/10/24~2024/12/17) (AFAS-P Collar 検出器)




図 2.1.3-50 ダブル値と温度の相関(2024/12/17~2025/3/11) (AFAS-P Collar 検出器)




図 2.1.3-51 シングル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-P Collar 検出器)




図 2.1.3-52 シングル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-P Collar 検出器)




図 2.1.3-53 シングル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-P Collar 検出器)




図 2.1.3-54 シングル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-P Collar 検出器)




図 2.1.3-55 シングル値と湿度の相関(2024/12/17~2025/3/11) (AFAS-P Collar 検出器)




図 2.1.3-56 ダブル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-P Collar 検出器)




図 2.1.3-57 ダブル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-P Collar 検出器)




図 2.1.3-58 ダブル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-P Collar 検出器)




図 2.1.3-59 ダブル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-P Collar 検出器)




図 2.1.3-60 ダブル値と湿度の相関(2024/12/17~2025/3/11) (AFAS-P Collar 検出器)

③ AFAS-B Top Fork 検出器

・バックグラウンドのシングル値:図2.1.3-61~65

・シングル値と温度の相関:図2.1.3-66~70

・シングル値と湿度の相関:図2.1.3-71~75




図 2.1.3-61 バックグラウンドのシングル値(2024/4/5~2024/8/5) (AFAS-B Top Fork 検出器)




図 2.1.3-62 バックグラウンドのシングル値(2024/8/24~2024/10/9) (AFAS-B Top Fork 検出器)




図 2.1.3-63 バックグラウンドのシングル値(2024/10/11~2024/10/21) (AFAS-B Top Fork 検出器)




図 2.1.3-64 バックグラウンドのシングル値(2024/10/24~2024/12/17) (AFAS-B Top Fork 検出器)




図 2.1.3-65 バックグラウンドのシングル値(2024/12/17~2025/3/11) (AFAS-B Top Fork 検出器)




図 2.1.3-66 シングル値と温度の相関(2024/4/5~2024/8/5) (AFAS-B Top Fork 検出器)




図 2.1.3-67 シングル値と温度の相関(2024/8/24~2024/10/9) (AFAS-B Top Fork 検出器)




図 2.1.3-68 シングル値と温度の相関(2024/10/11~2024/10/21) (AFAS-B Top Fork 検出器)




図 2.1.3-69 シングル値と温度の相関(2024/10/24~2024/12/17) (AFAS-B Top Fork 検出器)




図 2.1.3-70 シングル値と温度の相関(2024/12/17~2025/3/11) (AFAS-B Top Fork 検出器)




図 2.1.3-71 シングル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-B Top Fork 検出器)




図 2.1.3-72 シングル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-B Top Fork 検出器)




図 2.1.3-73 シングル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-B Top Fork 検出器)




図 2.1.3-74 シングル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-B Top Fork 検出器)




図 2.1.3-75 シングル値と湿度の相関(2024/12/17~2025/3/11) (AFAS-B Top Fork 検出器)

④ AFAS-P Top Fork 検出器

・バックグラウンドのシングル値:図2.1.3-76~80

・シングル値と温度の相関:図 2.1.3-81~85 ・シングル値と湿度の相関:図 2.1.3-86~90




図 2.1.3-76 バックグラウンドのシングル値(2024/4/5~2024/8/5)

(AFAS-P Top Fork 検出器)




図 2.1.3-77 バックグラウンドのシングル値(2024/8/24~2024/10/9) (AFAS-P Top Fork 検出器)




図 2.1.3-78 バックグラウンドのシングル値(2024/10/11~2024/10/21) (AFAS-P Top Fork 検出器)




図 2.1.3-79 バックグラウンドのシングル値(2024/10/24~2024/12/17) (AFAS-P Top Fork 検出器)




図 2.1.3-80 バックグラウンドのシングル値(2024/12/17~2025/3/11) (AFAS-P Top Fork 検出器)




図 2.1.3-81 シングル値と温度の相関(2024/4/5~2024/8/5) (AFAS-P Top Fork 検出器)




図 2.1.3-82 シングル値と温度の相関(2024/8/24~2024/10/9) (AFAS-P Top Fork 検出器)




図 2.1.3-83 シングル値と温度の相関(2024/10/11~2024/10/21) (AFAS-P Top Fork 検出器)




図 2.1.3-84 シングル値と温度の相関(2024/10/24~2024/12/17) (AFAS-P Top Fork 検出器)




図 2.1.3-85 シングル値と温度の相関 (2024/12/17~2025/3/11) (AFAS-P Top Fork 検出器)




図 2.1.3-86 シングル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-P Top Fork 検出器)




図 2.1.3-87 シングル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-P Top Fork 検出器)




図 2.1.3-88 シングル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-P Top Fork 検出器)




図 2.1.3-89 シングル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-P Top Fork 検出器)




図 2.1.3-90 シングル値と湿度の相関 (2024/12/17~2025/3/11) (AFAS-P Top Fork 検出器)

⑤ AFAS-B Bottom Fork 検出器

・バックグラウンドのシングル値:図2.1.3-91~95

・シングル値と温度の相関:図2.1.3-96~100

・シングル値と湿度の相関:図2.1.3-101~105




図 2.1.3-91 バックグラウンドのシングル値(2024/4/5~2024/8/5) (AFAS-B Bottom Fork 検出器)




図 2.1.3-92 バックグラウンドのシングル値(2024/8/24~2024/10/9) (AFAS-B Bottom Fork 検出器)




図 2.1.3-93 バックグラウンドのシングル値(2024/10/11~2024/10/21) (AFAS-B Bottom Fork 検出器)




図 2.1.3-94 バックグラウンドのシングル値(2024/10/24~2024/12/17) (AFAS-B Bottom Fork 検出器)




図 2.1.3-95 バックグラウンドのシングル値(2024/12/17~2025/3/11) (AFAS-B Bottom Fork 検出器)




図 2.1.3-96 シングル値と温度の相関(2024/4/5~2024/8/5) (AFAS-B Bottom Fork 検出器)




図 2.1.3-97 シングル値と温度の相関(2024/8/24~2024/10/9) (AFAS-B Bottom Fork 検出器)




図 2.1.3-98 シングル値と温度の相関(2024/10/11~2024/10/21) (AFAS-B Bottom Fork 検出器)




図 2.1.3-99 シングル値と温度の相関(2024/10/24~2024/12/17) (AFAS-B Bottom Fork 検出器)




図 2.1.3-100 シングル値と温度の相関(2024/12/17~2025/3/11) (AFAS-B Bottom Fork 検出器)




図 2.1.3-101 シングル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-B Bottom Fork 検出器)




図 2.1.3-102 シングル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-B Bottom Fork 検出器)




図 2.1.3-103 シングル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-B Bottom Fork 検出器)




図 2.1.3-104 シングル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-B Bottom Fork 検出器)




図 2.1.3-105 シングル値と湿度の相関(2024/12/17~2025/3/11) (AFAS-B Bottom Fork 検出器)

⑥ AFAS-P Bottom Fork 検出器

・バックグラウンドのシングル値:図2.1.3-106~110

・シングル値と温度の相関:図2.1.3-111~115

・シングル値と湿度の相関:図2.1.3-116~120




図 2.1.3-106 バックグラウンドのシングル値(2024/4/5~2024/8/5) (AFAS-P Bottom Fork 検出器)




図 2.1.3-107 バックグラウンドのシングル値(2024/8/24~2024/10/9) (AFAS-P Bottom Fork 検出器)




図 2.1.3-108 バックグラウンドのシングル値(2024/10/11~2024/10/21) (AFAS-P Bottom Fork 検出器)




図 2.1.3-109 バックグラウンドのシングル値(2024/10/24~2024/12/17) (AFAS-P Bottom Fork 検出器)




図 2.1.3-110 バックグラウンドのシングル値(2024/12/17~2025/3/11) (AFAS-P Bottom Fork 検出器)




図 2.1.3-111 シングル値と温度の相関(2024/4/5~2024/8/5) (AFAS-P Bottom Fork 検出器)




図 2.1.3-112 シングル値と温度の相関(2024/8/24~2024/10/9) (AFAS-P Bottom Fork 検出器)




図 2.1.3-113 シングル値と温度の相関(2024/10/11~2024/10/21) (AFAS-P Bottom Fork 検出器)




図 2.1.3-114 シングル値と温度の相関(2024/10/24~2024/12/17) (AFAS-P Bottom Fork 検出器)




図 2.1.3-115 シングル値と温度の相関(2024/12/17~2025/3/11) (AFAS-P Bottom Fork 検出器)




図 2.1.3-116 シングル値と湿度の相関(2024/4/5~2024/8/5) (AFAS-P Bottom Fork 検出器)




図 2.1.3-117 シングル値と湿度の相関(2024/8/24~2024/10/9) (AFAS-P Bottom Fork 検出器)




図 2.1.3-118 シングル値と湿度の相関(2024/10/11~2024/10/21) (AFAS-P Bottom Fork 検出器)




図 2.1.3-119 シングル値と湿度の相関(2024/10/24~2024/12/17) (AFAS-P Bottom Fork 検出器)




図 2.1.3-120 シングル値と湿度の相関(2024/12/17~2025/3/11) (AFAS-P Bottom Fork 検出器)

(2) ²⁵²Cf 中性子線源測定

AFAS-B 及び AFAS-P において、2024 年 4 月から 2025 年 3 月までの間に月 1 回の頻度で 252 Cf 中性子線源の測定を行い、各月ごとに Fork 検出器においてはシングル値とその期待値、Collar 検出器においてはダブル値とその期待値の%Difference(図 $^{2.1.3-121}$ ~ 126 に示す)を求めた。各グラフ中の赤色の実線は各%Difference の平均値を示し、破線は管理限界値(各%Difference の誤差の二乗平均平方根を 3 倍した値を%Difference の平均値に加算及び減算し求めた値)を示す。管理限界値の上限値 UL 及び下限値 LL を求める式を (2.1) 及び (2.2) に示す。いずれの検出器においても各%Difference は、各管理限界値の範囲内となった。AFAS-P の各検出器については、UDL-1 による測定を合わせて行い、いずれも各管理限界値の範囲内となったが、AFAS-P の Collar 検出器においては、JSR-12 と比較して UDL-1 の%Difference が高くなる傾向がみられた。

また、各計数率の%Difference と温度及び湿度の相関の有無の確認結果を図 2.1.3-127 ~ 138 に示す。各測定において、温度は約 18 $^{\circ}$ $^{\circ}$ $^{\circ}$ $^{\circ}$ 、湿度は約 44 $^{\circ}$ $^{\circ}$ 64%の範囲で変動したが、これらの変動は計数率に影響を及ぼさなかった。

$$UL = \frac{1}{n} \sum_{i=1}^{n} \%Difference_i + 3 \times \sqrt{\frac{1}{n} \sum_{i=1}^{n} (\sigma\%Difference_i)^2} \cdot \cdot \cdot (2.1)$$

$$LL = \frac{1}{n} \sum_{i=1}^{n} \%Difference_i - 3 \times \sqrt{\frac{1}{n} \sum_{i=1}^{n} (\sigma\%Difference_i)^2} \cdot \cdot \cdot (2.2)$$

ただし、n は 2024 年 4 月から 2025 年 3 月までの間に月 1 回の頻度で測定した回数であり、

σ%Difference は計数率と期待値の%Difference の誤差である。(n = 1 は 2024 年 4 月のデータとする)




図 2.1.3-121 月ごとの%Difference の評価結果 (2024 年 4 月~2025 年 3 月) (AFAS-B Top Fork 検出器)




図 2.1.3-122 月ごとの%Difference の評価結果 (2024 年 4 月~2025 年 3 月) (AFAS-B Collar 検出器)




図 2.1.3-123 月ごとの%Difference の評価結果(2024 年 4 月~2025 年 3 月) (AFAS-B Bottom Fork 検出器)




図 2.1.3-124 月ごとの%Difference の評価結果(2024 年 4 月~2025 年 3 月) (AFAS-P Top Fork 検出器)




図 2.1.3-125 月ごとの%Difference の評価結果 (2024年4月~2025年3月) (AFAS-P Collar 検出器)




図 2.1.3-126 月ごとの%Difference の評価結果(2024 年 4 月~2025 年 3 月) (AFAS-P Bottom Fork 検出器)




図 2.1.3-127 %Difference と温度の相関(2024年4月~2025年3月) (AFAS-B Top Fork 検出器)




図 2.1.3-128 %Difference と温度の相関(2024 年 4 月~2025 年 3 月) (AFAS-B Collar 検出器)




図 2.1.3-129 %Difference と温度の相関(2024 年 4 月~2025 年 3 月) (AFAS-B Bottom Fork 検出器)




図 2.1.3-130 %Difference と温度の相関(2024年4月~2025年3月) (AFAS-P Top Fork 検出器)




図 2.1.3-131 %Difference と温度の相関(2024 年 4 月~2025 年 3 月) (AFAS-P Collar 検出器)




図 2.1.3-132 %Difference と温度の相関(2024 年 4 月~2025 年 3 月) (AFAS-P Bottom Fork 検出器)




図 2.1.3-133 %Difference と湿度の相関(2024年4月~2025年3月) (AFAS-B Top Fork 検出器)




図 2.1.3-134 %Difference と湿度の相関(2024 年 4 月~2025 年 3 月) (AFAS-B Collar 検出器)




図 2.1.3-135 %Difference と湿度の相関(2024年4月~2025年3月) (AFAS-B Bottom Fork 検出器)




図 2.1.3-136 %Difference と湿度の相関(2024年4月~2025年3月) (AFAS-P Top Fork 検出器)




図 2.1.3-137 %Difference と湿度の相関(2024 年 4 月~2025 年 3 月) (AFAS-P Collar 検出器)




図 2.1.3-138 %Difference と湿度の相関(2024 年 4 月~2025 年 3 月) (AFAS-P Bottom Fork 検出器)

2.1.4 考察

(1) バックグラウンド測定

AFAS-Pの Collar 検出器においてみられた、単発的な高計数率が全期間に複数回測定される事象については、シングル値及びダブル値において同時に確認されており、他の検出器では確認されなかったため、宇宙線や環境の変化による影響ではないと考えられる。本事象については、前年度(令和5年度)以前の性能確認試験(参考文献 2)においても同様の事象が確認されており、検出器の不具合により発生している事象である可能性があるため、LANL に報告済みである。なお、本事象については、通常、INCC ソフトウェアのQC Test 機能により排除されるため、測定に影響を及ぼすことはない。

AFAS-Bの Bottom Fork 検出器においてみられた、2024年6月から8月にかけてシングル値が徐々に増加する事象については、8月23日にデシカントを交換したことにより、図2.1.3-92に示すとおり、約2週間で6月以前のシングル値まで低下した。なお、デシカント交換時のインジケーターの指示値は、JAEAが定めるデシカントの交換基準を超過していた(*8)。また、図2.1.4-1に示すとおり、当該期間のシングル値と温度及び湿度に相関係数0.6程度の正の相関がみられた。以上のことから、本事象が発生した期間においてはAFAS-BのBottom Fork 検出器内部の湿度は高い状態にあり、測定環境の温度及び湿度の変動が測定に影響を及ぼしたと考えられる。

検出器内部の湿度が高い状態においては、検出器の種類に関わらず測定環境の温度及び湿度の変動が測定に影響を及ぼす可能性があるため、J-MOX においては AFAS 測定環境の温度及び湿度の変動範囲が当該測定と同等(温度:約23℃~26℃、湿度:約38%~72%)の場合には、検出器のデシカント交換を頻繁に行うことで測定に及ぼす影響を低減できると考えられる。

以上のように、いくつかの異常値が取得されたが、各検出器により得られたほぼ全てのシングル値は、誤差(3σ)の範囲内に収まっていた。このことから、測定期間中、各検出器は安定して動作していたと考えられる。

*8: 写真 2.1.4-1 に示すとおり、インジケーターには数字が印字されており、湿度が一定の値を超えると対応するエリアの色が青からピンクに変化する。本検出器は、2 か月に 1 回の頻度で行われた維持管理において、目視によりインジケーターの指示値を読み取り、JAEA が定めるデシカントの交換基準 (インジケーターの指示値が 60%) を超過しているものについてはデシカントを未使用のものと交換した。なお、本事象が発生するまでに維持管理は 4 月 5 日、6 月 5 日及び 8 月 2 日に実施されたが、いずれもインジケーターの指示値が 40%以下であり、交換基準を超過していなかった。



写真 2.1.4-1 検出器内の湿度確認用インジケーター (指示値が 40%の場合)




図 2.1.4-1 シングル値と温度及び湿度の相関(2024/4/5~2024/8/5) (AFAS-B Bottom Fork 検出器)

(2) ²⁵²Cf 中性子線源測定

いずれの検出器においても、各月の計数率と期待値の%Difference は、管理限界値の 範囲内となったことから、本期間において AFAS の性能は正常に維持されていたと考え られる。

また、AFAS-PのCollar 検出器においては、JSR-12と比較してUDL-1の測定結果から得られた%Difference が高くなる傾向がみられた。この原因は、表 2.1.4-1 に示すとおり、JSR-12と比較してクロック周波数の高いUDL-1ではパルス信号の数え落としが少なくなることでダブル値が高く計数されているためだと考えられる。

表 2.1.4-1 各計数装置のクロック周波数

計数装置	クロック周波数 (MHz)
JSR-12	4
UDL-1	100

2.2 パルス間隔分布の評価

2.2.1 目的

装置に不具合が発生した場合、その原因特定の判断材料となる、アンプ毎のパルス間隔分布を評価する。

2.2.2 方法

JAEA が所有するリストモジュール (PTR-32HV) を使用して ²⁵²Cf 中性子線源を測定し、AFAS の各検出器のアンプ毎のパルス間隔分布を評価する。また、本評価結果と前年度 (令和 5 年度) に実施した性能確認試験(^{参考文献 2)}における評価結果を比較する。令和 5 年度及び令和 6 年度の評価結果の例を図 2.2.2-1 に示す。

本評価は、AFAS-B及びAFAS-Pの各検出器について行う。

各検出器のアンプ数及び配置図を表 2.2.2-1 及び図 2.2.2-2~2.2.2-5 に示す。

本評価では、1.6 項の表 1.6-1 に示した ²⁵²Cf 中性子線源(T1-349)を使用する。




図 2.2.2-1 アンプ A のパルス間隔分布の例 (AFAS-B Collar 検出器)

検出器 アンプ個数 Top Fork 1 **AFAS-B** Collar 6 **Bottom Fork** 1 Top Fork 1 AFAS-P Collar 10 **Bottom Fork** 1

表 2.2.2-1 各検出器におけるアンプ数




図 2.2.2-2 AFAS-B Collar 検出器のアンプ配置図




図 2.2.2-3 AFAS-B Top Fork 及び Bottom Fork 検出器のアンプ配置図




図 2.2.2-4 AFAS-P Collar 検出器のアンプ配置図




図 2.2.2-5 AFAS-P Top Fork 及び Bottom Fork 検出器のアンプ配置図

詳細な試験手順を以下に示す。

- ① AFAS の各検出器を試験用架台に設置する。
- ② 線源固定治具を使用して、²⁵²Cf 中性子線源を任意の検出器にセットする。
- ③ 検出器にリストモジュールを接続する。
- ④ 検出器のジャンクションボックスの蓋を外す。
- ⑤ 任意のアンプ以外のアンプのジャンパーピン (写真 2.2.2-1 参照) を取り外し、計数機能を停止させる。
- ⑥ リストモジュール用ソフトウェアを使用してパルスを測定し、パルス間隔分布を 評価する。
 - ・AFAS-B 測定時間: Collar 検出器: 10 分、Fork 検出器: 100 分
 - ・AFAS-P 測定時間: Collar 検出器: 10 分、Fork 検出器: 100 分
- ⑦ 全てのアンプについて、⑤~⑥を実施する。
- ⑧ 全ての検出器について、②~⑦を実施する。

The second secon

ジャンパーピン

写真 2.2.2-1 AFAS のアンプ (AMPTEK A111)

2.2.3 結果




図 2.2.3-1 AFAS-B Collar 検出器 アンプ A のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-2 AFAS-B Collar 検出器 アンプ A のパルス間隔分布の比較 (0~100μs)




図 2.2.3-3 AFAS-B Collar 検出器 アンプ B のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-4 AFAS-B Collar 検出器 アンプ B のパルス間隔分布の比較 (0~100μs)




図 2.2.3-5 AFAS-B Collar 検出器 アンプ C のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-6 AFAS-B Collar 検出器 アンプ C のパルス間隔分布の比較 (0~100μs)




図 2.2.3-7 AFAS-B Collar 検出器 アンプ D のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-8 AFAS-B Collar 検出器 アンプ D のパルス間隔分布の比較 (0~100μs)




図 2.2.3-9 AFAS-B Collar 検出器 アンプEのパルス間隔分布の比較 (0~1000μs)




図 2.2.3-10 AFAS-B Collar 検出器 アンプEのパルス間隔分布の比較 (0~100μs)




図 2.2.3-11 AFAS-B Collar 検出器 アンプ F のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-12 AFAS-B Collar 検出器 アンプ F のパルス間隔分布の比較 (0~100μs)




図 2.2.3-13 AFAS-B Top Fork 検出器 パルス間隔分布の比較 (0~1000µs)




図 2.2.3-14 AFAS-B Top Fork 検出器 パルス間隔分布の比較 (0~100μs)




図 2.2.3-15 AFAS-B Bottom Fork 検出器のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-16 AFAS-B Bottom Fork 検出器のパルス間隔分布の比較 (0~100μs)




図 2.2.3-17 AFAS-P Collar 検出器 アンプ A のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-18 AFAS-P Collar 検出器 アンプ A のパルス間隔分布の比較 (0~100μs)




図 2.2.3-19 AFAS-P Collar 検出器 アンプ B のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-20 AFAS-P Collar 検出器 アンプ B のパルス間隔分布の比較 (0~100μs)




図 2.2.3-21 AFAS-P Collar 検出器 アンプ C のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-22 AFAS-P Collar 検出器 アンプ C のパルス間隔分布の比較 (0~100μs)




図 2.2.3-23 AFAS-P Collar 検出器 アンプ D のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-24 AFAS-P Collar 検出器 アンプ D のパルス間隔分布の比較 (0~100μs)




図 2.2.3-25 AFAS-P Collar 検出器 アンプEのパルス間隔分布の比較 (0~1000μs)




図 2.2.3-26 AFAS-P Collar 検出器 アンプEのパルス間隔分布の比較 (0~100μs)




図 2.2.3-27 AFAS-P Collar 検出器 アンプ F のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-28 AFAS-P Collar 検出器 アンプ F のパルス間隔分布の比較 (0~100μs)




図 2.2.3-29 AFAS-P Collar 検出器 アンプ G のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-30 AFAS-P Collar 検出器 アンプ G のパルス間隔分布の比較 (0~100μs)




図 2.2.3-31 AFAS-P Collar 検出器 アンプ H のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-32 AFAS-P Collar 検出器 アンプ H のパルス間隔分布の比較 (0~100μs)




図 2.2.3-33 AFAS-P Collar 検出器 アンプ I のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-34 AFAS-P Collar 検出器 アンプ I のパルス間隔分布の比較 (0~100μs)




図 2.2.3-35 AFAS-P Collar 検出器 アンプ J のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-36 AFAS-P Collar 検出器 アンプ J のパルス間隔分布の比較 (0~100μs)




図 2.2.3-37 AFAS-P Top Fork 検出器のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-38 AFAS-P Top Fork 検出器のパルス間隔分布の比較 (0~100μs)

AFAS-P Bottom Fork 検出器におけるアンプのパルス間隔分布を評価した。本評価結果と令和 5 年度(2023 年度)に実施した性能確認試験($^{(\otimes 5 + \chi m 2)}$)における評価結果を比較した結果($^{(o \sim 1000 \mu s)}$ 及び $^{(o \sim 1000 \mu s)}$ の分布)を図 2.2.3-39 及び図 2.2.3-40 に示す。各グラフ中の×印及び破線は令和 5 年度(2023 年度)に評価したパルス間隔分布及びその移動平均線を示し、 $^{(o \sim 1000 \mu s)}$ の印及び実線は令和 6 年度(2024 年度)に評価したパルス間隔分布及びその移動平均線を示す。




図 2.2.3-39 AFAS-P Bottom Fork 検出器のパルス間隔分布の比較 (0~1000μs)




図 2.2.3-40 AFAS-P Bottom Fork 検出器のパルス間隔分布の比較(0~100µs)

2.2.4 考察

本年度の試験では、前年度(令和5年度)の評価に使用した ²⁵²Cf 中性子線源と同じ線源を用いたため、線源の減衰により得られた計数値は減少したが、評価したパルス間隔分布は前年度に評価したパルス間隔分布とほぼ同一の形状となった。したがって、本年度の試験において得られたパルス間隔分布は正常であり、各アンプ、また、それぞれに接続されている ³He 比例計数管及びケーブルは健全に保たれていると考えられる。

2.3 プルトニウム 240 実効質量解析手法 Known M 法の適用性評価

2.3.1 目的

J-MOX における AFAS を使用した燃料集合体中のプルトニウム 240 実効質量(以下、 \lceil^{240} Pu 実効質量」という)の定量には、通常 Known α 法 $(^{*9})$ が適用される。

しかし、J-MOX において燃料集合体を長期保管した場合、アメリシウム 241 (以下、 $\lceil^{241}\text{Am}\rfloor$ という) の増加に伴い α 値が増加することにより、燃料集合体中の $\lceil^{240}\text{Pu}\rceil$ 実効質量を Known α 法により適切に評価できなくなる恐れがある。

そこで、本評価では 240 Pu 実効質量の定量に使用可能な別手法である Known M 法 $^{(*10)}$ の AFAS 測定への適用性を評価する。

- *9: Known α 法とは、 α 値を組成情報から計算することで既知とし、増倍 M 及び ²⁴⁰Pu 実効質量を解析する手法。不純物が混入しない MOX (PuO_2 - UO_2) 等に適用可能である。
- *10: Known M 法とは、増倍 M を核分裂性物質量から推定し、 α 値及び 240 Pu 実効質量 を解析する手法。 α 値の増加に影響を受けづらい手法である。

2.3.2 方法

本評価は、モンテカルロシミュレーションコード(以下、「MCNPX」という)を用いたシミュレーションにより行う。シミュレーションでは、J-MOX における AFAS による燃料集合体の測定を模擬したモデルを作成し、本モデルにおける AFAS の計数率を評価する。そして、シミュレーションにより得られた計数率を J-MOX において適用が想定される Known α 法及び別手法の Known m 法により解析し、両手法の m 実効質量の定量精度を比較することにより、AFAS 測定への Known m 法の適用性を評価する。なお、m 実効質量の各解析には、「INCC Software Users Manual m に記載されている式を使用する。

表 2.3.2-1 BWR 燃料集合体の同位体組成比

燃料集合体	²³⁸ Pu (%) ²³³ U (%)	²³⁹ Pu (%) ²³⁴ U (%)	²⁴⁰ Pu (%) ²³⁵ U (%)	²⁴¹ Pu (%) ²³⁶ U (%)	²⁴² Pu (%) ²³⁸ U (%)	²⁴¹ Am (%)
基準	1.257	61.085	25.759	7.301	4.598	1 126
左 毕	0.000	0.005	0.999	0.000	98.995	1.136
10 年保管	1.196	62.916	26.511	4.640	4.737	4.010
10 平休官	0.000	0.005	0.999	0.000	98.995	4.010
20 年保管	1.127	64.126	27.000	2.918	4.829	5 920
20 十休官	0.000	0.005	0.999	0.000	98.995	5.820

表 2.3.2-2 PWR 燃料集合体の同位体組成比

燃料集合体	²³⁸ Pu (%) ²³³ U (%)	²³⁹ Pu (%) ²³⁴ U (%)	²⁴⁰ Pu (%) ²³⁵ U (%)	²⁴¹ Pu (%) ²³⁶ U (%)	²⁴² Pu (%) ²³⁸ U (%)	²⁴¹ Am (%)
基準	1.784	58.817	25.076	8.461	5.862	1.900
本学 	0.000	0.006	0.720	0.000	99.275	1.900
10 年保管	1.707	60.883	25.937	5.404	6.069	5 264
10 平冰官	0.000	0.006	0.720	0.000	99.275	5.264
20 年保管	1.613	62.263	26.505	3.410	6.208	7.399
20 十休官	0.000	0.006	0.720	0.000	99.275	7.399

詳細な試験手順を以下に示す。

- ① MCNPX を用いて、J-MOX の AFAS 測定環境において燃料集合体が AFAS にて測定された場合の各検出器のシングル値及びダブル値を推定する。
- ② Known α 法を使用して、単位長さあたりの 240 Pu 実効質量 (g/cm) を評価し、有効長を乗じることにより、燃料集合体あたりの 240 Pu 実効質量 (g) を計算する。
- ③ Known M 法を使用して、単位長さあたりの 240 Pu 実効質量(g/cm)を評価し、有効長を乗じることにより、燃料集合体あたりの 240 Pu 実効質量(g/cm)を計算する。なお、解析に使用するキャリブレーションパラメータ(b_{kM} , c_{kM})は、増倍 M の異なる燃料集合体のモデルを複数作成し、キャリブレーションを行うことで評価する $^{(*11)}$ 。
- ④ ②及び③にて評価した 240 Pu 実効質量 (g) をそれぞれシミュレーションモデルに与えた 240 Pu 実効質量 (g) と比較する。
- ⑤ ①~④を AFAS-B 及び AFAS-P を対象に実施する。
- *11: 本評価では、MOX 燃料ペレットの密度 8.82(g/cm³)を基準として、密度を 8.72 から 8.92 まで 0.01 ずつ変動させた場合の燃料集合体のモデルをそれぞれ作成し、式(2.3) を使用してキャリブレーションを行った。評価したキャリブレーションパラメータを表 2.3.2-3 に示す。

$$M = 1 + b_{kM} m_{239e} + c_{kM} m_{239e}^2 \cdot \cdot \cdot (2.3)$$

ただし、 m_{239e} : 239 Pu 実効質量M: 中性子増倍

表 2.3.2-3 Known M 法による解析に使用したキャリブレーションパラメータ

	b_{kM}	c_{kM}
AFAS-B Collar 検出器	4.82E-06	0
AFAS-P Collar 検出器	2.03E-06	0

2.3.3 結果

各 AFAS の Collar 検出器による BWR 及び PWR 燃料集合体の測定について、シミュレーションから評価した 240 Pu 実効質量(以下、「シミュレーション値」という)とシミュレーションモデルに与えた 240 Pu 実効質量(以下、「モデル値」という)の%Difference を比較した。BWR 燃料集合体の測定結果を図 2.3.3-1 に、PWR 燃料集合体の測定結果を図 2.3.3-2 に示す。




図 2.3.3-1 BWR燃料集合体の²⁴⁰Pu 実効質量の%Difference の比較




図 2.3.3-2 PWR燃料集合体の²⁴⁰Pu 実効質量の%Difference の比較

2.3.4 考察

図 2.3.3-1 及び図 2.3.3-2 に示すとおり、Known M 法を使用して解析したシミュレーション値とモデル値の%Difference は、BWR 燃料集合体を測定した場合は 700%以上、PWR 燃料集合体を測定した場合は 800%以上だったことから、Known M 法では適切に解析できなかったと考えられる。

Known M 法において 240 Pu 実効質量を適切に解析できなかった理由は、「INCC Software Users Manual $^{(887)}$ 」に記載されている Known M 法の適用条件を満たしていないためである。「INCC Software Users Manual」に記載されている Known M 法の式(2.4)は、測定対象物全体を覆い測定する検出器のように、ダブル値の検出効率がシングル値の検出効率の 2 乗となる検出器に対して適用可能である。これに対して、燃料集合体を部分的に覆い測定する AFAS の Collar 検出器については、ダブル値の検出効率がシングル値の検出効率の 2 乗にはならないことを追加のシミュレーションにより確認しており、

「INCC Software Users Manual」に記載されている Known M 法の適用条件を満たさないため 240 Pu 実効質量を適切に求められなかった。参考として、シミュレーションにより評価した燃料集合体測定時における AFAS の Collar 検出器のシングル値とダブル値の検出効率を表 $^{2.3.4-1}$ に示す。

そこで、ダブル値の検出効率がシングル値の検出効率の2乗とならない場合においても適用可能な Known M 法の式を式(2.7)のとおり導出し、AFASの Collar 検出器を用いた燃料集合体測定に適用した。結果及び考察を2.3.5項及び2.3.6項に示す。

$$f_{M}(m_{239e}) = \frac{\varepsilon f_{d}M}{2} \left[k_{M} m_{239e} F_{0} \varepsilon M \nu_{s2} + \left(\frac{M-1}{\nu_{i1}-1} \right) \nu_{i2} S \right] - D = 0 \qquad \cdot \cdot \cdot (2.4)$$

$$Singles = mF_0 \varepsilon_S \nu_{s1} M(1+a) \qquad \qquad \cdot \cdot \cdot (2.5)$$

$$Doubles = \frac{1}{2} m F_0 \varepsilon_D f_d v_{s2} M^2 \left[1 + \left(\frac{M-1}{v_{i1}-1} \right) \frac{v_{s1} v_{i2}}{v_{s2}} (1+\alpha) \right] \qquad (2.6)$$

$$f_{M}(m_{239e}) = \frac{\varepsilon_{D}f_{d}M}{2\varepsilon_{S}} \left[k_{M}m_{239e}F_{0}\varepsilon_{S}M\nu_{s2} + \left(\frac{M-1}{\nu_{i1}-1}\right)\nu_{i2}S \right] - D = 0 \quad \cdot \cdot \cdot (2.7)$$

ただし、Singles (S): 増倍補正前のシングル値

Doubles (D): 増倍補正前のダブル値

m: 240Pu 実効質量

m_{239e}: ²³⁹Pu 実効質量

M: 中性子增倍

a: 自発核分裂と (α, n) 反応により生じた中性子の比

fa: ダブルゲートフラクション

 ν_{s1} , ν_{s2} : 自発核分裂中性子の階乗モーメント

ν_{i1}, ν_{i2}: 誘発核分裂中性子の階乗モーメント

F₀: ²⁴⁰Pu の自発核分裂率

 k_M : 240 Pu 実効質量と 239 Pu 実効質量の比(m/m_{239e})

表 2.3.4-1 シミュレーションから求めた各検出効率

	\mathcal{E}_{S}	$arepsilon_D$	
AFAS-B Collar 検出器において	0.027	0.005	
BWR 燃料集合体を測定した場合	0.027	0.005	
AFAS-P Collar 検出器において	0.025	0.004	
PWR 燃料集合体を測定した場合	0.025	0.004	

2.3.5 結果 (再解析後)

新たに導出した Known M 法の式を適用し、各 AFAS の Collar 検出器による BWR 及び PWR 燃料集合体の測定について、シミュレーション値とモデル値の%Difference を比較した。BWR 燃料集合体の測定結果を図 2.3.5-1 に、PWR 燃料集合体の測定結果を図 2.3.5-2 に示す。




図 2.3.5-1 BWR燃料集合体の²⁴⁰Pu 実効質量の%Difference の比較




図 2.3.5-2 PWR燃料集合体の²⁴⁰Pu 実効質量の%Difference の比較

2.3.6 考察 (再解析後)

図 2.3.5-1 及び図 2.3.5-2 に示すとおり、AFAS の Collar 検出器を用いた BWR 及び PWR 燃料集合体の測定に新たに導出した Known M 法の式を適用した場合、 240 Pu 実効質量の定量誤差を示す%Difference は、燃料集合体の保管年数の増加に伴い増加し、本評価において最長となる 20 年保管のとき、BWR 燃料集合体について約-0.89%及び PWR 燃料集合体について約-1.56%となった。また、Known M 法の%Difference は、J-MOX において適用予定の Known α 法と比較して同程度または小さかったことから、Known M 法のAFAS 測定への適用性を確認した。

なお、本評価では測定における計数誤差を考慮していないため、J-MOX における実際の測定では、計数誤差があることを留意する必要がある。

3. AVIS の性能確認試

3.1 パルス間隔分布の評価

3.1.1 目的

装置に不具合が発生した場合、その原因特定の判断材料となる、アンプ毎のパルス間隔分布を評価する。

3.1.2 方法

JAEA が所有するリストモジュール (PTR-32HV) を使用して ²⁵²Cf 中性子線源を測定し、AVIS の中性子検出器のアンプ毎のパルス間隔分布を評価する。また、本評価結果と前年度 (令和 5 年度) に実施した性能確認試験(^{参考文献 2)}における評価結果を比較する。令和 5 年度及び令和 6 年度の評価結果の例を図 3.1.2-1 に示す。

AVIS の中性子検出器のアンプ数及び配置図を表 3.1.2-1 及び図 3.1.2-2 に示す。本評価では、1.6 項の表 1.6-1 に示した 252 Cf 中性子線源(T1-349)を使用する。




図 3.1.2-1 AVIS の中性子検出器 アンプ A のパルス間隔分布の例

表 3.1.2-1 AVIS におけるアンプ数

7	アンプ 個数	
AVIS	Primary 検出器	7
	Secondary 検出器	7




図 3.1.2-2 AVIS のアンプ配置図

詳細な試験手順を以下に示す。

- ① 線源固定治具を使用して、252Cf中性子線源を検出器にセットする。
- ② 検出器にリストモジュールを接続する。
- ③ 検出器のジャンクションボックスの蓋を外す。
- ④ 任意のアンプ以外のアンプのジャンパーピンを取り外し、計数機能を停止させる。
- ⑤ リストモジュール用ソフトウェアを使用してパルスを 10 分間測定し、パルス間隔分布を評価する。
- ⑥ 全てのアンプについて、④~⑤を実施する。

3.1.3 結果




図 3.1.3-1 アンプ A のパルス間隔分布の比較(0~1000µs)




図 3.1.3-2 アンプAのパルス間隔分布の比較(0~100µs)




図 3.1.3-3 アンプBのパルス間隔分布の比較 (0~1000µs)




図 3.1.3-4 アンプBのパルス間隔分布の比較 (0~100µs)




図 3.1.3-5 アンプ C のパルス間隔分布の比較 (0~1000µs)




図 3.1.3-6 アンプ C のパルス間隔分布の比較 (0~100µs)




図 3.1.3-7 アンプ D のパルス間隔分布の比較 (0~1000µs)




図 3.1.3-8 アンプ D のパルス間隔分布の比較 (0~100µs)




図 3.1.3-9 アンプEのパルス間隔分布の比較 (0~1000µs)




図 3.1.3-10 アンプEのパルス間隔分布の比較 (0~100µs)




図 3.1.3-11 アンプFのパルス間隔分布の比較 (0~1000µs)




図 3.1.3-12 アンプFのパルス間隔分布の比較 (0~100µs)




図 3.1.3-13 アンプ G のパルス間隔分布の比較 (0~1000µs)




図 3.1.3-14 アンプ G のパルス間隔分布の比較 (0~100µs)




図 3.1.3-15 アンプ H のパルス間隔分布の比較 (0~1000µs)




図 3.1.3-16 アンプ H のパルス間隔分布の比較 (0~100µs)




図 3.1.3-17 アンプIのパルス間隔分布の比較 (0~1000µs)




図 3.1.3-18 アンプIのパルス間隔分布の比較 (0~100µs)




図 3.1.3-19 アンプJのパルス間隔分布の比較 (0~1000µs)




図 3.1.3-20 アンプJのパルス間隔分布の比較 (0~100µs)




図 3.1.3-21 アンプ K のパルス間隔分布の比較 (0~1000µs)




図 3.1.3-22 アンプ K のパルス間隔分布の比較 (0~100µs)




図 3.1.3-23 アンプLのパルス間隔分布の比較 (0~1000µs)




図 3.1.3-24 アンプLのパルス間隔分布の比較 (0~100µs)




図 3.1.3-25 アンプ M のパルス間隔分布の比較 (0~1000µs)




図 3.1.3-26 アンプ M のパルス間隔分布の比較 (0~100µs)




図 3.1.3-27 アンプ N のパルス間隔分布の比較 (0~1000µs)




図 3.1.3-28 アンプNのパルス間隔分布の比較 (0~100µs)

3.1.4 考察

2.2 項の AFAS における評価と同様に、本年度の試験では、前年度(令和 5 年度)の評価に使用した ²⁵²Cf 中性子線源と同じ線源を用いたため、線源の減衰により得られた計数値は減少したが、本年度評価したパルス間隔分布は前年度とほぼ同一の形状となり、異常な分布は確認されなかった。したがって、本年度の試験において得られたパルス間隔分布は正常であり、各アンプ、また、それぞれに接続されている ³He 比例計数管及びケーブルは健全に保たれていると考えられる。

4. IPCA2 の性能確認試験

4.1 長期管理限界の妥当性確認

4.1.1 目的

IPCA2 を構成する中性子検出器、HPGe(高純度ゲルマニウム)ガンマ線検出器及びロードセル(荷重変換器)から得られる計測値の長期管理限界の妥当性確認(長期管理限界内に収まっているか否かの確認)を定期的に行う。

4.1.2 方法

中性子検出器については、連続的なバックグラウンド測定及び定期的な Pu 線源及び Cm-244 線源の測定を実施し、検出器の長期安定性の評価及び得られた計数値とこれらから評価された検出器の中性子検出効率が長期管理限界内に収まっていることを確認する。また、本測定期間における測定環境の温度及び湿度を測定し、計数値または中性子検出効率との相関の有無を確認する。さらに、IAEA の要望に従い、適宜 IAEA から支給された UDL-1 による性能確認試験を行う。

HPGe ガンマ線検出器ついては、令和2年度から発生している全てのガンマ線検出器において性能が低下する事象の確認を実施したため、測定は行われなかった。

ロードセルについては、定期的な分銅測定を実施し、得られた重量が長期管理限界内 に収まっていることを確認する。

詳細な試験手順を以下に示す。

- (1) 中性子検出器におけるバックグラウンド測定
 - ① JSR-15 を使用して連続的なバックグラウンド測定を実施する。
 - ② ①により取得した計数値における異常値の有無を確認し、検出器の長期安定性を評価する。
 - ③ ①~②を月1回の頻度で実施する。
- (2) 中性子検出器における Pu 線源及び Cm-244 線源測定
 - ① 線源 (Pu 線源または Cm-244 線源) を検出器にセットする。
 - ② JSR-15 及び UDL-1 を使用して線源測定を実施する。また、測定環境の温湿度測定を実施する。
 - ③ Pu 線源測定により取得したシングル値から求めた中性子検出効率及び Cm-244 線源測定により取得したシングル値が長期管理限界内に収まっていることを確 認する。また、測定結果と温湿度の相関の有無を評価する。
 - ④ ①~③を定期的に実施する。

- (3) ロードセルにおける分銅測定
 - ① 分銅をロードセルにセットする。
 - ② 分銅測定を実施する。
 - ③ ①~②を定期的に実施する。

4.1.3 結果

令和6年4月から令和7年3月の期間において、LANLが実施したIPCA2の長期管理限界の妥当性確認結果及びUDL-1による性能確認試験結果を添付資料に示す。いずれの項目においても測定結果は管理値以内であった。

4.1.4 考察

実施した試験の各測定結果は、いずれも管理値以内であったことから、本期間において IPCA2 の性能は正常に維持されており、長期間安定して動作していたと考えられる。

4.2 ハードウェアの修理及び動作確認

4.2.1 目的

IPCA2 に接続されている Power Supply Box について不具合が確認されたため、修理及び修理後の動作確認を行う。

4.2.2 方法

- ① Power Supply Box を新品に交換する。
- ② 新品の Power Supply Box に AC 電源を接続する。
- ③ 交換した機器が正常に動作することを確認する。

4.2.3 結果

LANL が実施した IPCA2 のハードウェアの修理及び動作確認の結果を添付資料に示す。

令和 5 年 7 月に発生した Cm-244 線源測定の際に計数率が低下した事象について、 LANL が原因調査を行った。調査結果に基づき、不具合が生じていた Power Supply Box を新品に交換することで上記事象は解消した。

4.2.4 考察

LANL が実施した調査の結果、Power Supply Box を接続していた UPS が故障したため、Power Supply Box に供給される交流電圧が不十分となり、Power Supply Box が故障したことを特定した。なお、令和7年度にLANLからJ-MOX~IPCA2の出荷が予定されているため、UPSの交換については実施していない。

4.3 測定パラメータの評価

4.3.1 目的

IPCA2 の性能を確認するため、測定パラメータの評価を行う。

4.3.2 方法

- ① HV プラトーの測定を行い、適切な供給電圧を評価する。
- ② 各ゲート幅 (10,20,40,80,160μs) におけるダブル値を測定し、適切なゲート幅及 びダイアウェイタイムを評価する。
- ③ 評価した測定パラメータと IPCA2 の動作設定に用いられる標準的な測定パラメータを比較する。

4.3.3 結果

LANL が実施した測定パラメータの評価結果を添付資料に示す。評価した測定パラメータと IPCA2 の動作設定に用いられる標準的な測定パラメータを比較した結果、概ね一致した。

4.3.4 考察

評価した測定パラメータは標準的な測定パラメータと概ね一致したため、IPCA2 の性能は一定に保たれていることを確認した。

5. 結論

AFAS の性能確認試験では、「長期管理限界の妥当性確認」、「パルス間隔分布の評価」、 及び「プルトニウム 240 実効質量解析手法 Known M 法の適用性評価」を実施した。

「長期管理限界の妥当性確認」では、AFAS の長期安定性を確認するため、令和 6 年 4 月から令和 7 年 3 月までの期間において連続的なバックグラウンド測定及び定期的な ²⁵²Cf 中性子線源測定を実施した。その結果、各検出器において得られたほぼ全てのバックグラウンドの測定値及び ²⁵²Cf 中性子線源測定の計数率は長期管理限界の範囲内に収まったことから、本測定期間中、AFAS は正常な状態を維持し、長期間安定して動作していたことを確認した。

「パルス間隔分布の評価」では、装置の不具合の原因を特定する際の判断材料となるアンプ毎のパルス間隔分布を取得するため、リストモジュールを使用して ²⁵²Cf 中性子線源を測定した。その結果、得られたパルス間隔分布には、異常は確認されず、また、令和 5 年度に得られたパルス間隔分布と同様の傾向を示したことから、本評価において得られたパルス間隔分布は正常であり、各アンプ、また、それぞれに接続されている ³He 比例計数管及びケーブルは健全に保たれていることを確認した。

「プルトニウム 240 実効質量解析手法 Known M 法の適用性評価」では、長期保管された燃料集合体の測定に対して Known M 法の適用性を確認するため、モンテカルロシミュレーション(MCNPX)による評価を実施し、Known α 法及び Known M 法を使用して解析した燃料集合体中のプルトニウム 240 実効質量を比較した。その結果、AFAS の Collar 検出器において燃料集合体を測定する場合、Known M 法においてプルトニウム 240 実効質量を適切に解析できなかった。したがって、AFAS の Collar 検出器を用いた BWR 及び PWR 燃料集合体の測定に適用可能な Known M 法の式を新たに導出し、当該式を適用することで Known M 法の AFAS 測定への適用性を確認した。

AVIS の性能確認試験「パルス間隔分布の評価」では、AFAS における評価と同様に、リストモジュールを使用して ²⁵²Cf 中性子線源を測定した。その結果、得られたパルス間隔分布には、異常は確認されず、令和 5 年度に得られたパルス間隔分布と同様の傾向を示したことから、本評価において得られたパルス間隔分布は正常であり、各アンプ、また、それぞれに接続されている ³He 比例計数管及びケーブルは健全に保たれていることを確認した。

IPCA2 の性能確認試験では、「長期管理限界の妥当性確認」、「ハードウェアの修理及び動作確認」及び「測定パラメータの評価」を実施した。

「長期管理限界の妥当性確認」では、IPCA2 の長期安定性及び IPCA2 の各検出器及びロードセルにおいて得られた計測値が長期管理限界内に収まることを確認するため、令和6年4月から令和7年3月までの期間において、中性子検出器については、連続的なバックグラウンド測定及び定期的な中性子線源測定を、ロードセルについては、定期的な分銅測定を実施した。その結果、中性子検出器及びロードセルにおいて取得された計数値及び重量測定値は長期管理限界の範囲内に収まったことから、本測定期間中、

IPCA2 は正常な状態を維持し、長期間安定して動作したことを確認した。また、UDL-1 については、定期的な中性子線源測定において JSR-15 と同様の測定性能であることを確認した。

「ハードウェアの修理及び動作確認」では、令和5年7月に発生した不具合事象の原因を特定し、令和7年1月に IPCA2 に接続されている Power Supply Box の交換を行った。その結果、交換した機器は正常に動作し、不具合事象が発生しないことを確認した。

「測定パラメータの評価」では、IPCA2の性能を確認するため、供給電圧、ゲート幅及びダイアウェイタイムの評価を行い、IPCA2の動作設定に用いられる標準的な測定パラメータと比較した。その結果、評価した測定パラメータは標準的な測定パラメータと概ね一致したため、IPCA2の性能は一定に保たれていることを確認した。

以上

【参考文献】

- 参考文献 1: M. T. Swinhoe, J. B. Marlow and H.O. Menlove, "Technical Specification for the Improved Plutonium Canister Assay System (IPCA2)" (2009)
- 参考文献 2: 大型混合酸化物燃料加工施設保障措置機器性能確認試験報告書 (令和 5 年度)
- 参考文献 3: "INCC Software Users Manual" (2009) (LA-UR-01-6761)
- 参考文献 4: "User Requirements AFAS Version No.1, IAEA" (2007) (SG-TE-GNRL-ZZ-1010)
- 参考文献 5: M. T. Swinhoe, "Simulation of BWR MOX Fuel in AFAS" (2015) (LA-CP-15-20176)
- 参考文献 6: M. T. Swinhoe, H.O. Menlove, C. D. Real and J. B. Marlow, "Factory Acceptance Test Report for the Advanced Fuel Assembly Assay System (AFAS) PWR" (2008) (LA-CP-08-0730)

添付資料

【線源情報】

線源証明書



Tel 661·309·1010 Fax 661·257·8303

NOMINAL SOURCE CERTIFICATE

Customer: Daiichi Clarity Company Limited

Purchase Order No.: CA619

Model No.: N-252

Catalog No.: CF230140100U

Capsule Type: A3014

Active Diameter/Mass: 1.6 mm (0.062 ")

Cover: Stainless steel

Backing: Stainless steel

Certificate Date: 04-Oct-10

Quantity: 1

SS&DR No.: CA406S102S

ISO Classification: ISO/99/C66535

Special Form No.: USA/0351/S Rev 6

Nuclide Half Life: 2.645 ± 0.008 years

Recommended Working Life: 15 years

Nuclide	Source No.	Activity	Neutron Output [neutrons/second]	Reference Date
Cf-252	H4-694	100 μCi/3.7 MBq	4.24E+05	15-Oct-10

Impurities: See Technical Data sheet.

Leak Test Information is on Reverse Side:

Remarks:

- This document uses the numerical convention where 1.000 = 1 and $1,000 = 10^3$.
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.

Name Signature Date

Notebook Page: 1465-40

- ISO 9001 CERTIFIED -

Medical Imaging Laboratory

24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory

1800 North Keystone Street Burbank, California 91504

Tel 661·309·1010 Fax 661·257·8303

Cf-252 Technical data

The Cf-252 used to prepare your order was taken from Eckert & Ziegler Isotope Products Laboratories Lot #5343201 and it had the following composition as of 20 Sep 10.

Mass %	Activity %
9.013	0.0886
11.904	3.1129
3.956	0.0151
75.126	96.7830
0.00002	0.00044
	9.013 11.904 3.956 75.126

The Cm-248 decay product was last separated on 17 Sep 09

Isotopic composition provided by Oak Ridge National Laboratory

If you have any questions, please contact Eckert & Ziegler Isotope Products Technical Service: 661-309-1010

ISO 9001 CERTIFIED



Tel 661·309·1010 Fax 661·257·8303

NOMINAL SOURCE CERTIFICATE

Customer: Daiichi Clarity Co., Ltd.

Purchase Order No.: DN125

Model No.: N-252

Catalog No.: CF230140100U

Capsule Type: A3014-01
Active Diameter: 0.062" (1.6 mm)

Cover: Stainless Steel

Backing: Stainless Steel

Certificate Date: 2020-09-25

Quantity: 1

SS&DR No.: CA0406S102S

ISO/ANSI Classification: ANSI 77C66535

Special Form No.: USA/0351/S-96 Rev 10

Nuclide Half Life: 2.645 ± 0.008 years Recommended Working Life: 15 years

	Nuclide	Source No.	Activity	Radiation Output	Reference Date
L	Cf-252	T1-349	3.7 MBq (100 μCi)	3.92 E+05 n/s	2020-10-01

Impurities: See Technical Data Sheet

Leak Test Information is on Reverse Side:

Remarks:

- This document uses the numerical convention where 1.000 = 1 and 1,000 = 10³.
- This document uses the date convention YYYY-MM-DD in accordance with ISO 8601.
- Nuclear data were taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- ANSI classification is equivalent to ISO2919.

		2020-09-25
Name	Signature	Date

Notebook Page: 2190-01

ISO 9001 CERTIFIED -

Industrial Gauging Laboratory

1800 North Keystone Street Burbank, California 91504



Tel 661•309•1010 Fax 661•257•8303 www.ezag.com

Cf-252 Technical Data

The Cf-252 used to prepare your order of source T1-349 was taken from Eckert & Ziegler Isotope Products Laboratories Lot #6050711. It had the following composition as of 2020-08-19.

Nuclide Nuclide	Mass %	Activity %
Cf-249	14.572	0.1723
Cf-250	17.660	5.553
Cf-251	6.859	0.0314
Cf-252	60.910	94.243

The Cm-248 decay product was last separated on 2018-04-19.

Isotopic composition provided by Oak Ridge National Laboratory

If you have any questions, please contact Eckert & Ziegler Isotope Products Technical Service: 661-309-1010



ISO 9001 CERTIFIED

【AFAS 性能確認試験】

(1) 2.1 長期管理限界の妥当性確認

Results

Cycle

8 9 10

Cycle

234567

8 9 10

Passive cycle raw data

Passive cycle rate data

Singles 76. 183 76. 050 76. 317 76. 383 75. 850 75. 933 77. 517 77. 833 77. 700 78. 800

```
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                           24. 04. 05 14:36:13
44503613. VER
                                                                                                           BWR TF
                                                                                                           Pu
0.000
Verification
Review disk file
                                                                                                           On
Sample method
Measured
     Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am date:
                                                                                                        0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.04.05
0.0000 +-
24.04.05
                                                                                                                                                                    0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                     0.0000
                                                                                                                                                                                                                                                                           0.0000
                                                                                                               1. 50
64. 00
64. 00
1720
50. 0000
0. 0080
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                               Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                                            1.0000 +-
   Normalization constant:
                                                                                                                                                                                   0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                           0.000 +-
0.000 +-
0.000 +-
0.000
0.000
                                                                                                                                                                                    0.000
0.000
0.000
```

202404_AFAS-B_Bottom Fork.txt

(1)

10 60

INCC 5.1.2	
Facility: Material balance area: Detector type: Detector id: Electronics id: Inventory change code: I/O code: Measurement date: Results file name: Inspection number: Ltem id: Stratum id: Material type: Original declared mass: Measurement option: Data source: CC tests: Error calculation: Accidentals method: Inspector name: Passive comment:	JMOX JM26 AFAS JSR 02 JSR-12 24.04.05 15:04:21 445P0421.VER BWR BF XXXXX Pu 0.000 Verification Review disk file On Sample method Measured
Isotopics id: Isotopics source code: Pu238: Pu239:	Default 0D 0.0000 +- 0.0000 0.0 0.00

Number passive cycles: Count time (sec):

Passive error messages Known alpha analysis error

ISOLODICS IG.	peraurt			
Isotopics source code:	OD			
Pu238:	0.0000 +-	0.0000	0.0000 +-	0.0000
Pu239:	0.0000 +-	0.0000	0.0000 +-	0.0000
Pu240:	100.0000 +-	0.0000	100.0000 +-	0.0000
Pu241:	0.0000 +-	0.0000	0.0000 +-	0.0000
Pu242:	0.0000 +-	0.0000	0.0000 +-	0.0000
Pu date:	00. 01. 01		24. 04. 05	
Am241:	0.0000 +-	0.0000	0.0000 +-	0.0000
Am date:	00. 01. 01		24. 04. 05	
Predelav:	1. 50			

Predelay: 1.50
Gate length: 64.00
2nd gate length: 64.00
High voltage: 1720
Die away time: 50.0000
Efficiency: 0.0060
Multiplicity deadtime: 0.0000
Coefficient A deadtime: 0.0000
Coefficient B deadtime: 0.0000
Coefficient C deadtime: 0.0000
Doubles gate fraction: 0.0001
Normalization constant: 1.0000 ←

Passive singles bkgrnd: 0.000 +- 0.000
Passive doubles bkgrnd: 0.000 +- 0.000
Passive triples bkgrnd: 0.000 +- 0.000
Passive scaler1 bkgrnd: 0.000
Passive scaler2 bkgrnd: 0.000

Number passive cycles: 10 Count time (sec): 60

Passive error messages

No known alpha calibration

202404_AFAS-B_Bottom Fork.txt

(2)

76. 857 +-0. 383 +-0. 000 +-132. 323 +-0. 000 +-

Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000

> Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests Pass Pass

Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

> Doubles 0.367 0.300 0.567 0.183 0.217 0.417 0.300 0.617 0.500

Results						
		Sing Doub Trip Scale Scale	les: les: er 1:	136. 997 1. 320 0. 000 4. 405 0. 000	+- 0.0 +- 0.0 +- 0.0	93 00 99
Passive	cycle raw da	ata				
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 8205 8225 8240 8164 8120 8368 8204 8134 8270 8268	R+A 147 143 172 144 140 155 169 190 156 133	A 76 74 85 80 76 68 74 73 78 73	Scaler1 263 300 270 260 247 279 284 248 250 242	Scaler2 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Passive	cycle rate o	data				
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 136, 750 137, 083 137, 333 136, 067 135, 333 139, 467 136, 733 135, 567 137, 833 137, 800	Doub1 1. 1 1. 1 1. 4 1. 0 1. 4 1. 5 1. 9 1. 3	83 50 50 67 67 50 83 50	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Pass Pass

0.0000

Passive cycle raw data

Results

8 9 10

202404_AFAS-B_Bottom Fork.txt

```
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                        24. 04. 05 15:04:21
445P0421. VER
                                                                                        BWR BF
                                                                                        Pu
0.000
Verification
Review disk file
                                                                                        On
Sample method
Measured
                                                                                      Isotopics id:
     Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu date:
Am date:
                                                                                                                                                                          0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.04.05
                                                                                                                                      0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                        0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                        0.0000
                                                                                                                                                                                                                          0.0000
                                                                                                                                                                           0.0000
24.04.05
                                                                                           1. 50
64. 00
64. 00
1720
50. 0000
0. 0060
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                          Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                      1.0000 +-
   Normalization constant:
                                                                                                                                                  0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                    0.000 +
0.000 +
0.000 +
0.000 +
0.000
                                                                                                                                                   0.000
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                 10
60
Passive error messages
No known alpha calibration
```

Scaler2 0 0 0 0 0 0 0 0 0 Singles 8205 Cycle R+A 147 143 172 144 140 155 169 190 156 133 Scaler1 QC Tests 76 74 85 80 76 68 74 73 78 73 263 300 270 260 247 279 284 248 250 242 8225 8240 8164 8120 8368 8204 8134 8270 8 9 10 8268 Passive cycle rate data Singles 136, 750 137, 083 137, 333 136, 067 135, 333 139, 467 136, 733 135, 567 137, 833 137, 800 QC Tests Pass Pass Doubles 1. 183 1. 150 Cycle Triples Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 234567 1. 450 1. 067 1. 067 1. 450 1. 583 1. 950 1. 300 1. 000

Singles: Doubles: Triples: Scaler 1: Scaler 2:

136. 997 +-1. 320 +-0. 000 +-4. 405 +-0. 000 +-

0. 385 0. 093 0. 000 0. 099 0. 000

Pass

202404_AFAS-P_Top Fork.txt

(1)

Facility: Material balance area: Detector type: Detector id: Electronics id:	JMOX JM2G AFAS JSR_02 JSR-12
Inventory change code:	0311-12
I/O code:	
Measurement date:	24. 04. 05
Results file name:	445P1825
Inspection number:	
Item id:	PWK IF

INCC 5.1.2

05 15:18:25 25. VER PWR TF XXXX Pu 0.000 Verification Review disk file On Sample method Stratum id: Stratum id.
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Passive comment:

Sample method Measured

Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.04.05 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient Geadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0.0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

10 60

Passive error messages

No known alpha calibration

202404_AFAS-P_Top Fork.txt

(2)

Results

2. 762 +-0. 000 +-0. 000 +-120. 037 +-0. 000 +-0.069 0.000 0.000 0.290 0.000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4 5 6 7 8 9	Singles 178 163 157 161 142 178 172 172 183 151	R+A 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	A 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	Scaler1 7255 7240 7262 7193 7204 7105 7197 7259 7123 7184	0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Passive	cycle rate	data	Ů	,,,,,	v	. 400

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	2. 967	0.000	0.000	0.000	Pass
2	2, 717	0.000	0.000	0.000	Pass
3	2, 617	0.000	0.000	0.000	Pass
4	2, 683	0.000	0.000	0.000	Pass
5	2, 367	0.000	0.000	0.000	Pass
6	2, 967	0.000	0.000	0.000	Pass
7	2. 867	0.000	0.000	0.000	Pass
8	2. 867	0.000	0.000	0.000	Pass
9	3, 050	0.000	0.000	0.000	Pass
10	2, 517	0.000	0.000	0.000	Pass

202404_AFAS-P Top Fork(UDL-1).txt

INCC 6.23.2.9 HM32 Not Validated

6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.04.05 15:
Results file name: 445P2035.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx 15:20:35

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 2. 370 0. 000 0. 000 0. 000 0. 000 10 60. 000

Summed raw data

Number of good cycles: Total count time: Shift register singles sum: Shift register accidentals sum: Shift register accidentals sum: Shift register 1st scaler sum: Shift register 1st scaler sum: 600 71972 1232 532 0

Summed multiplicity distributions

R+A sums A sums

Results

Singles: Doubles: Triples: Scaler 1: 117. 583 1. 167 0. 000 0. 000 0. 000 0. 305 0. 072 0. 000 0. 000 0. 000 Scaler 2

(1)

202404_AFAS-P_Collar.txt

INCC 5.1.2

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS 24. 04. 05 15:34:29 445P3429. VER PWR COLLAR XXXXX Pu 0.000 Verification Review disk file On Sample method Sample method Measured

0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.04.05 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000 +-24.04.05

0.0000

1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1 0000 +-Normalization constant:

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

Passive error messages

No passive calibration curve calibration No known alpha calibration

Normalization results for reference source: H4-694

60.000

Current normalization constant: 1.000 + 0

Cf252 expected singles rate: 116.321 + 0

Cf252 measured singles rate: 117.583 + 0

Singles rate expected/measured: 0.989 + 0

New normalization constant: 1.000 + 0

Normalization test -- data quality is good.

Cycle raw data Count time (sec):

Cycle 1 2 3 4 5 6 7 8 9 10	Singles 7216 7222 7296 7180 7185 7262 7129 7179 7195 7108	R+A 108 141 133 142 109 122 115 115 115 113	A 56 61 58 49 60 53 57 37 43 58	Scaler1 0 0 0 0 0 0 0 0 0 0	0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Cycle D	TC rate data					
Cycle	Singles	Doubles		Triples	QC Tests	

3 119. 230 1. 250 0. 000 Pass 4 117. 297 1. 550 0. 000 Pass	sts
5 117, 380 0, 817 0, 000 Pass 6 118, 663 1, 150 0, 000 Pass 7 116, 447 0, 967 0, 000 Pass 8 117, 280 1, 300 0, 000 Pass 9 117, 547 1, 167 0, 000 Pass 10 116, 097 1, 267 0, 000 Pass	sts

(2)

202404_AFAS-P_Collar.txt

Results

		Dou Tri Scal	ngles: ubles: ples: er 1: er 2:	2055. 722 350. 522 0. 000 17. 933 0. 000	+- +- +-	3. 0 1. 5 0. 0 0. 1 0. 0	93 00 95
Passiv	e cycle raw	data					
Cycle 1 2	Singles 123594 122597	R+A 37518 36892	A 16373 16240	Scaler1 1096 1052	Scale	r2 0 0	QC Tes Pass Pass

Cycle	STREES	Ν±Ν	A	Scaleri	Scalerz	WO LESTS
1	123594	37518	16373	1096	0	Pass
2	122597	36892	16240	1052	0	Pass
3	122948	37107	15973	1041	0	Pass
4	123033	37335	16001	1042	0	Pass
5	122520	36782	15982	1041	0	Pass
6	124376	37562	16497	1104	0	Pass
7	123652	37301	16050	1150	0	Pass
8	123331	36751	16320	1075	0	Pass
9	123914	37522	16518	1052	0	Pass
10	123249	37459	16111	1107	0	Pass

Passive cycle rate data

ycle	Singles	Doubles	Triples	Mass	QC Tests
1	2060, 267	352, 668	0.000	0.000	Pass
2	2043, 644	344, 443	0.000	0.000	Pass
3	2049, 496	352, 483	0.000	0.000	Pass
4	2050, 914	355, 819	0.000	0.000	Pass
5	2042, 361	346, 912	0.000	0.000	Pass
6	2073, 305	351, 335	0.000	0.000	Pass
7	2061. 234	354, 436	0.000	0.000	Pass
8	2055, 882	340, 759	0.000	0.000	Pass
9	2065, 602	350, 317	0.000	0.000	Pass
10	2054 515	356 053	0.000	0.000	Pacc

```
202404_AFAS-P Collar(UDL-1).txt
 INCC 6.23.2.9 HM32 Not Validated
                   6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.04.05 15:
Results file name: 445P3710.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
0C tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
                                                                                                                                                15:37:10
                  Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                 1. 500
64. 000
1720
50. 0000
                                                                                                                 0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
                    Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
Number of cycles:
Count time (sec):
                                                                                                                     14. 202
0. 007
0. 000
0. 000
0. 000
                                                                                                                      10
60. 000
Summed raw data
Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
                                                                                                                                                        372646
162580
Summed multiplicity distributions
                                 R+A sums
                                                                           A sums
Messages
Normalization test -- data quality is inadequate
```

Results

2041. 404 +-350. 352 +-0. 000 +-Singles: Doubles: Triples: 2. 815 1. 210 0. 000

(1)

202404_AFAS-P_Bottom Fork.txt

INCC 5.1.2

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS 24. 04. 05 15:48:33 445P4833. VER PWR BF XXXX Pu 0.000 Verification Review disk file On Sample method Measured Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 24 04 05 0.0000 0.0000 0.0000 24.04.05 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0 0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

Passive error messages

No passive calibration curve calibration No known alpha calibration

Normalization results for reference source: H4-694

1,000 +- 0,000 344,821 +- 0,523 3: 350,352 +- 1,210 1: 0,984 +- 0,004 1: 1,000 +- 0,000 data quality is inadequate. 0,345 0,300 0,000 Current normalization constant:
 Cf252 expected doubles rate:
 Cf252 measured doubles rate:
 Doubles rate expected/measured:
 New normalization constant:
 Normalization test - Measured percent precision:
 Required percent precision:
Repeat measurement for at least:

Cycle raw data

Count time (sec): 60.000 Singles 123313 123028 122513 123362 122445 124166 123629 123614 123690 123384 Scaler2 QC Tests O Pass O Pass R+A 37339 37202 A 16089 16335 16131 16416 Cycle Scaler1 36801 37649 Pass Pass O Pass O Pass O Pass O Pass O Pass O Pass 36950 37357 37493 15820 16498 16187 36979 37296 37580 16229 16433 16442 8 9 10

Cycle DTC rate data

(2)

202404_AFAS-P_Bottom Fork.txt

Results

180. 228 +-2. 482 +-0. 000 +-126. 682 +-0. 000 +-1. 367 0. 201 0. 000 0. 283 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4 5	Singles 10802 10871 11519 10776 10797 10677	R+A 241 301 533 255 261 273	A 132 109 296 115 127 121	Scaler1 7593 7622 7624 7517 7631 7507	Scaler2 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass
					0	
5					0	Pass
6					0	Pass
7	10699	242	107	7645	0	Pass
8	10721	266	137	7604	0	Pass
9	10623	259	142	7588	0	Pass
10	10652	259	115	7678	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	180, 033	1. 817	0, 000	0.000	Pass
2	181, 183	3, 200	0.000	0.000	Pass
3	191, 983	3, 950	0.000	0.000	Pass
4	179, 600	2, 333	0.000	0.000	Pass
5	179. 950	2. 233	0.000	0.000	Pass
6	177. 950	2, 533	0.000	0.000	Pass
7	178. 317	2. 250	0.000	0.000	Pass
8	178, 683	2. 150	0.000	0.000	Pass
9	177, 050	1. 950	0.000	0.000	Pass
10	177.533	2 400	0.000	0.000	Pacc

202404_AFAS-P Bottom Fork(UDL-1).txt 202404_AFAS-P Bottom Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6. 23. 2. 9 HM32 NOT VAIIDATED

Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24. 04. 05 15:
Results file name: 44595148. NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 124. 663 125. 202 0. 996 1. 000 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 7665 7589 7602 7616 Scaler2 0 0 QC Tests Cycle Α Scaler1 67 81 72 54 65 54 67 65 58 61 Pass Pass Pass Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 7394 7649 7561 7681 7663 7662 8 9 10 149 133 129 Cycle DTC rate data Singles 126.148 124.882 125.098 125.332 121.632 125.882 124.415 126.415 126.098 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Doubles 1. 367 0. 633 1. 133 1. 350 1. 217 1. 400 QC Tests Pass Pass Cvcle Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 602 0. 000 0. 000 0. 000 0. 000 Pass Pass Pass Pass 350 400 Pass 60.000 8 9 10 Pass 1. 250 1. 133 Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 76082 1378 644 Summed multiplicity distributions R+A sums A sums Messages Normalization test -- data quality is inadequate Results 125. 202 +-1. 223 +-0. 000 +-Singles: Doubles: Triples: 0. 451 0. 073 0. 000 (1) (2) 202405_AFAS-B_Top Fork.txt 202405_AFAS-B_Top Fork.txt INCC 5.1.2 Results Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory charge code: 76. 168 +0. 390 +0. 000 +130. 315 +0. 000 +-0. 325 0. 030 0. 000 0. 404 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2: JM2G AFAS Detector Id.
Electronics id.
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id.
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Passive cycle raw data 24. 05. 01 09:35:02 451J3502. VER Singles 4495 4492 4608 4583 4649 4609 4535 4488 Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 7925 7733 7747 7697 7785 7869 7814 7859 7852 7908 Cvcle QC Tests 20 16 19 13 27 15 18 22 21 26 Pass Pass BWR TF XXXX Pu 0.000 Verification Review disk file On 49 45 38 42 52 41 45 37 41 Sample method Measured 9 10 4617 4625 Passive cycle rate data Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Singles 74, 917 74, 867 76, 800 76, 383 77, 483 76, 817 75, 583 74, 800 76, 950 77, 083 Doubles 0. 483 0. 483 0. 317 0. 483 0. 417 0. 433 0. 450 0. 250 0. 333 0. 250 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 QC Tests Pass Pass Isotopics id: Cvcle Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.05.01 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 Pass Pass Pass Pass Pass Pass Pass 0.0000 0.0000 0.0000 24.05.01 10 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0080 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Passive error messages

Known alpha analysis error

Normalization constant:

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0 0000

0.000 0.000 0.000

1 0000 +-

0.000 +-0.000 +-0.000 +-0.000 0.000

> 10 60

INCC 5.1.2 Results

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Isspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    2468. 041 +-
526. 483 +-
0. 000 +-
9. 047 +-
0. 000 +-
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         Singles:
Doubles:
Triples:
Scaler 1:
Scaler 2:
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 2. 565
1. 357
0. 000
0. 139
0. 000
                                                                                                                                                                                                                                                                                                                                                                                           Passive cycle raw data
                                                                                                24. 05. 01 09:48:06
451J4806. VER
                                                                                                                                                                                                                                                                                                                                                                                                                        Singles
148118
147513
147940
148526
148257
148083
147038
147885
148118
148760
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            Scaler2
0
0
0
0
0
0
0
0
0
                                                                                                                                                                                                                                                                                                                                                                                           Cycle
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               QC Tests
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     Scaler1
                                                                                                                                                                                                                                                                                                                                                                                                                                                                          R+A
55018
54844
54670
55168
55422
55007
54773
54639
55123
55510
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  23579
23307
23418
23512
23345
23417
23149
23530
23524
24003
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    508
540
562
520
563
568
495
559
547
566
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               BWR COLLAR
                                                                                                Pu
0.000
Verification
Review disk file
                                                                                                                                                                                                                                                                                                                                                                                                   8
9
10
                                                                                                On
Sample method
Measured
                                                                                                                                                                                                                                                                                                                                                                                           Passive cycle rate data
                                                                                              Singles
2469. 612
2459. 521
2466. 643
2476. 417
2471. 930
2469. 028
2451. 598
2465. 725
2469. 612
2480. 320
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        Doubles 524. 815 526. 447 521. 692 528. 439 535. 466 527. 335 527. 897 519. 305 527. 486 525. 953
                                        Isotopics id:
                                                                                                                                                                                                                                                                                                                                                                                           Cycle
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   QC Tests
     Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu date:
Am date:
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        Mass
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    Pass
Pass
                                                                                                                                                                                          0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.05.01
                                                                                                                                                   0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                             0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     Pass
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     Pass
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   Pass
Pass
Pass
Pass
Pass
Pass
Pass
                                                                                                                                                                                                                                                                                                                                                                                                       567
                                                                                                                                                     0.0000
                                                                                                                                                                                                                                               0.0000
                                                                                                                                                                                                                                                                                                                                                                                                   8
9
10
                                                                                                                                                                                            0.0000
24.05.01
                                                                                                    1. 50
64. 00
64. 00
1720
50. 0000
0. 1970
160. 0000
0. 6419
0. 1030
0. 0000
0. 0001
0. 0001
                            Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                               1.0000 +-
   Normalization constant:
                                                                                                                                                                0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                              0.000 +
0.000 +
0.000 +
0.000 +
0.000
                                                                                                                                                                 0.000
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                            10
60
Passive error messages
Known alpha analysis error
                                                                                                                                                                    (1)
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               (2)
```

202405_AFAS-B_Bottom Fork.txt

INCC 5.1.2 Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
I tem id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Facility: JM2G AFAS 24. 05. 01 10:01:34 451K0134. VER

BWR BF XXXX Pu 0.000 Verification Review disk file On Sample method Measured

Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 24 05 01 0.0000 0.0000 0.0000 24.05.01

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0.0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

10 60

Passive error messages

No known alpha calibration

202405_AFAS-B_Bottom Fork.txt

Results 135. 203 +-1. 448 +-0. 000 +-4. 483 +-0. 000 +-0. 468 0. 081 0. 000 0. 083 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2: Passive cycle raw data

Scaler2 0 0 0 0 0 0 0 0 0 Scaler 1 255 282 286 257 296 282 253 263 258 258 Singles 8048 Cvcle QC Tests 76 70 77 86 69 78 77 70 83 71 Pass Pass 8049 159 152 172 187 161 143 167 151 166 Pass Pass Pass Pass Pass Pass 8102 8163 8228 7917 8122 8178 8159 8156 9 10 Pass Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	134. 133	1. 533	0.000	0.000	Pass
2	134, 150	1, 483	0.000	0.000	Pass
3	135, 033	1, 250	0.000	0.000	Pass
4	136, 050	1, 433	0.000	0.000	Pass
5	137, 133	1. 967	0.000	0.000	Pass
6	131, 950	1. 383	0.000	0.000	Pass
7	135, 367	1, 100	0.000	0.000	Pass
8	136, 300	1.617	0.000	0.000	Pass
9	135, 983	1, 133	0.000	0.000	Pass
10	135, 933	1. 583	0.000	0.000	Pass

```
202405_AFAS-P_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                 24. 05. 01 10:15:38
451K1538. VER
                                                                                                  PWR TF
                                                                                                  Pu
0.000
Verification
Review disk file
                                                                                                 On
Sample method
Measured
     Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                               0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.05.01
0.0000 +-
24.05.01
                                                                                                                                                       0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                   0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                        0.0000
                                                                                                                                                                                                                                                    0.0000
 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:
                                                                                                      1. 50
64. 00
64. 00
1720
50. 0000
0. 0126
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                                                                                                                  1.0000 +-
   Normalization constant:
                                                                                                                                                                   0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                 0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
```

(1)

202405_AFAS-P Top Fork(UDL-1).txt

```
INCC 6.23.2.9 HM32 Not Validated
```

Passive error messages No known alpha calibration

> Facility: PPFF
> Material balance area: XXXX
> Detector type: PASS
> Detector id: AFAS-P TOP
> Electronics id: UDL-1
> Measurement date: 24.05.01 10:
> Results file name: 451K1922.NOR
> Inspection number:
> Measurement option: Normalization
> Data source: IAFA DataZ file
> Detector configuration: Passive
> 00 tests: On
> Accidentals method: Measured
> Inspector name:
> Comment:
> Ending comment: xx Facility: PPFF 10:19:22

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1.500 64.000 1720 50, 0000 0. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 2. 227 0. 000 0. 000 0. 000 0. 000 0. 045 0. 000 0. 000

60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

114. 965 +-1. 103 +-0. 000 +-

(1)

Singles: Doubles: Triples: Scaler 1: Scaler 2: 2. 785 +-0. 000 +-0. 000 +-117. 117 +-0. 000 +-

Passive cycle raw data

Cycle 1 2 3 4 5 6 7 8 9	Singles 149 172 171 165 182 150 176 159 168 179	R+A 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	A 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	Scaler1 7007 6975 7034 6991 6968 7188 7097 7010 6897 7103	Scaler2 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
-------------------------	---	---	---	---	---	---

Passive cycle rate data

Cycle 1 2 3 4 5 6 7 8 9	Singles 2. 483 2. 867 2. 850 2. 750 3. 033 2. 500 2. 933 2. 650 2. 800 2. 983	Doubles 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	QC Tests Pass Pass Pass Pass Pass Pass Pass P
--	---	--	---	---	---

202405_AFAS-P Top Fork(UDL-1).txt

(2)

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Cycle raw data

Count	time (sec):	60.000			
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 6964 7017 7029 6945 7041 7157 7089 6995 6949 7129	R+A 126 127 124 99 110 123 113 122 123 123	A 45 52 59 43 58 53 58 57 54 49	Scaler1 0 0 0 0 0 0 0 0 0	Scaler2 QC Tests

Cycle DTC rate data

Cycle	Singles	Doubles 5 8 1	Triples	QC Tests
1	113. 840	1, 350	0,000	Pass
2	114, 723	1, 250	0.000	Pass
3	114. 923	1. 083	0.000	Pass
4	113. 523	0. 933	0.000	Pass
5	115. 123	0. 867	0.000	Pass
6	117. 057	1. 167	0.000	Pass
7	115. 923	0. 917	0.000	Pass
8	114. 357	1. 083	0.000	Pass
9	113. 590	1. 150	0.000	Pass
10	116. 590	1. 233	0.000	Pass

(2)

202405_AFAS-P_Collar.txt 202405_AFAS-P_Collar.txt

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                    24. 05. 01 10:31:13
451K3113. VER
                                                                                                    PWR COLLAR
                                                                                                    Pu
0.000
Verification
Review disk file
                                                                                                    On
Sample method
Measured
                                                                                                 Isotopics id:
    Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                                                 0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.05.01
0.0000 +-
24.05.01
                                                                                                                                                          0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                             Pu date:
Am241:
Am date:
                                                                                                                                                           0.0000
                                                                                                                                                                                                                                                           0.0000
 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:
                                                                                                         1. 50
64. 00
64. 00
1720
50. 0000
0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
                                                                                                                   1.0000 +-
   Normalization constant:
                                                                                                                                                                       0.0000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                   0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
```

Passive error messages

No passive calibration curve calibration No known alpha calibration

(1)

202405_AFAS-P Collar(UDL-1).txt

INCC 6.23.2.9 HM32 Not Validated

Facility: PPFF

Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.05.01 10:33:57
Results file name: 451K3357.NOR
Inspection number:
Measurement option: Normalization
Data source: IAFA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: XX Facility: PPFF Ending comment: xx

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1.500 64.000 1720 50, 0000 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001

15. 100 0. 002 0. 000 0. 000 0. 000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 0. 169 0. 014 0. 000

60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 1212203 364515 157100

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

2005 591 +-

Results

Singles: Doubles: Triples: Scaler 1: Scaler 2: 2. 496 1. 782 0. 000 0. 197 0. 000

Passive cycle raw data

Cycle Singles Scaler1 36521 36726 36852 35866 36188 35452 36773 36588 121493 121228 121822 120641 120796 120577 121716 121236 121681 120775 15660 15670 15942 15475 15428 15408 15607 15856 16206 15567 1066 1056 9 10 36669 36103

Passive cycle rate data

Mass QC Tests
0.000 Pass
0.000 Pass Singles 2025. 238 2020. 820 2030. 723 2011. 033 2013. 617 2009. 966 2028. 956 2020. 953 2028. 372 2013. 267 Doubles 347, 927 351, 179 348, 745 340, 086 346, 241 334, 299 353, 014 345, 775 341, 289 342, 505 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Cvcle 234567 8 9 10

(2)

202405_AFAS-P Collar(UDL-1).txt

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant:
Cf252 expected doubles rate:
Cf252 measured doubles rate:
Doubles rate expected/measured:
New normalization constant:
Normalization test—
Measured percent precision:
Required percent precision:
Repeat measurement for at least: 1.000 338.500 345.932 0.979 1.000 0.000 0.514 2.098 0.006 0.000 data quality is inadequate. 0.606 0.300 0.000

Cycle raw data

Count time (sec): 60.000

Cycle 1 2 3 4 5 6 7 8	Singles 121414 121376 121555 121331 120753 120651 120504 121824 121185	R+A 36444 36643 36912 36619 36066 36102 35558 36976 36456	A 15844 15983 15829 15749 15538 15511 15514 15558 15993	Scaler1 0 0 0 0 0 0 0 0	Scaler2 OC Tests
	121185 121610			0	0 Pass 0 Pass

Cycle DTC rate data

Cycle 1 2	Singles 2008.821 2008.187	Doubles 343. 572 344. 573	Triples 0.000 0.000	QC Tests Pass Pass
3	2011. 172	351. 628	0.000	Pass
4	2007. 437	348. 075	0.000	Pass
5	1997. 800	342. 370	0.000	Pass
6	1996. 100	343. 420	0.000	Pass
7	1993. 649	334. 297	0.000	Pass
8	2015. 656	357. 216	0.000	Pass
9	2005. 003	341. 287	0.000	Pass
10	2012. 089	352. 879	0.000	Pass

0. 414 0. 069 0. 000 0. 549 0. 000

QC Tests

QC Tests

Pass

Pass

Pass Pass Pass Pass Pass Pass Pass

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                                                                                                                                                                                                                                Results
                                                                                                                                                                                                                                                                                                                                                                                                               Singles:
Doubles:
Triples:
Scaler 1:
Scaler 2:
                                                                                                                                                                                                                                                                                                                                                                                                                                                                    175. 655 +-
1. 998 +-
0. 000 +-
125. 293 +-
0. 000 +-
                                                                                 24. 05. 01 10:51:19
451K5119. VER
                                                                                                                                                                                                                                                                                                                                Passive cycle raw data
                                                                                                                                                                                                                                                                                                                                                                                                R+A
256
220
242
213
232
257
249
265
246
242
                                                                                                                                                                                                                                                                                                                                                                                                                                                                 Scaler1
7534
7733
7445
7547
7480
7379
7453
7645
7501
7459
                                                                                                                                                                                                                                                                                                                                                         Singles
                                                                                                                                                                                                                                                                                                                                Cycle
                                                                                                                                                                                                                                                                                                                                                                                                                                       A
132
118
124
112
112
117
129
125
130
124
                                                                                 PWR BF
                                                                                                                                                                                                                                                                                                                                                                10657
10576
                                                                                 Pu
0.000
Verification
Review disk file
                                                                                                                                                                                                                                                                                                                                                                10486
                                                                                                                                                                                                                                                                                                                                                                10575
10557
                                                                                                                                                                                                                                                                                                                                                                10601
10504
                                                                                 On
Sample method
Measured
                                                                                                                                                                                                                                                                                                                                      9
10
                                                                                                                                                                                                                                                                                                                                                                10521
10365
                                                                                                                                                                                                                                                                                                                                Passive cycle rate data
                                                                              Isotopics id:
    Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                                                                                                                                                                                                                                                              Doubles
2.067
1.700
1.967
1.683
2.000
2.333
2.000
2.333
1.933
1.967
                                                                                                                                                                                                                                                                                                                                                                  Singles
177.617
                                                                                                                                                                                                                                                                                                                                Cvcle
                                                                                                                                                                                                                                                                                                                                                                                                                                                           Triples
                                                                                                                                                           0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.05.01
0.0000 +-
24.05.01
                                                                                                                            0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                       0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                 0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
                                                                                                                                                                                                                                                                                                                                                                  177. 617
176. 267
175. 850
174. 767
176. 250
175. 950
176. 683
175. 067
175. 350
172. 750
                                                                                                                                                                                                                                                                                                                                          234567
                                                 Pu date:
Am241:
Am date:
                                                                                                                             0.0000
                                                                                                                                                                                                         0.0000
                                                                                    1. 50
64. 00
64. 00
1720
50. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                        Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                              1.0000 +-
   Normalization constant:
                                                                                                                                       0.0000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                             0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                         10
60
Passive error messages
No passive calibration curve calibration
No known alpha calibration
                                                                                                                                          (1)
                                                                                                                                                                                                                                                                                                                                                                                                                                                                          (2)
                                                                                    202405_AFAS-P Bottom Fork(UDL-1).txt
INCC 6.23.2.9 HM32 Not Validated
                                                                                                                                                                                                                                                                                                                                                                                                                                     Scaler 1:
Scaler 2:
```

Facility: PPFF
Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24, 05, 01 10::
Results file name: 451K5104.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Facility: PPFF 10:51:04 Ending comment: xx Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1.500 64.000 1720 50, 0000

0. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 618 0. 000 0. 000 0. 000 0. 000 0.065 0.000 0.000 60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

123. 648 +-1. 235 +-0. 000 +-

(1)

202405_AFAS-P Bottom Fork(UDL-1).txt

Normalization results for reference source: H4-694

Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test —-0. 000 0. 201 0. 432 0. 040 0. 000 122. 378 123. 648 0. 990 1. 000 data quality is inadequate.

Cycle raw data

Count time (sec): 60,000 Singles 7454 7457 7606 7463 7647 Scaler2 QC O Pass O Pass O Pass O Pass O Pass O Pass QC Tests Cvcle Α Scaler1 141 130 127 129 154 49 49 61 66 50 66 58 66 76 O Pass O Pass O Pass O Pass O Pass O Pass 7625 7515 7506 7468 7419 119 122 141 140 141 10

Cycle DTC rate data

Cycle	Singles	Doubles	Triples	QC Tests
1	122, 615	1, 533	0,000	Pass
2	122, 665	1, 350	0,000	Pass
3	125, 148	1, 100	0,000	Pass
4	122, 765	1, 050	0,000	Pass
5	125, 832	1, 533	0,000	Pass
6	125, 465	1, 150	0,000	Pass
7	123, 632	0. 933	0.000	Pass
8	123, 482	1. 383	0.000	Pass
9	122, 848	1. 233	0.000	Pass
10	122, 032	1. 083	0.000	Pass

(2)

Cycle

8 9 10

Cycle

234567

8 9 10

Passive cycle raw data

Passive cycle rate data

Singles 72. 433 73. 617 73. 500 74. 917 73. 400 74. 050 75. 850 76. 200 74. 983 71. 800

```
202406_AFAS-B_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                     24. 06. 05 11:03:30
465L0330. VER
                                                                                                     BWR TF
                                                                                                    Pu
0.000
Verification
Review disk file
                                                                                                    On
Sample method
Measured
     Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am date:
                                                                                                  0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.06.05
0.0000 +-
24.06.05
                                                                                                                                                           0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                                                                                                                           0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                             0.0000
                                                                                                                                                                                                                                                             0.0000
 Predlay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                          1. 50
64. 00
64. 00
1720
50. 0000
0. 0080
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                                                                                                                     1.0000 +-
   Normalization constant:
                                                                                                                                                                         0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                    0.000 +-
0.000 +-
0.000 +-
0.000
0.000
                                                                                                                                                                          0.000
0.000
0.000
```

10 60

Number passive cycles: Count time (sec):

Normalization constant:

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

Passive error messages Known alpha analysis error

Passive error messages Known alpha analysis error

202406_AFAS-B_Collar.txt

(1)

INCC 5.1.2				
Facility: Material balance area: Detector type: Detector id: Electronics id: Inventory change code: 1/0 code: Measurement date: Results file name: Inspection number: Item id: Stratum id: Stratum id: Material type: Original declared mass: Measurement option: Data source: Q tests: Error calculation: Accidentals method: Inspector name: Passive comment:	JMOX JM2G AFAS JSR 01 JSR-12 24.06.05 11:17 465L1734.VER BWR COLLAR XXXXX Pu 0.000 Verification Review disk fil On Sample method Measured			
Isotopics id: Isotopics source code: Pu238: Pu239: Pu240: Pu241: Pu242: Pu date: Am241: Am date:	0.0000 +- 0 100.0000 +- 0 0.0000 +- 0 0.0000 +- 0 00.01.01	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000	0.0000 +- 0.0000 +- 100.0000 +- 0.0000 +- 0.0000 +- 24.06.05 0.0000 +- 24.06.05	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000
Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:	1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 160. 0000 0. 6419 0. 1030 0. 0000 0. 00001 0. 00001			

1 0000 +-

0.000 +-0.000 +-0.000 +-0.000 0.000

> 10 60

202406_AFAS-B_Collar.txt

(2)

74. 075 +-0. 380 +-0. 000 +-126. 635 +-0. 000 +-

Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000

> Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests Pass Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

> Doubles 0, 300 0, 417 0, 400 0, 250 0, 433 0, 300 0, 433 0, 450

			20240	6_AFAS-B_Co	llar.txt	
Results						
		Doul		2409. 878 514. 217 0. 000 9. 028 0. 000	+- 1.9 +- 0.0 +- 0.1	57 100 09
Passive	cycle raw da	ata				
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 144523 144905 144753 143839 144222 144867 145039 144219 144377 144624	R+A 53312 53752 53092 52519 52933 53414 53214 52548 52827 53072	A 22307 22329 22746 22156 22044 22225 22673 22089 22048 22013	Scaler1 515 563 534 556 547 526 521 522 559 574	Scaler2 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Passive	cycle rate o	data				
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 2409. 648 2416. 020 2413. 485 2398. 239 2404. 628 2415. 386 2418. 255 2404. 578 2407. 213 2411. 333	Doub 517. 1 524. 1 506. 2 506. 1 515. 1 520. 1 509. 2 508. 4	550 529 551 830 612 623 807 434	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000	Pass Pass Pass Pass Pass Pass Pass Pass

(1)

0.0000

0.000 0.000 0.000

```
INCC 5.1.2
```

INCC 5.1.2

Facility:

Material balance area:
 Detector type:
 Detector id:
 Electronics id:
 Inventory change code:
 I/O code:
 Measurement date:
 Results file name:
 Isnection number:
 Item id:
 Stratum id:
 Material type:
 Original declared mass:
 Measurement option:
 Data source:
 GC tests:
 Error calculation:
 Accidentals method:
 Inspector name:
 Passive comment: 24. 06. 05 11:33:41 465L3341. VER BWR BF Pu 0.000 Verification Review disk file On Sample method Measured Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id: Isotopics source code: Pu238: Pu239: Pu240: Pu241: Pu date: Am date: 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +24.06.05 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction: 1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000 0.000 0.000 0.000 Number passive cycles: Count time (sec): 10 60

(1)

0.0000 0.0000 0.0000 0.0000 0.0000

0.0000

0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.06.05

0.0000

0. 0000 0. 0000 0. 0000 0. 0000 0. 0000

0.0000

202406_AFAS-P_Top Fork.txt

Facility:

INCC 5.1.2

Passive error messages No known alpha calibration

JM2G AFAS

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
I tem id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 24. 06. 05 13:26:11 465N2611. VER PWR TF XXXX Pu 0.000 Verification Review disk file On Sample method

Sample method Measured

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient Geadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:

1. 50 64. 00 64. 00 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Normalization constant: 1 0000 +-

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

10 60

Passive error messages

No known alpha calibration

Results

132. 170 +-1. 433 +-0. 000 +-4. 447 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Scaler2 0 0 0 0 0 0 0 0 0 Singles Cycle Scaler1 88 67 59 87 59 70 63 73 60 66 287 273 261 240 272 274 272 274 267 248 7863 7858 8006 7967 7891 7927 7916 7922 8031 161 149 148 158 167 8 9 10

Passive cycle rate data

Cycle 1 2 3 4 5 6 7	Singles 132. 017 131. 050 130. 967 133. 433 132. 783 131. 517 132. 117	Doubles 1, 450 1, 450 1, 717 1, 233 1, 500 1, 300 1, 583 1, 567	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000	QC Tests Pass Pass Pass Pass Pass Pass Pass P
9 10	132. 033 133. 850	1. 317 1. 217	0. 000 0. 000	0. 000 0. 000	Pass Pass

202406_AFAS-P_Top Fork.txt

(2)

Results

2. 737 +-0. 000 +-0. 000 +-113. 997 +-0. 000 +-0. 079 0. 000 0. 000 0. 456 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle	Singles	R+A	Α.	Scaler1	Coolors	QC Tests
Cycle		N±N	Α .		Scalerz	
- 1	143	0	0	6811	0	Pass
2	167	0	0	6740	0	Pass
3	160	0	0	6804	0	Pass
4	146	0	0	6845	0	Pass
5	172	0	0	6947	0	Pass
6	161	0	0	6965	0	Pass
7	183	0	0	6739	0	Pass
8	185	0	0	6904	0	Pass
9	149	0	0	6740	0	Pass
10	176	0	0	6903	0	Pass

Passive cycle rate data

QC Tests Pass Pass
Pass Pass Pass Pass Pass Pass
Pass Pass Pass

0.0000

202406_AFAS-P Top Fork(UDL-1).txt 202406_AFAS-P Top Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.06.05 13:
Results file name: 465N3006.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 111. 346 112. 045 0. 994 1. 000 13:30:06 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 6996 6774 Scaler2 0 0 QC Tests Pass Pass Pass Scaler1 Cycle Α 49 64 32 53 47 57 50 48 63 43 102 125 114 126 126 130 111 Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 6801 6747 1. 500 64. 000 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 6886 6939 6951 8 9 10 109 113 6694 Cycle DTC rate data Singles 114.575 110.875 111.325 110.425 112.742 113.625 113.825 109.792 113.725 109.542 Doubles
1. 433
0. 633
1. 550
1. 017
1. 317
1. 150
1. 333
1. 050
0. 767
1. 167 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 QC Tests Pass Pass Cvcle Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 2. 025 0. 000 0. 000 0. 000 0. 000 234567 Pass Pass Pass Pass 10 60. 000 Pass 8 9 10 Pass Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 68442 1191 506 Summed multiplicity distributions A sums R+A sums Messages Normalization test -- data quality is inadequate Results 112.045 +-1.142 +-0.000 +-Singles: Doubles: Triples: 0. 592 0. 091 0. 000 (1) (2) 202406_AFAS-P_Collar.txt 202406_AFAS-P_Collar.txt INCC 5.1.2 Results Facility: Material balance area: Detector type: Detector id: Electronics id: 1968. 762 +-336. 050 +-0. 000 +-17. 335 +-0. 000 +-Singles: Doubles: Triples: 1. 876 1. 821 0. 000 0. 143 0. 000 JM2G AFAS Scaler 1: Scaler 2: 24. 06. 05 13:41:17 465N4117. VER Passive cycle raw data Singles 117942 117849 117761 118435 118300 117854 118807 118078 118339 117691 Scaler2 0 0 0 0 0 0 0 0 0 Cvcle R+A 34793 35280 34832 35800 35080 34754 35345 34973 34741 34531 Scaler1 QC Tests A 14686 14710 14879 15064 15002 14654 14965 14767 14990 14919 1067 1029 1010 1050 1091 1048 1002 8 9 10 Passive cycle rate data Singles 1966. 034 1964. 484 1963. 016 1974. 254 1972. 003 1964. 567 1980. 456 1968. 302 1972. 653 1961. 849 Mass OC Tests
0.000 Pass
0.000 Pass Doubles 335. 345 343. 066 332. 776 345. 836 334. 862 335. 228 339. 899 336. 996 329. 408 327. 088 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Cycle

8 9 10

Inspection number: Inspection number: Stratum id: Material type: Original declared mass: Measurement option: QC tests: Error calculation: Accidentals method: Inspector name: Passive comment:	PWR COLLAR XXXXX Pu 0.000 Verification Review disk fi On Sample method Measured	ile		
Isotopics id: Isotopics source code: Pu238: Pu239: Pu240: Pu241: Pu242: Pu date: Am241: Am date:	0.0000 +- 100.0000 +- 0.0000 +-	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000	0.0000 +- 0.0000 +- 100.0000 +- 0.0000 +- 0.0000 +- 24.06.05 0.0000 +- 24.06.05	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000
Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:	1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001			
Normalization constant:	1.0000 +-	0.0000		

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000

Passive error messages

No passive calibration curve calibration No known alpha calibration

添付-18

```
202406_AFAS-P Collar(UDL-1).txt
 INCC 6.23.2.9 HM32 Not Validated
                   6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24,06.05 13:
Results file name: 465N4842.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
                  Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                    1. 500
64. 000
1720
50. 0000
                                                                                                                    0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
                    Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
Number of cycles:
Count time (sec):
                                                                                                                        13. 737
0. 013
0. 000
0. 000
0. 000
                                                                                                                         10
60. 000
Summed raw data
Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
                                                                                                                                                        600
1182508
351352
149245
Summed multiplicity distributions
                                  R+A sums
                                                                            A sums
Messages
Normalization test -- data quality is inadequate
```

(1)

202406_AFAS-P_Bottom Fork.txt

Singles: Doubles: Triples:

1957. 446 +-337. 061 +-0. 000 +-

1. 745 1. 822 0. 000

INCC 5.1.2

Results

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS 24. 06. 05 13:55:20 465N5520. VER PWR BF XXXX Pu 0.000 Verification Review disk file On Sample method Measured Isotopics id:

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.06.05 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000 24.06.05 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0 0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

Passive error messages

No passive calibration curve calibration No known alpha calibration

Normalization results for reference source: H4-694

330.073 + 0.501 2: 337.061 + 1.822 1: 0.979 + 0.005 1.000 + 0.000 data quality is inadequate. 0.540 0.300 0.000 Current normalization constant:
 Cf252 expected doubles rate:
 Cf252 measured doubles rate:
 Doubles rate expected/measured:
 New normalization constant:
 Normalization test - Measured percent precision:
 Required percent precision:
Repeat measurement for at least:

Cycle raw data

Count time (sec): 60.000 Cycle Singles 1 118185 2 118449 R+A 35620 35549 34808 35287 35253 Scaler2 QC Tests O Pass O Pass Scaler1 A 14824 14907 14656 15109 15157 14904 14802 14686 15067 15133 117717 Pass Pass 118894 118350 117891 118129 35253 34424 34872 35045 35230 35264 Pass Pass Pass 118124 118274 118495 0 Pass 0 Pass 0 Pass 8 9 10

Cycle DTC rate data

Singles 1956. 349 1960. 750 1948. 546 1968. 170 1959. 100 1951. 447 1955. 415 1955. 332 1957. 833 1961. 517 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Doubles 346. 823 344. 255 336. 081 336. 517 335. 149 325. 541 334. 715 339. 534 336. 266 335. 733 QC Tests Pass Pass Cvcle 234567 Pass Pass Pass Pass Pass 8 9 10 Pass Pass Pass

(2)

202406_AFAS-P_Bottom Fork.txt

Results

171. 620 +-2. 055 +-0. 000 +-122. 457 +-0. 000 +-0. 623 0. 073 0. 000 0. 646 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4 5	Singles 10305 10185 10355 10288 10320	R+A 252 237 244 218 255	A 117 114 109 110 126	Scaler1 7262 7281 7375 7348 7142	Scaler2 0 0 0 0	QC Tests Pass Pass Pass Pass Pass
6 7 8 9 10	10158 10372 10335 10528 10126	246 219 270 261 226	123 103 122 147 124	7357 7338 7602 7464 7305	0 0 0 0	Pass Pass Pass Pass Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	171, 750	2, 250	0.000	0.000	Pass
2	169, 750	2, 050	0.000	0.000	Pass
3	172, 583	2, 250	0.000	0.000	Pass
4	171, 467	1, 800	0.000	0.000	Pass
5	172, 000	2, 150	0.000	0.000	Pass
6	169, 300	2, 050	0.000	0.000	Pass
7	172, 867	1, 933	0.000	0.000	Pass
8	172, 250	2, 467	0.000	0.000	Pass
9	175, 467	1, 900	0.000	0.000	Pass
10	168. 767	1. 700	0.000	0.000	Pass

202406_AFAS-P Bottom Fork(UDL-1).txt 202406_AFAS-P Bottom Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6. 23. 2. 9 HM32 NOT VAIIDATED

Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24. 06. 05 14:
Results file name: 46500103. NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 119. 331 121. 247 0. 984 1. 000 14:01:03 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 7260 Scaler2 0 0 QC Tests Cycle Α Scaler1 58 56 65 60 50 49 58 64 53 55 Pass 7381 7334 7155 7353 7346 7606 7467 Pass Pass Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 8 9 10 123 102 125 7292 7433 Cycle DTC rate data Singles 119. 535 121. 552 120. 768 117. 785 121. 085 120. 968 125. 302 122. 985 120. 068 122. 418 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Doubles
1. 050
1. 150
1. 083
1. 300
1. 467
1. 417
1. 317
0. 983
0. 817
1. 167 QC Tests Pass Pass Cvcle Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 465 0. 000 0. 000 0. 000 0. 000 234567 Pass Pass Pass Pass Pass 60.000 8 9 10 Pass Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 568 Summed multiplicity distributions R+A sums A sums Messages Normalization test -- data quality is inadequate Results 121. 247 +-1. 175 +-0. 000 +-Singles: Doubles: Triples: 0. 647 0. 064 0. 000 (1) (2) 202407_AFAS-B_Top Fork.txt 202407_AFAS-B_Top Fork.txt INCC 5.1.2 Results INCC 5.1.2

Facility:

Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error cal culation:
Accidentals method:
Inspector name:
Passive comment: 71. 913 +-0. 282 +-0. 000 +-124. 983 +-0. 000 +-0. 428 0. 029 0. 000 0. 412 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2: JM2G AFAS Passive cycle raw data 24. 07. 03 09:24:17 473J2417. VER Singles 4379 4422 4328 4390 4300 4179 4185 4299 Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 7516 7423 7555 7548 7607 7339 7446 7548 7524 7484 Cvcle QC Tests Pass Pass BWR TF XXXX Pu 0.000 Verification Review disk file On 43 28 45 34 39 30 47 22 27 22 27 15 14 20 27 20 21 Sample method Measured 9 10 4306 4360 Passive cycle rate data Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Singles 72. 983 73. 700 72. 133 73. 167 71. 667 69. 650 69. 750 71. 650 71. 767 72. 667 Doubles 0. 350 0. 267 0. 100 0. 300 0. 317 0. 417 0. 333 0. 317 0. 250 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 QC Tests Pass Pass Isotopics id: Cvcle Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.07.03 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 Pass Pass Pass Pass Pass Pass Pass 0.0000 0.0000 .0000 0.0000 24.07.03 10 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0080 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Normalization constant: 1 0000 +-0 0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

Passive error messages

Known alpha analysis error

10 60

```
INCC 5.1 2
```

INCC 5. 1. 2	Results			
Facility: JMOX Material balance area: JM26 Detector type: AFAS Detector id: JSR_01 Electronics id: JSR_12 Inventory change code: 1/0 code:	Passive cycle raw d	Singles: Doubles: Triples: Scaler 1: Scaler 2:	2361. 763 - 504. 541 - 0. 000 - 8. 587 - 0. 000 -	+- 2.314 +- 0.000 +- 0.131
Measurement date: 24,07.03 09:40:21 Results file name: 473,4021.VER Inspection number: Item id: BWR COLLAR Stratum id: XXXX Material type: Pu Original declared mass: 0.000 Measurement option: Verification Data source: Review disk file OC tests: On Error calculation: Sample method Accidentals method: Measured Inspector name: Passive comment:	Cycle Singles 1 141404 2 141513 3 141333 4 142348 5 141541 6 141212 7 141871 8 142557 9 141788 10 140954 Passive cycle rate	R+A A 51333 21366 51475 21735 51399 21211 51614 21418 51911 21531 51487 21051 52325 21457 52433 21729 51851 21543 51068 21641	Scaler1 552 486 472 498 516 524 524 547 516 517	Scaler2 OC Tests
Isotopics id: Default Isotopics source code: 0D	Cycle Singles 1 2357. 625 2 2359. 443 3 2356. 441 4 2373. 370 5 2359. 910 6 2354. 423 7 2365. 414 8 2376. 856 9 2364. 030 10 2350. 119	Doubles 500, 206 496, 418 503, 895 504, 034 507, 101 508, 034 516, 116 512, 515 505, 900 491, 190	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass QC Tests 0.000 Pass
Predelay: 1.50 Gate length: 64.00 2nd gate length: 64.00 High voltage: 1720 Die away time: 50.0000 Efficiency: 0.1970 Multiplicity deadtime: 160.0000 Coefficient A deadtime: 0.6419 Coefficient B deadtime: 0.1030 Coefficient C deadtime: 0.0000 Doubles gate fraction: 0.0001 Triples gate fraction: 0.0001	10 2350.119	491. 190	0.000	0.000 Pass
Normalization constant: 1.0000 + 0.0000				
Passive singles bkgrnd: 0.000 +- 0.000 Passive doubles bkgrnd: 0.000 +- 0.000 Passive triples bkgrnd: 0.000 +- 0.000 Passive scaler! bkgrnd: 0.000 0.000 Passive scaler2 bkgrnd: 0.000				
Number passive cycles: 10 Count time (sec): 60				
Passive error messages				
Known alpha analysis error				

202407_AFAS-B_Bottom Fork.txt

(1)

Results

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 129. 123 +-1. 455 +-0. 000 +-4. 287 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2: 0. 358 0. 047 0. 000 0. 125 0. 000 JM2G AFAS JSR_02 JSR-12 Passive cycle raw data 24. 07. 03 09:57:32 473J5732. VER Singles 7830 7704 7639 7735 7806 7776 7702 7664 7809 7809 Scaler1 290 252 263 206 281 249 241 254 276 Cvcle R+A 138 148 153 143 163 160 135 149 150 155 65 60 57 49 65 62 56 69 70 68

Scaler2 0 0 0 0 0 0 0 0 0 BWR BF XXXXX Pu 0.000 Verification Review disk file On Sample method Measured 8 9 10 Passive cycle rate data Cycle

Singles 130, 500 128, 400 127, 317 128, 917 130, 100 129, 600 128, 367 127, 733 130, 150 Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Mass OC Tests
0.000 Pass
0.000 Pass Doubles
1. 217
1. 467
1. 600
1. 567
1. 633
1. 633
1. 317
1. 333
1. 333
1. 450 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.07.03 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000

0.0000 0.0000 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 1. 0000 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:

1.0000 +-Normalization constant: 0.0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

(1)

10 60

Passive error messages No known alpha calibration

INCC 5.1.2

(2)

202407_AFAS-B_Bottom Fork.txt

Results

Cycle

8 9 10

Cycle

8 9 10

Passive cycle raw data

Singles

Singles 3.083 2.633

2. 633 2. 900 2. 933 2. 433 3. 000 3. 117 2. 950 2. 750

Passive cycle rate data

```
202407_AFAS-P_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                        24. 07. 03 10:12:37
473K1237. VER
                                                                                        PWR TF
                                                                                        Pu
0.000
Verification
Review disk file
                                                                                        On
Sample method
Measured
                                                                                     Default

0D

0.0000 +-

0.0000 +-

100.0000 +-

0.0000 +-

0.0000 +-

00.01.01

0.0000 +-

00.01.01
                                     Isotopics id:
    Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                           0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.07.03
                                                                                                                                       0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                     Pu date:
Am241:
Am date:
                                                                                                                                        0.0000
                                                                                                                                                                                                                           0.0000
                                                                                                                                                                            0.0000
24.07.03
                                                                                            1. 50
64. 00
64. 00
1720
50. 0000
0. 0126
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                          Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                      1.0000 +-
   Normalization constant:
                                                                                                                                                   0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                     0.000 +
0.000 +
0.000 +
0.000 +
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                  10
60
Passive error messages
No known alpha calibration
                                                                                                                                                      (1)
```

202407_AFAS-P Top Fork(UDL-1).txt

```
INCC 6.23.2.9 HM32 Not Validated
```

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.07.03 10:
Results file name: 473K1554.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
0 C tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Facility: PPFF 10:15:54 Ending comment: xx 1.500 64.000

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1720 50, 0000 0. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 2. 323 0. 000 0. 000 0. 000 0. 000

0. 086 0. 000 0. 000

60.000 Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 67630 1136 480

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

110. 393 +-1. 093 +-0. 000 +-

(1)

202407_AFAS-P Top Fork(UDL-1).txt Scaler 1: Scaler 2:

(2)

2. 850 +--0. 002 +-0. 000 +-112. 772 +-0. 000 +-

6779 6819

6752 6761

6639 6759

6907

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests

QC Tests

Pass Pass

Pass Pass Pass Pass Pass Pass

Pass Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles -0.017 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

Normalization results for reference source: H4-694

Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test —-1. 000 109. 142 110. 393 0. 989 1. 000 0. 000 0. 244 0. 356 0. 040 0. 000

data quality is inadequate.

Cycle raw data

Count time (sec): 60,000 Scaler2 QC O Pass O Pass O Pass O Pass O Pass O Pass Singles 6710 QC Tests Cvcle R+A Α Scaler1 93 132 102 123 107 54 63 40 36 46 6630 6805 O Pass O Pass O Pass O Pass O Pass O Pass 6778 6879 6743 6740 6759 58 42 54 39 48 10

Cycle DTC rate data

Cycle	Singles	Doubles	Triples	QC Tests
1	109, 510	0. 650	0.000	Pass
2	110. 993	1. 150	0.000	Pass
3	110. 793	1. 033	0.000	Pass
4	108. 177	1. 450	0.000	Pass
5	111. 093	1. 017	0.000	Pass
6	110. 643	0. 817	0.000	Pass
7	112. 327	1. 433	0.000	Pass
8	110.060	1. 217	0.000	Pass
9	110.010	1. 117	0.000	Pass
10	110. 327	1. 050	0.000	Pass

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                                                                                                                                                                                                                                                           Results
                                                                                                                                                                                                                                                                                                                                                                                                                                                 Singles:
Doubles:
Triples:
Scaler 1:
Scaler 2:
                                                                                        24. 07. 03 10:30:39
473K3039. VER
                                                                                                                                                                                                                                                                                                                                                           Passive cycle raw data
                                                                                                                                                                                                                                                                                                                                                           Cycle
                                                                                                                                                                                                                                                                                                                                                                                      Singles
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         Scaler1
                                                                                        PWR COLLAR
                                                                                                                                                                                                                                                                                                                                                                                                                                    R+A
34218
33561
33695
33827
34673
34483
34343
34581
                                                                                                                                                                                                                                                                                                                                                                                                                                                                          14446
14024
14384
14261
14463
14315
14583
14513
                                                                                                                                                                                                                                                                                                                                                                                           115181
115466
116247
116409
                                                                                        Pu
0.000
Verification
Review disk file
                                                                                                                                                                                                                                                                                                                                                                                          116118
116685
115946
116074
                                                                                        On
Sample method
Measured
                                                                                                                                                                                                                                                                                                                                                                  9
10
                                                                                                                                                                                                                                                                                                                                                           Passive cycle rate data
                                                                                     Isotopics id:
                                                                                                                                                                                                                                                                                                                                                                                              Singles
1938. 341
1918. 068
1920. 002
1924. 754
1937. 775
1940. 475
1935. 624
1945. 077
1932. 756
1934. 890
     Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                                                                                                                                                                                                                                                                                               Doubles
329. 754
325. 833
322. 064
326. 317
337. 059
336. 359
329. 554
334. 692
333. 756
322. 382
                                                                                                                                                                                                                                                                                                                                                           Cvcle
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 Triples
                                                                                                                                                                         0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.07.03
                                                                                                                                       0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
0. 000
                                                                                                                                                                                                                                                                                                                                                                      234567
                                                     Pu date:
Am241:
Am date:
                                                                                                                                       0.0000
                                                                                                                                                                                                                          0.0000
                                                                                                                                                                           0.0000
24.07.03
                                                                                                                                                                                                                                                                                                                                                                  8
9
10
                                                                                            1. 50
64. 00
64. 00
1720
50. 0000
0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
                          Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                       0. 0001
                                                                                                     1.0000 +-
   Normalization constant:
                                                                                                                                                  0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                     0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
Passive error messages
No passive calibration curve calibration
No known alpha calibration
                                                                                                                                                      (1)
```

202407_AFAS-P Collar(UDL-1).txt

```
INCC 6.23.2.9 HM32 Not Validated
```

Facility: PPFF

Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.07.03 10:48:52
Results file name: 473K4852.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
QC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: XX Facility: PPFF Ending comment: xx 1.500 64.000

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 0. 1620 86. 5000 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001 15. 303 0. 017 0. 000 0. 000 0. 000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 0. 146 0. 004 0. 000 60.000

1720 50, 0000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 1158511 340688 142342

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

1915. 871 +-330. 781 +-0. 000 +-

(1)

202407_AFAS-P Collar(UDL-1).txt

(2)

Scaler 1: Scaler 2:

2. 834 1. 758 0. 000 0. 225 0. 000

QC Tests

Mass QC Tests
0.000 Pass

1001 1047

1088 1040

Normalization results for reference source: H4-694

Current normalization constant:
Cf252 expected doubles rate:
Cf252 measured doubles rate:
Doubles rate expected/measured:
New normalization constant:
Normalization test—
Measured percent precision:
Required percent precision:
Repeat measurement for at least: 0.000 0.491 2.102 0.006 0.000 323. 537 330. 781 0. 978 1. 000 1.000 +- 0.000 data quality is inadequate. 0.636 0.300 0.000

Cycle raw data Count time (sec):

Singles 115747 116043 R+A 34029 34028 33579 34619 Cvcle Scaler1

60.000

Scaler2 QC 0 Pass 0 Pass QC Tests A 14112 14297 14075 14367 14191 14368 14188 14527 14195 14022 115359 116213 115771 115596 Pass Pass O Pass 33952 33892 33636 34015 115424 116176 115731 116451 34236 34702 10

Cycle DTC rate data

Triples QC Tests 0.000 Pass Singles 1914. 135 1919. 070 1907. 666 1921. 904 1914. 535 1911. 618 1908. 750 Doubles 332. 155 329. 053 325. 266 337. 743 329. 553 325. 600 324. 332 325. 001 334. 223 344. 881 Cvcle 1921. 288 1913. 868 1925. 872 10

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                             24. 07. 03 10:58:46
473K5846. VER
                                                                                              PWR BF
                                                                                              Pu
0.000
Verification
Review disk file
                                                                                              On
Sample method
Measured
     Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                           0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.07.03
                                                                                                                                                0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                       0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                 0.0000
                                                                                                                                                                                                                                        0.0000
                                                                                                                                                                                       0.0000
24.07.03
 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:
                                                                                                  1. 50
64. 00
64. 00
1720
50. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                                                                                                             1.0000 +-
   Normalization constant:
                                                                                                                                                           0.0000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                            0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
Passive error messages
No passive calibration curve calibration
No known alpha calibration
```

202407_AFAS-P Bottom Fork(UDL-1).txt

(1)

```
INCC 6.23.2.9 HM32 Not Validated
```

```
Facility: PPFF
Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24.07.03 11:07:38
Results file name: 473.0738.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: XX
Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                                                                                               1.500
64.000
                                                                                                                                                                                          1720
50, 0000
                                                                                                                                                                                                 0. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
```

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 687 0. 000 0. 000 0. 000 0. 000 0.064 0.000 0.000

60.000

Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

117. 857 +-1. 212 +-0. 000 +-

(1)

Results

Singles: Doubles: Triples: Scaler 1: Scaler 2: 170. 633 +-2. 220 +-0. 000 +-119. 815 +-0. 000 +-0. 559 0. 106 0. 000 0. 625 0. 000

Passive cycle raw data

Cycle 1 2 3 4 5 6 7 8 9	Singles 10119 10346 10265 10248 10394 10022 10241 10211 10241 10293	R+A 221 258 226 237 235 235 248 253 269 205	99 112 87 124 103 104 92 116 107	Scaler1 7302 7062 7215 7067 7393 7056 7200 7265 7253 7076	Scaler2 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
-------------------------	---	---	--	---	---	---

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	168. 650	2. 033	0.000	0.000	Pass
2	172. 433	2. 433	0.000	0.000	Pass
3	171. 083	2. 317	0.000	0.000	Pass
4	170. 800	1. 883	0.000	0.000	Pass
5	173. 233	2. 200	0.000	0.000	Pass
6	167. 033	2. 183	0.000	0.000	Pass
7	170. 683	2. 600	0.000	0.000	Pass
8	170. 183	2. 283	0.000	0.000	Pass
9	170. 683	2. 700	0.000	0.000	Pass
10	171. 550	1. 567	0.000	0.000	Pass

(2)

202407_AFAS-P Bottom Fork(UDL-1).txt

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Cycle raw data

Count	time (sec):	60.000			
Cycle 1 2 3 4 5 6 7 8 9	Singles 7173 7175 7147 7175 7331 7160 7086 7216 7225 7038	R+A 127 129 136 131 129 130 134 127 132	A 64 53 68 57 57 51 62 57 54 63	Scaler1 0 0 0 0 0 0 0 0 0	Scaler2 QC Tests

Cycle DTC rate data

Cycle	Singles	Doubles 5 8 1	Triples	QC Tests
1	117. 863	1, 050	0,000	Pass
2	117, 897	1, 267	0.000	Pass
3	117, 430	1, 133	0.000	Pass
4	117, 897	1, 233	0.000	Pass
5	120, 497	1, 200	0.000	Pass
6	117, 647	1, 317	0.000	Pass
7	116, 413	1, 200	0.000	Pass
8	118, 580	1, 167	0.000	Pass
9	118, 730	1, 300	0.000	Pass
10	115. 613	1. 250	0.000	Pass

Cycle

8 9 10

Cycle

234567

8 9 10

Passive cycle raw data

Passive cycle rate data

Singles 72. 967 71. 367 70. 817 72. 867 70. 200 71. 800 68. 517 70. 500 70. 583 68. 400

```
202408_AFAS-B_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                           24. 08. 02 10:38:00
482K3800. VER
                                                                                                           BWR TF
                                                                                                         Pu
0.000
Verification
Review disk file
On
Sample method
Measured
     Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am date:
                                                                                                       0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.08.02
0.0000 +-
24.08.02
                                                                                                                                                                    0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                                                                                                                                        0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                     0.0000
                                                                                                                                                                                                                                                                          0.0000
 Predlay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                               1. 50
64. 00
64. 00
1720
50. 0000
0. 0080
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                                                                                                                                                                                  0.0000
                                                                                                                           1.0000 +-
   Normalization constant:
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                          0.000 +-
0.000 +-
0.000 +-
0.000
0.000
                                                                                                                                                                                   0.000
0.000
0.000
```

10 60

Number passive cycles: Count time (sec):

Passive error messages Known alpha analysis error

INCC 5.1.2

(2)

202408_AFAS-B_Collar.txt

(1)

11100 0.1.2				
Racility: Material balance area: Detector type: Detector id: Inventory change code: I/O code: Measurement date: Results file name: Inspection number: Stratum id: Material type: Original declared mass: Measurement option: Qt tests: Error calculation: Accidentals method: Inspector name: Passive comment:	JMOX JM26 AFAS JSR 01 JSR-12 24.08.05 12: 485M2457.VER BWR COLLAR XXXX Pu 0.000 Verification Review disk f On Sample method Measured			
Isotopics id: Isotopics source code: Pu238: Pu239: Pu240: Pu241: Pu242: Pu date: Am241: Am date:	Default 0D 0.0000 +- 0.0000 +- 100.0000 +- 0.0000 +- 0.0000 +- 0.0000 +- 0.0000 +- 0.0000 +- 0.0000 +- 0.01.01 0.0000 +- 00.01.01	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000	0.0000 +- 0.0000 +- 100.0000 +- 0.0000 +- 0.0000 +- 24.08.05 0.0000 +- 24.08.05	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000
Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:	1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 0. 0000 0. 6419 0. 1030 0. 0000 0. 0001 0. 0001			
Normalization constant:	1.0000 +-	0.0000		
Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:	0.000 +- 0.000 +- 0.000 +- 0.000 0.000	0.000 0.000 0.000		

Passive error messages Known alpha analysis error 202408_AFAS-B_Collar.txt

70. 802 +-0. 362 +-0. 000 +-122. 203 +-0. 000 +-

Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000

0. 492 0. 024 0. 000 0. 468 0. 000

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests Pass Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles 0. 417 0. 467 0. 283 0. 383 0. 267 0. 450 0. 283 0. 333 0. 433 0. 300

Results						
		Dou Tri Scal	gles: bles: ples: er 1: er 2:	2306. 397 491. 601 0. 000 8. 747 0. 000	+- 2.1 +- 0.0 +- 0.0	33 00 99
Passive	cycle raw da	ata				
Cycle 1 2 3 4 5 6 7 8 9	Singles 139011 137942 138113 138180 138815 138800 138657 137275 138283 138250	R+A 50779 49374 49498 49781 49916 50346 49957 49137 49870 50138	A 20679 20209 20447 20721 20246 20374 20599 20055 20633 20309	Scaler1 524 543 546 530 509 492 533 507 549 515	Scaler2 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Passive	cycle rate o	data				
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 2317, 712 2299, 882 2302, 734 2303, 852 2314, 443 2314, 1807 2288, 757 2305, 570 2305, 019	Doub 502. 486. 484. 485. 500. 495. 485. 485. 487.	414 801 900 050 235 276 027 413	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Pass Pass Pass Pass Pass Pass Pass Pass

```
INCC 5.1.2
```

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 24. 08. 02 11:10:43 482L1043. VER BWR BF Pu 0.000 Verification Review disk file On Sample method Measured Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +24.08.02
0.0000 +24.08.02 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction: 1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000 0.000 0.000 0.000 Number passive cycles: Count time (sec): 10 60 Passive error messages

No known alpha calibration

(1)

202408_AFAS-P_Top Fork.txt

INCC 5.1.2

INCC 5.1.2

Facility:

Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error cal culation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS 24. 08. 02 11:26:47 482L2647. VER PWR TF XXXX Pu 0.000 Verification Review disk file On Sample method Sample method Measured Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.08.02 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 24.08.02 0.0000 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction 1. 50 64. 00 64. 00 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Normalization constant: 1 0000 +-0.0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

Passive error messages

No known alpha calibration

10 60

Results

126. 058 +-1. 287 +-0. 000 +-4. 233 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4 5 6 7 8 9 10	Singles 7674 7559 7626 7612 7630 7479 7648 7494 7485 7428	R+A 163 128 140 144 140 115 135 126 138 136	A 67 58 48 62 63 65 46 60 60 64	Scaler 1 265 237 243 264 268 266 241 246 254 256	Scaler2 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
----------------------------	---	---	---	--	---	---

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	127. 900	1. 600	0.000	0.000	Pass
2	125. 983	1. 167	0.000	0.000	Pass
3	127. 100	1. 533	0.000	0.000	Pass
4	126. 867	1. 367	0.000	0.000	Pass
3	127. 100	1. 533	0.000	0.000	Pass
7	127. 467	1. 483	0. 000	0.000	Pass
8	124. 900	1. 100	0. 000	0.000	Pass
9	124. 750	1. 300	0. 000	0.000	Pass
10	123. 800	1. 200	0. 000	0.000	Pass

(2)

202408_AFAS-P_Top Fork.txt

Results

Singles:	3. 240 +-	0. 073
Doubles:	0. 000 +-	0. 000
Triples:	0. 000 +-	0. 000
Scaler 1:	109. 593 +-	0. 584
Scaler 2:	0. 000 +-	0. 000

Passive cycle raw data

1 4331 40	Cycle law	Jaca				
Cycle 1 2 3 4 5 6 7 8 9 10 Passive	Singles 173 207 200 221 190 183 184 195 188 203	R+A 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	A 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	Scaler1 6738 6467 6427 6604 6424 6576 6542 6640 6626 6712	Scaler2 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Cycle 1	Singles 2.883	Doubles 0.000	1	Triples 0.000	Mass 0.000	QC Tests Pass

Cvcle	Singles	Doubles	Triples	Mass	QC Tests
Cycle					
1	2. 883	0.000	0.000	0.000	Pass
2	3, 450	0.000	0.000	0.000	Pass
3	3, 333	0.000	0.000	0.000	Pass
4	3, 683	0.000	0.000	0.000	Pass
5	3, 167	0.000	0.000	0.000	Pass
6	3, 050	0.000	0.000	0.000	Pass
7	3. 067	0.000	0.000	0.000	Pass
8	3. 250	0.000	0.000	0.000	Pass
9	3, 133	0.000	0.000	0.000	Pass
10	3, 383	0.000	0.000	0.000	Pass

202408_AFAS-P Top Fork(UDL-1).txt 202408_AFAS-P Top Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6.23.2.9 HM32 NOT VAIIDATED

Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.08.02 11:
Results file name: 482L3948.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 106. 813 107. 353 0. 995 1. 000 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 6633 6655 6395 Scaler2 0 0 QC Tests Cycle Α Scaler1 48 52 43 40 45 47 46 48 49 52 Pass 128 92 107 Pass Pass Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 6585 6472 6474 6565 6528 6710 6671 1.500 64.000 1720 50.0000 Pass Pass O Pass O Pass O Pass O Pass O Pass O Pass 106 114 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 8 9 10 109 142 Cycle DTC rate data Singles 108, 423 108, 790 104, 457 107, 623 105, 740 105, 773 107, 290 106, 673 109, 707 109, 057 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 QC Tests Pass Pass Cvcle Doubles 1.145 1. 145 1. 228 0. 778 1. 078 0. 962 0. 945 1. 095 0. 978 0. 962 1. 462 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 2. 127 0. 038 0. 000 0. 000 0. 000 234567 Pass Pass Pass Pass Pass 60.000 8 9 10 Pass Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 65688 1131 470 Summed multiplicity distributions A sums R+A sums Messages Normalization test -- data quality is inadequate Results Singles: Doubles: Triples: 0. 538 0. 068 0. 000 107.353 1. 063 (1) (2) 202408_AFAS-P_Collar.txt 202408_AFAS-P_Collar.txt INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: INCC 5.1.2 Results 1890. 399 +-323. 396 +-0. 000 +-16. 842 +-0. 000 +-Singles: Doubles: Triples: 1. 552 1. 377 0. 000 0. 235 0. 000 JM2G AFAS Scaler 1: Scaler 2: 24. 08. 02 13:27:15 482N2715. VER Passive cycle raw data Singles 113271 113313 113486 113572 113539 113030 113948 113476 112906 113513 Scaler2 0 0 0 0 0 0 0 0 0 PWR COLLAR XXXX Pu 0.000 Verification Review disk file On Cvcle R+A 32897 32889 33035 33527 33354 32932 33326 32831 32832 33190 Scaler1 QC Tests 13929 13738 13419 13645 13744 1048 1027 972 1061 953 979 978 1073 1045 969 13660 13904 13583 13437 13843 Sample method Measured 8 9 10 Passive cycle rate data Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-Isotopics id: Doubles 316. 340 319. 392 327. 147 331. 584 327. 047 321. 409 323. 913 321. 010 323. 460 322. 661 Mass OC Tests
0.000 Pass
0.000 Pass Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Cycle Singles 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 1888. 158 1888. 858 1891. 743 1893. 176 1893. 1/6 1892. 626 1884. 140 1899. 445 1891. 576 1882. 073 1892. 193 00. 01. 01 0. 0000 +-00. 01. 01 24 08 02 0.0000 0.0000 0.0000 8 9 10 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

Passive error messages

No passive calibration curve calibration No known alpha calibration

Normalization constant:

0 0000

0.000 0.000 0.000

1 0000 +-

```
202408_AFAS-P Collar(UDL-1).txt
 INCC 6.23.2.9 HM32 Not Validated
                   6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.08.02 13:
Results file name: 482N3744.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
                                                                                                                                               13:37:44
                  Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                1. 500
64. 000
1720
50. 0000
0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
                   Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
Number of cycles:
Count time (sec):
                                                                                                                   14. 062
0. 038
0. 000
0. 000
0. 000
                                                                                                                    10
60. 000
Summed raw data
Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
                                                                                                                                                  600
1136254
332469
138222
Summed multiplicity distributions
                                R+A sums
                                                                         A sums
Messages
Normalization test -- data quality is inadequate
Results
                                                                                                                 Singles:
Doubles:
Triples:
                                                                                                                                                                                                                   2. 536
1. 094
0. 000
                                                                                                                                                       1880.005
```

202408_AFAS-P_Bottom Fork.txt

(1)

323. 919 +-0. 000 +-

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS

INCC 5.1.2

24. 08. 02 13:44:20 482N4420. VER PWR BF XXXX Pu 0.000 Verification Review disk file On Sample method Sample method Measured

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 24 08 02 0.0000 0.0000 0.0000 +-24.08.02

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0.0000 0.000 0.000 0.000

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

Passive error messages

No passive calibration curve calibration No known alpha calibration

Normalization results for reference source: H4-694

Current normalization constant:
 Cf252 expected doubles rate:
 Cf252 measured doubles rate:
 Doubles rate expected/measured:
 New normalization constant:
 Normalization test - Measured percent precision:
 Required percent precision:
Repeat measurement for at least:

Cycle raw data

Count time (sec): 60.000

Cycle	Singles	R+A	Α	Scaler1	Scaler2 QC Tests
1	113396	33047	13787	0	0 Pass
2	114482	33463	13910	0	0 Pass
3	114286	33579	14111	0	0 Pass
4	113249	32879	13599	0	0 Pass
5	113227	32886	13626	0	0 Pass
6	113475	33172	14000	0	0 Pass
7	113600	33646	13983	0	0 Pass
8	113527	33402	13568	0	0 Pass
9	113010	32952	13597	0	0 Pass
10	114002	33443	14041	0	0 Pass

Cycle DTC rate data

Cycle Singles 1 1876-181 2 1894-286 3 1891-019 4 1873-730 5 1873-363 6 1877-498 7 1879-582 8 1878-365 9 1869-745 10 1886-284	Doubles 321.172 326.060 324.642 321.505 321.171 319.704 327.893 330.745 322.755 323.541	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	QC Tests Pass Pass Pass Pass Pass Pass Pass P
--	---	---	---

(2)

202408_AFAS-P_Bottom Fork.txt

Results

166. 265 +-2. 170 +-0. 000 +-117. 542 +-0. 000 +-0. 516 0. 114 0. 000 0. 535 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4 5 6	Singles 9931 9915 10085 10168 9975 9816	R+A 235 214 248 260 253 223	A 105 114 107 125 106 99	Scaler1 7116 7051 7066 7126 7138 7027	Scaler2 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass
	10168	260	125	7126	0	Pass
5	9975	253	106	7138	0	Pass
6	9816	223	99	7027	0	Pass
7	9941	260	86	6809	0	Pass
8	9934	206	98	7023	0	Pass
9	9964	241	108	7165	0	Pass
10	10030	215	105	7004	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	165. 517	2. 167	0.000	0.000	Pass
2	165, 250	1, 667	0.000	0.000	Pass
3	168, 083	2, 350	0.000	0.000	Pass
4	169. 467	2. 250	0.000	0.000	Pass
5	166. 250	2. 450	0.000	0.000	Pass
6	163, 600	2. 067	0.000	0.000	Pass
7	165, 683	2, 900	0.000	0.000	Pass
8	165, 567	1, 800	0.000	0.000	Pass
9	166. 067	2. 217	0.000	0.000	Pass
10	167. 167	1. 833	0.000	0.000	Pass

202408_AFAS-P Bottom Fork(UDL-1).txt 202408_AFAS-P Bottom Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6. 23. 2. 9 HM32 NOT VAIIDATED

Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24. 08. 02 14:
Results file name: 48200122. NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1.000 +-114.467 +-115.915 +-0.988 +-1.000 +-14:01:22 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 7245 7002 7001 Scaler2 0 0 QC Tests Cycle Α Scaler1 45 60 51 56 49 47 46 67 59 Pass Pass Pass Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 6884 7057 7079 7160 7159 6912 105 141 8 9 10 120 117 101 6988 Cycle DTC rate data Singles 119, 187 115, 137 115, 120 113, 170 116, 053 116, 420 117, 770 117, 753 113, 637 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Doubles 1. 495 1. 012 1. 045 1. 045 0. 928 1. 545 0. 845 QC Tests Pass Pass Cvcle Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 563 0. 038 0. 000 0. 000 0. 000 0. 074 0. 033 0. 000 Pass 1. 045 1. 045 0. 928 1. 545 0. 845 0. 928 0. 795 Pass Pass Pass Pass 60.000 8 9 10 Pass 113.637 114.903 Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 70487 1195 531 Summed multiplicity distributions A sums R+A sums Messages Normalization test -- data quality is inadequate Results Singles: Doubles: Triples: 0. 608 0. 087 0. 000 115.915 +-1. 068 (1) (2) 202409_AFAS-B_Top Fork.txt 202409_AFAS-B_Top Fork.txt INCC 5.1.2 Results Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory charge code: 69. 918 +-0. 365 +-0. 000 +-117. 778 +-0. 000 +-0. 190 0. 037 0. 000 0. 458 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2: JM2G AFAS Detector Id.
Electronics id.
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id.
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Passive cycle raw data 24. 09. 24 09:58:57 490J5857. VER Singles 4143 4238 4225 4194 4206 4248 4198 Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 7163 6950 7062 7011 7159 7032 6993 7186 6977 7134 Cvcle QC Tests Pass Pass BWR TF XXXX Pu 0.000 Verification Review disk file On 38 48 45 41 32 44 49 49 33 18 20 18 20 17 20 19 17 25 20 Sample method Measured 9 10 4186 4169 Passive cycle rate data Doubles 0. 333 0. 233 0. 500 0. 417 0. 400 0. 200 0. 417 0. 533 0. 400 0. 217 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-Singles QC Tests Pass Pass Isotopics id: Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Cvcle Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 69.050 70.633 70.417 69.900 70.100 70.800 69.967 69.067 69.767 69.483 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 Pass Pass Pass Pass Pass Pass Pass 00. 01. 01 0. 0000 +-00. 01. 01 24 09 24 0.0000 0.0000 0.0000 10 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0080 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001

Passive error messages

Known alpha analysis error

Normalization constant:

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 1 0000 +-

0.000 +-0.000 +-0.000 +-0.000 0.000

> 10 60

0 0000

0.000 0.000 0.000

```
INCC 5.1.2
                                                                                                     Results
```

11100 0.1.2				Resurts					
Facility: Material balance area: Detector type: Detector id: Electronics id: Inventory change code:	JM2G AFAS JSR_01					Singl Double Triple Scaler Scaler	es: es: 1:	2223. 502 474. 811 0. 000 7. 962 0. 000	+- 0.000 +- 0.140
Measurement date: Results file name: Results file name: Inspection number: Stratum id: Material type: Original declared mass: Measurement option: Data source:	XXXX			Passive Cycle 1 2 3 4 5 6 7	cycle raw days a singles 133702 133510 133565 133435 132988 133234 134066	R+A 47879 47234 47568 47165 46979 47565 47732	A 19099 18929 18949 19205 18358 18717 19168	Scaler1 452 486 522 450 484 457 498	Scaler2 QC Tests
QC tests: Error calculation: Accidentals method: Inspector name:	On Sample method Measured			8 9 10	132773 133570 132782	47034 47000 46861	18818 18549 18745	504 482 442	0 Pass 0 Pass 0 Pass
Passive comment:				Passive	cycle rate	data			
Isotopics id: Isotopics source code; Pu238: Pu239: Pu240: Pu240: Pu241: Pu242: Pu date: Am241: Am date:	0D 0.0000 +- 0.0000 0.0000 +- 0.0000 100.0000 +- 0.0000 0.0000 +- 0.0000 0.0000 +- 0.0000 00.01.01 0.0000 +- 0.0000	0.0000 +- 100.0000 +- 0.0000 +- 0.0000 +- 24.09.24	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000	Cycle 1 2 3 4 5 6 7 8 9	Singles 2229. 164 2225. 962 2226. 879 2224. 711 2217. 255 2221. 358 2235. 235 2213. 670 2226. 962 2213. 820	Double 480. 35- 472. 42- 477. 66- 466. 66- 477. 69- 481. 48- 476. 75- 470. 93- 474. 86- 86- 86- 87- 88- 88- 88- 88- 88- 88- 88- 88- 88	4 5 6 6 6 6 6 0 5 2	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass OC Tests 0.000 Pass
Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient C deadtime: Coefficient C deadtime: Triples gate fraction:	1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 160. 0000 0. 6419 0. 1030 0. 0000 0. 0000 0. 0001			10	2213. 820	469. 26	D.	0.000	0.000 Pass
Normalization constant:	1.0000 +- 0.0000)							
Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:	0.000 +- 0.000 0.000 +- 0.000 0.000 +- 0.000 0.000 0.000								
Number passive cycles: Count time (sec):	10 60								
Passive error messages									

Known alpha analysis error

INCC 5.1.2

(1)

202409_AFAS-B_Bottom Fork.txt 202409_AFAS-B_Bottom Fork.txt Results

Facility:	
Material balance area:	JM2G
Detector type:	AFAS
Detector id:	
Electronics id:	JSR-12
Inventory change code:	
I/O code:	
	24. 09. 24 10:30:48
Results file name:	490K3048. VER
Inspection number:	
Item id:	
Stratum id:	XXXX
Material type:	
Original declared mass:	0.000
Measurement option:	
Data source	Review disk file
QC tests:	0n
Error calculation:	Sample method
Accidentals method:	Measured
Inspector name:	

Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Isotopics id: Default
Isotopics source code: 0D
Pu238: 0.0000 +Pu239: 0.0000 +Pu240: 100.0000 +Pu241: 0.0000 +Pu242: 0.0000 +Pu242: 0.0000 +Pu242: 0.0000 +Amadate: 00.01.01 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +24.09.24
0.0000 +24.09.24 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000

0.0000

0.0000

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 1.0000 +-Normalization constant:

0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000 0.000 0.000 0.000

10 60

Passive error messages No known alpha calibration

121. 247 +-1. 267 +-0. 000 +-4. 075 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2: 0. 510 0. 072 0. 000 0. 089 0. 000

(2)

Passive cycle raw data

	0: 1	D 4		0 1 4		00 T .
Cycle	Singles	R+A	Α	Scaler1	Scaler2	QC Tests
Ţ	7355	150	74	250	0	Pass
2 3	7123	121	68	252	0	Pass
	7222	137	51	258	0	Pass
4	7312	116	42	259	0	Pass
4 5	7243	137	55	226	0	Pass
6	7271	156	60	221	0	Pass
7	7304	123	64	252	0	Pass
8	7342	144	52	257	0	Pass
9	7436	130	60	255	0	Pass
10	7140	122	50	215	0	Pass
Passive	cycle rate	data				
Cycle	Singles	Doubles 5		Triples	Mass	QC Tests
1	122. 583	1. 267		0.000	0.000	Pass
2	118, 717	0. 883		0.000	0.000	Pass

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	122. 583	1. 267	0.000	0.000	Pass
2	118. 717	0. 883	0.000	0.000	Pass
3	120. 367	1. 433	0.000	0.000	Pass
4	121. 867	1. 233	0.000	0.000	Pass
5	120. 717	1. 367	0.000	0.000	Pass
6	121. 183	1. 600	0.000	0.000	Pass
7	121. 733	0. 983	0.000	0.000	Pass
8	122. 367	1. 533	0.000	0.000	Pass
9	123. 933	1. 167	0.000	0.000	Pass
10	119.000	1. 200	0.000	0.000	Pass

Cycle

8 9 10

Cycle

234567

8 9 10

Passive cycle raw data

Singles

128 160

Passive cycle rate data

Singles 2, 783 2, 133 2, 667 2, 267 2, 383 2, 450 2, 733 2, 200 2, 083 2, 067

```
202409_AFAS-P_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                              24. 09. 24 10:45:52
490K4552. VER
                                                                              PWR TF
                                                                              Pu
0.000
Verification
Review disk file
                                                                              On
Sample method
Measured
    Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                            0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.09.24
0.0000 +-
24.09.24
                                                                                                                        0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                 0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                        0.0000
                                                                                                                                                                                                  0.0000
 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:
                                                                                 1. 50
64. 00
64. 00
1720
50. 0000
0. 0126
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                                                                                          1.0000 +-
   Normalization constant:
                                                                                                                                  0.0000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                          0.000 +-
0.000 +-
0.000 +-
0.000
0.000
    Number passive cycles:
Count time (sec):
Passive error messages
No known alpha calibration
                                                                                                                                     (1)
                                                                                      202409_AFAS-P Top Fork(UDL-1).txt
INCC 6.23.2.9 HM32 Not Validated
                                                        Facility: PPFF
```

```
Material balance area: XXXX
Detector type: PASS
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.09.24 11:03:08
Results file name: 490L0308.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
```

Predelay:	1.500
Gate length:	64.000
High voltage:	1720
Die away time:	50.0000
Efficiency:	0.0126
Multiplicity deadtime:	0.0000
Coefficient A deadtime:	0.0000
Coefficient B deadtime:	0.0000
kCoefficient C deadtime:	0.0000
Doubles gate fraction:	0. 0001
Triples gate fraction:	0. 0001
Passiva singles hkarnd:	2 127

Passive singles bkgrnd:	2, 127	+-	0.045
Passive doubles bkgrnd:	0.000	+-	0.000
Passive triples bkgrnd:	0.000	+-	0.000
Passive scaler1 bkgrnd:	0.000		
Passive scaler2 bkgrnd:	0.000		
Number of cycles:	10		
Count time (sec):	60.000		

Summed raw data

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test — data quality is inadequate

Results

 Singles:
 103.755 + 0.448

 Doubles:
 0.993 + 0.068

 Triples:
 0.000 + 0.000

(1)

Scaler 1: 0.000 +- 0.000 Scaler 2: 0.000 +- 0.000

(2)

202409_AFAS-P Top Fork(UDL-1).txt

2. 377 +-0. 000 +-0. 000 +-105. 612 +-0. 000 +-

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 QC Tests Pass Pass

Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

Normalization results for reference source: H4-694

Current normalization constant:	1.000 +- 0.000
Cf252 expected singles rate:	102.829 +- 0.230
Cf252 measured singles rate:	103.755 +- 0.448
Singles rate expected/measured:	0.991 +- 0.040
New normalization constant:	1.000 +- 0.000
Normalization test -- data quality is inadequate.	

Cycle raw data

1 6501 94 48 0 0 Pass 2 6331 101 47 0 0 Pass	Count 1	ime (sec):	60.000			
3 0300 92 49 0 0 7 ass 4 6 244 93 46 0 0 Pass 5 6303 111 44 0 0 Pass 6 6330 113 40 0 0 0 Pass 7 6340 106 37 0 0 Pass 8 6410 96 47 0 0 Pass 9 6238 113 42 0 0 Pass 10 6452 116 39 0 0 Pass	1 2 3 4 5 6 7 8 9	6501 6331 6380 6244 6303 6330 6340 6410 6238	94 101 92 93 111 113 106 96 113	48 47 49 46 44 40 37 47 42	Scaler1 0 0 0 0 0 0 0 0 0	O Pass

Cycle DTC rate data

Cycle	Singles	Doubles 5 8 1	Triples	QC Tests
1	106, 223	0. 767	0,000	Pass
2	103, 390	0.900	0,000	Pass
3	104, 207	0. 717	0,000	Pass
4	101, 940	0. 783	0,000	Pass
5	102, 923	1, 117	0,000	Pass
6	103, 373	1, 217	0,000	Pass
7	103, 540	1, 150	0,000	Pass
8	104, 707	0. 817	0,000	Pass
9	101, 840	1, 183	0,000	Pass
10	105. 407	1. 283	0.000	Pass

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                                                                                                                                                                                                                                                                        Results
                                                                                                                                                                                                                                                                                                                                                                                                                                                                 Singles:
Doubles:
Triples:
Scaler 1:
Scaler 2:
                                                                                           24. 09. 24 11:03:26
490L0326. VER
                                                                                                                                                                                                                                                                                                                                                                        Passive cycle raw data
                                                                                                                                                                                                                                                                                                                                                                                                     Singles
                                                                                                                                                                                                                                                                                                                                                                        Cycle
                                                                                                                                                                                                                                                                                                                                                                                                                                                  R+A
31019
31555
31448
31558
31259
31491
31951
31173
31518
31570
                                                                                           PWR COLLAR
                                                                                                                                                                                                                                                                                                                                                                                                          108438
109276
109153
109254
109352
109407
                                                                                           Pu
0.000
Verification
Review disk file
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           12898
12720
12807
12823
13045
12793
                                                                                          On
Sample method
Measured
                                                                                                                                                                                                                                                                                                                                                                                                          109852
108997
                                                                                                                                                                                                                                                                                                                                                                                9
10
                                                                                                                                                                                                                                                                                                                                                                        Passive cycle rate data
                                                                                         Isotopics id:
    Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                                                                                                                                                                                                                                                                                                                Doubles
312, 412
310, 929
312, 480
315, 549
306, 210
313, 047
319, 269
306, 026
308, 078
313, 148
                                                                                                                                                                                                                                                                                                                                                                                                             Singles
1807.582
                                                                                                                                                                                                                                                                                                                                                                        Cvcle
                                                                                                                                                                               0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.09.24
0.0000 +-
24.09.24
                                                                                                                                            0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                 0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                                                                                                                                                           1807. 582
1821. 553
1819. 503
1821. 187
1822. 821
1823. 738
1831. 157
1816. 902
1829. 556
1824. 488
                                                                                                                                                                                                                                                                                                                                                                                   234567
                                                       Pu date:
Am241:
Am date:
                                                                                                                                             0.0000
                                                                                                                                                                                                                                   0.0000
 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:
                                                                                               1. 50
64. 00
64. 00
1720
50. 0000
0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
                                                                                                         1.0000 +-
   Normalization constant:
                                                                                                                                                       0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                         0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
Passive error messages
No passive calibration curve calibration
No known alpha calibration
```

202409_AFAS-P Collar(UDL-1).txt

(1)

INCC 6.23.2.9 HM32 Not Validated

Facility: PPFF

Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.09.24 11:19:26
Results file name: 490L1926.NOR
Inspection number:
Measurement option: Normalization
Data source: IAFA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: XX Facility: PPFF Ending comment: xx 1.500 64.000 1720 50, 0000

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001 13. 117 0. 005 0. 000 0. 000 0. 000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 0. 128 0. 009 0. 000 60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 1091881 314524 126613

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

1806. 971 +-313. 377 +-0. 000 +-

202409_AFAS-P Collar(UDL-1).txt

(2)

1821. 849 +-311. 715 +-0. 000 +-16. 187 +-0. 000 +-

Scaler1

Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000

Mass QC Tests
0.000 Pass

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant:
Cf252 expected doubles rate:
Cf252 measured doubles rate:
Doubles rate expected/measured:
New normalization constant:
Normalization test—
Measured percent precision:
Required percent precision:
Repeat measurement for at least: 0. 000 0. 463 1. 439 0. 005 0. 000 304. 827 313. 377 0. 973 1. 000 data quality is inadequate. 0.459 0.300 0.000

Cycle raw data

Count time (sec):

oount	LTIIIC (300).	00.000			
Cycle	Singles	R+A	Α	Scaler1	Scaler2 QC Tests
- 1	108863	31035	12429	0	0 Pass
2	109125	31205	12736	0	0 Pass
3	108430	31182	12632	0	0 Pass
4	109496	31764	12865	0	0 Pass
5	109045	31464	12771	0	0 Pass
6	109280	31612	12646	0	0 Pass
7	109198	31299	12569	0	0 Pass
8	109607	31680	12537	0	0 Pass
9	109768	31963	12677	0	0 Pass
10	109069	31320	12751	0	0 Pass

Cycle DTC rate data

Cycle	Singles	Doubles	Triples	QC Tests
1	1801. 551	310. 290	0.000	Pass
2	1805. 919	308. 005	0.000	Pass
3	1794. 332	309. 355	0.000	Pass
4	1812. 105	315. 177	0.000	Pass
5	1804. 586	311. 741	0.000	Pass
6	1808. 503	316. 294	0.000	Pass
7	1807. 136	312. 358	0.000	Pass
8	1813. 955	319. 247	0.000	Pass
9	1816. 639	321. 632	0.000	Pass
10	1804. 986	309. 673	0.000	Pass

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                                                                                                                                                                                                                                         Results
                                                                                    24. 09. 24 11:19:31
490L1931. VER
                                                                                                                                                                                                                                                                                                                                         Passive cycle raw data
                                                                                                                                                                                                                                                                                                                                         Cycle
                                                                                    PWR BF
                                                                                    Pu
0.000
Verification
Review disk file
                                                                                   On
Sample method
Measured
                                                                                                                                                                                                                                                                                                                                                9
10
                                                                                 Isotopics id:
    Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.09.24
0.0000 +-
24.09.24
                                                                                                                                0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                              0.0000
0.0000
0.0000
0.0000
0.0000
                                                   Pu date:
Am241:
Am date:
                                                                                                                                 0.0000
                                                                                                                                                                                                               0.0000
                                                                                       1. 50
64. 00
64. 00
1720
50. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                         Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                1.0000 +-
   Normalization constant:
                                                                                                                                           0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                            10
60
Passive error messages
No passive calibration curve calibration
No known alpha calibration
```

(1)

202409_AFAS-P Bottom Fork(UDL-1).txt

INCC 6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24, 09, 24 11::
Results file name: 490L3618.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Facility: PPFF 11:36:18 Ending comment: xx 1.500 64.000 1720 50, 0000

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 0. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 487 0. 000 0. 000 0. 000 0. 000 0. 045 0. 000 0. 000 60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 67510 1180 439

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

111. 030 +-1. 235 +-0. 000 +-

(1)

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Scaler1 6744 Singles 88 98 85 97 102 101 97 93 9538 9650 9546 245 251 214 221 229 232 202 6602 6766 9636 9575 9664 9710 9524 6738 6902 6764

Passive cycle rate data

1 158, 967 2, 617 0.000 2 160, 833 2, 550 0.000 3 159, 100 2, 150 0.000 4 157, 500 2, 067 0.000 5 160, 600 2, 117 0.000 6 159, 583 2, 183 0.000 7 161, 067 1, 750 0.000 8 161, 833 2, 033 0.000 9 158, 733 1, 367 0.000 10 156, 467 2, 483 0.000	0.000 Pass 0.000 Pass 0.000 Pass 0.000 Pass 0.000 Pass 0.000 Pass 0.000 Pass
--	--

(2)

202409_AFAS-P Bottom Fork(UDL-1).txt Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test —-1.000 110.204 111.030 0.993 1.000 0.000 0.181 0.546 0.040 0.000 data quality is inadequate.

Cycle raw data

Count time (sec): 60,000 Scaler2 QC O Pass O Pass O Pass O Pass O Pass O Pass Singles 6835 QC Tests Cvcle Α Scaler1 129 125 129 107 126 117 100 49 42 43 41 36 51 38 42 49 48 6715 6677 6527 6723 O Pass O Pass O Pass O Pass O Pass O Pass 6733 6883 112 122 113 6765 6858 10 6794

Cycle DTC rate data

Cycle	Singles	Doubles	Triples	QC Tests
1	112. 430	1. 333	0.000	Pass
2	110. 430	1. 383	0.000	Pass
3	109. 797	1. 433	0.000	Pass
4	107. 297	1. 100	0.000	Pass
5	110. 563	1. 500	0.000	Pass
6	110. 730	1. 100	0.000	Pass
7	113. 230	1. 033	0.000	Pass
8	111. 263	1. 167	0.000	Pass
9	112, 813	1, 217	0.000	Pass
10	111. 747	1. 083	0.000	Pass

```
INCC 5.1.2
```

24. 10. 11 10:22:06 4ABK2206. VER BWR TF

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Ispection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Pu 0.000 Verification Review disk file On Sample method Measured

Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +24.10.11
0.0000 +24.10.11 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000

1. 50 64. 00 64. 00 1720 50. 0000 0. 0080 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:

Normalization constant: Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 1.0000 +0.000 +0.000 +0.000 +0.000
0.000 0.0000 0.000 0.000 0.000

10 60

Number passive cycles: Count time (sec):

Passive error messages Known alpha analysis error

Passive results

(1)

202410_AFAS-B_Collar.txt

INCC 5.1.2

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS

24. 10. 11 10:39:11 4ABK3911. VER BWR COLLAR XXXX Pu 0.000 Verification Review disk file On Sample method Sample method Measured

Isotopics id:

Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.10.11 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 0. 1970 160. 0000 0. 6419 0. 1030 0. 0000 0. 0001 0. 0001

Normalization constant: Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 1.0000 +-0.000 +-0.000 +-0.000 +-0.000 0.000 0.0000 0.000 0.000 0.000

Number passive cycles: Count time (sec):

Passive error messages

Known alpha analysis error

Singles: Doubles: Triples: Scaler 1: Scaler 2: Passive cycle raw data

Passive cycle rate data

Cycle 1 2 3 4 5 6 7 8 9	Singles 67. 317 70. 283 68. 233 68. 083 68. 483 68. 767 70. 763 68. 883 67. 783	Doubles 0. 483 0. 400 0. 500 0. 367 0. 417 0. 283 0. 167 0. 317 -0. 033	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	QC Tests Pass Pass Pass Pass Pass Pass Pass P
--	--	--	---	---	---

(2)

202410_AFAS-B_Collar.txt

2199. 620 +-467. 046 +-0. 000 +-8. 138 +-0. 000 +-3. 212 2. 020 0. 000 0. 116 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4 5	Singles 132037 131633 132555 132267 131840 131811	R+A 46724 46493 46878 46711 46796 46533	A 18964 18607 18517 18684 18410	Scaler1 474 455 527 500 489 494	Scaler2 0 0 0 0 0	QC Test Pass Pass Pass Pass Pass Pass
6	131811	46533	18211	494	Ŏ	Pass
7	130596	45526	18359	490	0	Pass
8 9	131857 132887	46263 47255	18465 18944	492 456	0	Pass Pass
10	131823	46426	18612	506	0	Pass

Passive cycle rate data

9 2215. 571 472. 522 0. 000 0. 000 Pas						Pass Pass Pass Pass Pass Pass Pass Pass
--	--	--	--	--	--	--

```
INCC 5.1.2
```

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 24. 10. 11 10:55:13 4ABK5513. VER BWR BF Pu 0.000 Verification Review disk file Sample method Measured Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +24.10.11
0.0000 +24.10.11 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction: 1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000 0.000 0.000 0.000 Number passive cycles: Count time (sec): 10 60

(1)

202410_AFAS-P_Top Fork.txt

INCC 5.1.2

Passive error messages No known alpha calibration

INCC 5.1.2

Facility:

Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error cal culation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS 24. 10. 11 11:12:18 4ABL1218. VER PWR TF XXXX Pu 0.000 Verification Review disk file On Sample method Sample method Measured

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.10.11 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction 1. 50 64. 00 64. 00 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Normalization constant: 1 0000 +-

0.0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

10 60

Passive error messages

No known alpha calibration

Results

120. 603 +-1. 268 +-0. 000 +-4. 307 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Passive cycle rate data

Cycle 1 2 3 4 5 6 7 8 9	Singles 119, 750 121, 267 121, 183 119, 033 121, 367 120, 167 119, 633 122, 417 120, 950 120, 267	Doubles 1, 400 1, 417 1, 000 1, 100 1, 350 1, 167 1, 433 1, 117 1, 250 1, 450	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000	QC Tests Pass Pass Pass Pass Pass Pass Pass P
--	---	---	---	--	---

202410_AFAS-P_Top Fork.txt

(2)

Results

Singles:	2. 537 +-	0.066
Doubles:	0.000 +-	0.000
Triples:	0.000 +-	0.000
Scaler 1:	104. 392 +-	0.396
Scaler 2:	0.000 +-	0.000

Passive cycle raw data

Cycle	Singles	R+A	Α	Scaler1	Scaler2	QC Tests
1	142	0	0	6156	0	Pass
2	135	0	0	6296	0	Pass
3	134	0	0	6276	0	Pass
4	162	0	0	6385	0	Pass
5	157	0	0	6246	0	Pass
6	166	0	0	6188	0	Pass
7	148	0	0	6322	0	Pass
8	148	0	0	6305	0	Pass
9	168	0	0	6301	0	Pass
10	162	0	0	6160	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	2. 367	0.000	0.000	0.000	Pass
2	2, 250	0.000	0.000	0.000	Pass
3	2, 233	0.000	0.000	0.000	Pass
4	2, 700	0.000	0.000	0.000	Pass
5	2, 617	0.000	0.000	0.000	Pass
6	2, 767	0.000	0.000	0.000	Pass
7	2, 467	0.000	0.000	0.000	Pass
8	2, 467	0.000	0.000	0.000	Pass
9	2, 800	0.000	0.000	0.000	Pass
10	2. 700	0.000	0.000	0.000	Pass

202410_AFAS-P Top Fork(UDL-1).txt 202410_AFAS-P Top Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.10.11 11:
Results file name: 4ABL3411.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 101. 581 102. 450 0. 992 1. 000 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 6307 6196 6193 Scaler2 0 0 QC Tests Cycle Α Scaler1 51 50 43 43 40 35 27 45 35 57 Pass Pass Pass 110 114 104 92 81 107 101 Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 6284 6314 6267 6349 6188 6243 6414 8 9 10 88 107 Cycle DTC rate data Singles 102.975 101.125 101.075 102.592 103.092 102.308 103.675 100.992 101.908 104.758 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Doubles
0. 883
1. 000
1. 183
1. 017
0. 867
0. 767
1. 333
0. 933
0. 883
0. 833 QC Tests Pass Pass Cvcle Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 2. 142 0. 000 0. 000 0. 000 0. 000 234567 Pass Pass Pass Pass 10 60. 000 Pass 8 9 10 Pass Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 62755 1008 426 0 0 Summed multiplicity distributions A sums R+A sums Messages Normalization test -- data quality is inadequate Results 102. 450 0. 970 0. 000 Singles: Doubles: Triples: 0. 393 0. 055 0. 000 (1) (2) 202410_AFAS-P_Collar.txt 202410_AFAS-P_Collar.txt INCC 5.1.2 Results Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type: 1799. 173 +-306. 244 +-0. 000 +-16. 062 +-0. 000 +-Singles: Doubles: Triples: 1. 729 1. 448 0. 000 0. 173 0. 000 JM2G AFAS Scaler 1: Scaler 2: 24. 10. 11 11:33:26 4ABL3326. VER Passive cycle raw data Singles 108327 108056 107622 107459 107768 108283 107518 107886 108212 108205 Scaler2 0 0 0 0 0 0 0 0 0 PWR COLLAR XXXX Pu 0.000 Verification Review disk file On Cvcle R+A 30970 30765 30618 30514 30624 30867 30531 31162 30597 31022 Scaler1 QC Tests 974 968 961 1013 928 942 968 982 999 Stratum io. Material type: Original declared mass: Measurement option: Data source: QC tests: 12500 12514 12085 12370 12395 12387 12324 12668 12446 8 9 10 Passive cycle rate data Singles 1805, 732 1801, 214 1793, 978 1791, 261 1796, 412 1804, 998 1792, 244 1798, 380 Mass OC Tests
0.000 Pass
0.000 Pass Doubles 310. 544 304. 606 301. 921 307. 340 304. 422 308. 059 302. 587 314. 162 299. 003 309. 793 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Cycle 00 00 00 00 00 00 8 9 10 1803. 815 1803. 698

Error calculation: Accidentals method: Inspector name: Passive comment:	Sample method Measured			
Isotopics id: Isotopics source code: Pu238: Pu239: Pu240: Pu241: Pu242: Pu date: Am241: Am date:	Default 0D	0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000	0.0000 +- 0.0000 +- 100.0000 +- 0.0000 +- 24.10.11 0.0000 +- 24.10.11	0. 000 0. 000 0. 000 0. 000 0. 000
Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:	1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001			
Normalization constant:	1.0000 +-	0.0000		
Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:	0.000 +- 0.000 +- 0.000 +- 0.000 0.000	0.000 0.000 0.000		

Passive error messages

No passive calibration curve calibration No known alpha calibration

```
INCC 6.23.2.9 HM32 Not Validated
```

```
6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.10.11 12:
Results file name: 4ABM0108.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
                                                                                                                                                                                                                                                                       12:01:08
Predelay:
Gate length:
High voltage:
Die away time
Efficiency:
Multiplicity deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                                                                                                           1. 500
64. 000
1720
50. 0000
0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
   Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
Number of cycles:
Count time (sec):
                                                                                                                                                                                                                  14. 530
0. 002
0. 000
0. 000
0. 000
```

Summed raw data

Number of good cycles:	10
Total count time:	600
Shift register singles sum:	1081445
Shift register reals + accidentals sum:	309283
Shift register accidentals sum:	124549
Shift register 1st scaler sum:	0
Shift register 2nd scaler sum:	0

10 60. 000

Summed multiplicity distributions

R+A sums A sums

Results

Singles:	1788. 159	+-	2. 180
Doubles:	308.080	+-	0.873
Triples:	0.000	+-	0.000
Scaler 1:	0.000	+-	0.000
Scaler 2:	0.000	+-	0.000

(1)

202410_AFAS-P_Bottom Fork.txt

INCC 5.1.2

```
INCC 5.1.2

Facility:

Material balance area:
Detector type:
AFAS
Detector id:
Electronics id:
JO code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
XXX
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
CError calculation:
Accidentals method:
Inspector name:
Passive comment:

Jostopics id:
Defaut
                                                                                                                                                                                                                                                                                                     24. 10. 11 14:25:19
4AB02519. VER
                                                                                                                                                                                                                                                                                                  PWR BF
XXXXX
Pu
0.000
Verification
Review disk file
On
Sample method
Measured
            0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.10.11
0.0000 +-
24.10.11
                                                                                                                                                                                                                                                                                                                                                                                                                                                                     0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                        0.0000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 0.0000
                                                                                                                                                                                                                                                                                                  1. 01

1. 50

64. 00

1720

50. 0000

0. 0127

0. 0000

0. 0000

0. 0000

0. 0000

0. 0000

0. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1. 0000

1
   Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:
                                                                                                                                                                                                                                                                                                                                                      1 0000 +-
         Normalization constant:
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          0.0000
      Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                                                                                                                                                                                                                                                   0.000 +-
0.000 +-
0.000 +-
0.000
0.000
                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              0. 000
0. 000
0. 000
```

Passive error messages

No passive calibration curve calibration No known alpha calibration

Normalization results for reference source: H4-694

Cycle raw data

Count ti	me (sec):	60.000				
Cycle 1 2 3 4 5 6 7 8 9	Singles 107989 108307 107275 107881 108576 108379 108000 108137 108774 108127	R+A 30891 30916 30461 31070 30973 31088 30790 30891 31169 31034	A 12582 12382 12179 12511 12682 12531 12206 12604 12430 12442		1 Scal 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	er2 QC Tests O Pass
Cycle DT Cycle 1 2 3 4 5 6 7 8 9 10	Singles 1785.567 1790.868 1773.663 1783.766 1795.353 1782.069 1785.750 1788.034 1798.654 1787.867	Doub 16 305, 33 309, 09 304, 88 309, 50 305, 03 309, 47 309, 92 304, 97 312, 51	88 97 97 95 55 25 1	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	QC Tests Pass Pass Pass Pass Pass Pass Pass P	

(2)

202410_AFAS-P_Bottom Fork.txt

Results

		Sing Doubl Tripl Scaler Scaler	es: es: 1:	157. 347 1. 903 0. 000 111. 408 0. 000	+- 0.00 +- 0.00 +- 0.5	77 00 45
Passive	cycle raw d	ata				
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 9366 9323 9526 9378 9415 9503 9379 9528 9417 9573	R+A 216 204 229 217 190 219 225 236 202 215	A 102 96 117 120 79 101 117 87 102 90	Scaler1 6616 6679 6626 6506 6834 6812 6775 6737 6609 6651	Scaler2 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Passive	cycle rate	data				
Cycle 1 2 3 4 5 6 7 8 9	Singles 156, 100 155, 383 158, 767 156, 300 156, 917 158, 383 156, 317 158, 800 156, 950 159, 550	Double 1. 90 1. 80 1. 80 1. 61 1. 85 1. 96 1. 85 2. 44 1. 66 2. 08	00 00 57 57 50 57 00 33	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0. 000 0. 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P

INCC 6.23.2.9 HM32 Not Validated

6. 23. 2. 9 HM32 NOT VAIIDATED

Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24. 10. 11 14:
Results file name: 4AB04546. NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 657 0. 000 0. 000 0. 000 0. 000

10 60. 000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 67000 1185 495

Summed multiplicity distributions

A sums R+A sums

Messages

Normalization test -- data quality is inadequate

Results

110.010 +-1.150 +-0.000 +-Singles: Doubles: Triples: 0. 585 0. 078 0. 000

(1)

202411_AFAS-B_Top Fork.txt

INCC 5.1.2

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS

24. 11. 05 11:53:12 4B5L5312. VER BWR TF XXXX Pu 0.000 Verification Review disk file On

Sample method Measured

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.11.05 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 0. 1970 160. 0000 0. 6419 0. 1030 0. 0000 0. 0001 0. 0001

1.0000 +-0.000 +-0.000 +-0.000 +-0.000 0.000 Normalization constant: Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.0000 0.000 0.000 0.000

Number passive cycles: Count time (sec):

Passive error messages

Known alpha analysis error

Normalization results for reference source: H4-694

60.000

Current normalization constant: Gf252 expected singles rate: Gf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 108. 858 110. 010 0. 990 1. 000 data quality is inadequate

Cycle raw data Count time (sec):

Singles 6840 6625 6679 6611 6695 6605 6521 6825 Scaler2 0 0 0 QC Tests Pass Scaler1 Cycle Α 59 49 60 40 41 25 64 55 53 Pass Pass O Pass 122 107 8 9 10 97 130 137 6825 6806 6793

Cycle DTC rate data

Singles 112, 343 108, 760 109, 660 108, 527 109, 927 108, 427 107, 027 112, 093 111, 777 111, 560 Doubles 1, 117 1, 217 1, 050 1, 017 1, 183 1, 350 1, 367 0, 550 1, 250 1, 400 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 QC Tests Pass Pass Cvcle Pass Pass Pass Pass Pass 8 9 10 Pass Pass Pass

(2)

202411_AFAS-B_Top Fork.txt

66. 452 +-0. 328 +-0. 000 +-114. 907 +-0. 000 +-0. 248 0. 044 0. 000 0. 364 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

	-,					
Cycle	Singles	R+A	Α	Scaler1	Scaler2	QC Tests
1	3975	44	13	6984	0	Pass
2	3992	30	18	6885	0	Pass
3	3958	33	18	6960	0	Pass
4	4049	40	17	6857	0	Pass
5	4017	34	15	6980	0	Pass
6	3962	38	19	6818	0	Pass
7	4006	33	20	6832	0	Pass
8	4017	47	10	6927	0	Pass
9	4015	33	18	6910	0	Pass
10	3880	30	17	6791	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	66. 251	0. 517	0.000	0.000	Pass
2	66, 534	0. 200	0.000	0.000	Pass
3	65, 967	0. 250	0.000	0.000	Pass
4	67. 484	0. 383	0.000	0.000	Pass
5	66. 951	0.317	0.000	0.000	Pass
6	66. 034	0.317	0.000	0.000	Pass
7	66. 767	0. 217	0.000	0.000	Pass
8	66. 951	0.617	0.000	0.000	Pass
9	66. 917	0. 250	0.000	0.000	Pass
10	64 667	0.217	0.000	0.000	Pass

Passive cycle raw data

Singles

129556 129079

Passive cycle rate data

2160. 082 2155. 229 2159. 832 2158. 914 2159. 732 2156. 596 2153. 994 2154. 862 2160. 015 2152. 060

Cycle

9 10

Cycle

8 9 10

```
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                             24. 11. 05 10:25:48
4B5K2548. VER
                                                                                                             BWR COLLAR
                                                                                                             Pu
0.000
Verification
Review disk file
                                                                                                             On
Sample method
Measured
                                                                                                           Isotopics id:
     Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu date:
Am date:
                                                                                                                                                                                                                   0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.11.05
0.0000 +-
24.11.05
                                                                                                                                                                        0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                              0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                          0.0000
                                                                                                                                                                                                                                                                                 0.0000
                                                                                                                  1. 50
64. 00
64. 00
1720
50. 0000
0. 1970
160. 0000
0. 6419
0. 1030
0. 0000
0. 0001
0. 0001
                                Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
 Normalization constant:
Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                             1.0000 +-
0.000 +-
0.000 +-
0.000 +-
0.000
0.000
                                                                                                                                                                                        0.0000
0.000
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                                             10
60
```

Passive error messages Known alpha analysis error

Passive results

(1)

202411_AFAS-B_Bottom Fork.txt

INCC 5.1.2

```
Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
I tem id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                           Facility:
                                                                                                         JM2G
AFAS
                                                                                                         24. 11. 05 10:38:48
4B5K3848. VER
                                                                                                        BWR BF
XXXX
Pu
0.000
Verification
Review disk file
On
Sample method
                                                                                                         Sample method
Measured
                                                                                                      Isotopics id:
     Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                                                                                                 0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                            0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.11.05
                                                                                                                                                                                                                                                                    0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                                   0.0000
                                                                                                                                                                                                            0.0000
                                                                                                                                                                                                                                                                      0.0000
  Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction:
                                                                                                              1. 50
64. 00
64. 00
1720
50. 0000
0. 0060
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
   Normalization constant:
                                                                                                                           1 0000 +-
                                                                                                                                                                                 0.0000
                                                                                                                         0.000 +-
0.000 +-
0.000 +-
0.000
0.000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                                                                                 0.000
0.000
0.000
```

10 60

Passive error messages No known alpha calibration 202411_AFAS-B_Bottom Fork.txt

(2)

Results 116. 957 +-1. 215 +-0. 000 +-3. 845 +-0. 000 +-0. 386 0. 099 0. 000 0. 065 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Singles: Doubles: Triples: Scaler 1: Scaler 2:

A 17847 17860

Doubles 465. 445 457. 349 465. 445 470. 952 462. 908 458. 234 464. 325 461. 972 455. 431 453. 059

0. 936 1. 721 0. 000 0. 136 0. 000

QC Tests

QC Tests

Pass Pass

Pass Pass Pass Pass Pass Pass Pass

Pass Pass

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

Scaler1

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

Passive cycle raw data

Cycle	Singles	R+A	Α	Scaler1	Scaler2	QC Tests
1	7094	94	55	234	0	Pass
2	6997	134	43	235	0	Pass
3	6862	118	46	256	0	Pass
4	7033	125	63	229	0	Pass
5	6980	108	48	219	0	Pass
6	6978	130	48	236	0	Pass
7	7136	140	57	213	0	Pass
8	7029	149	47	240	Ō	Pass
9	7030	117	61	221	Ō	Pass
10	7035	131	49	224	0	Pass
Passive	e cycle rate	data				

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	118. 233	0. 650	0.000	0.000	Pass
2	116, 617	1, 517	0.000	0.000	Pass
3	114, 367	1, 200	0.000	0.000	Pass
4	117. 217	1. 033	0.000	0.000	Pass
5	116. 333	1. 000	0.000	0.000	Pass
6	116. 300	1. 367	0.000	0.000	Pass
7	118, 933	1, 383	0.000	0.000	Pass
8	117. 150	1. 700	0.000	0.000	Pass
9	117, 167	0. 933	0.000	0.000	Pass
10	117, 250	1. 367	0.000	0.000	Pass

Results

Cycle

8 9 10

Cycle

234567

8 9 10

Passive cycle raw data

Singles

118 131

Passive cycle rate data

Singles

2. 000 1. 967

2. 183 2. 533 2. 517 2. 200 2. 483 2. 383 2. 117 2. 283

```
202411_AFAS-P_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                          PWR TF
                                                                                          Pu
0.000
Verification
Review disk file
                                                                                          On
Sample method
Measured
     Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                       0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.11.05
0.0000 +-
24.11.05
                                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                             0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                          0.0000
                                                                                                                                                                                                                               0.0000
                                                                                              1. 50
64. 00
64. 00
1720
50. 0000
0. 0060
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                           Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                        1.0000 +-
   Normalization constant:
                                                                                                                                                     0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                       0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
Passive error messages
No known alpha calibration
```

(2)

```
202411_AFAS-P Top Fork(UDL-1).txt
```

(1)

```
INCC 6.23.2.9 HM32 Not Validated
```

```
Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.11.05 11:
Results file name: 4B5L1928.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
                                                       Facility: PPFF
                                                                                                                                                            11:19:28
                                    Ending comment: xx
                                                  Predelay:
                                                                                                                        1.500
```

	64.000	Gate Tength:
	1720	High voltage:
	50.0000	Die away time:
	0. 0126	Efficiency:
	0.0000	Multiplicity deadtime:
	0.0000	Coefficient A deadtime:
	0.0000	Coefficient B deadtime:
	0.0000	*Coefficient C deadtime:
	0.0001	Doubles gate fraction:
	0.0001	Triples gate fraction:
+	1. 908	Passive singles bkgrnd:
+	0 000	Passive doubles bkgrnd:

0. 052 0. 000 0. 000 Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 0. 000 0. 000 0. 000 0. 000

60.000

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 61567 1036 403

Summed multiplicity distributions

R+A sums A sums

Summed raw data

Normalization test -- data quality is inadequate

Results

100. 703 +-1. 055 +-0. 000 +-

(1)

2. 267 +-0. 002 +-0. 000 +-102. 442 +-0. 000 +-

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests

Pass Pass

Pass Pass

Pass Pass Pass Pass Pass Pass Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles 0.017 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

202411_AFAS-P Top Fork(UDL-1).txt Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test —-0.000 0.223 0.316 0.040 0.000 99. 776 100. 703 0. 991 1. 000 data quality is inadequate.

Cycle raw data

Count time (sec): 60,000 Singles 6123 6135 6304 6206 6126 Scaler2 QC 1 0 Pass 0 Pass 0 Pass 0 Pass 0 Pass QC Tests Cvcle Α Scaler1 109 40 46 40 32 38 43 41 41 37 45 109 103 121 97 6096 6143 6135 6162 6137 O Pass O Pass O Pass O Pass O Pass O Pass 82 104 10

Cycle DTC rate data

Cycle	Singles	Doubles 5 8 1	Triples	QC Tests
1	100, 142	1, 150	0,000	Pass
2	100, 342	1, 050	0.000	Pass
3	103, 158	1, 050	0.000	Pass
4	101, 525	1, 483	0.000	Pass
5	100, 192	0. 983	0.000	Pass
6	99.692	0, 650	0.000	Pass
7	100. 475	1, 050	0.000	Pass
8	100. 342	1. 117	0.000	Pass
9	100. 792	1. 250	0.000	Pass
10	100, 375	0. 767	0.000	Pass

202411_AFAS-P_Collar.txt 202411_AFAS-P_Collar.txt

INCC 5.1.2 INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 24. 11. 05 13:24:33 4B5N2433. VER PWR COLLAR Pu 0.000 Verification Review disk file On Sample method Measured Isotopics id: Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242: 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +24.11.05
0.0000 +24.11.05 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 Pu date: Am241: Am date: 0.0000 0.0000 1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction: 0. 0001 1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000

Number passive cycles: Count time (sec): Passive error messages

No passive calibration curve calibration No known alpha calibration

(1)

202411_AFAS-P Collar(UDL-1).txt

INCC 6.23.2.9 HM32 Not Validated

Facility: PPFF

Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.11.05 13:48:53
Results file name: ABSM4853.NOR
Inspection number:
Measurement option: Normalization
Data source: IAFA DataZ file
Detector configuration: Passive
QC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: XX Facility: PPFF Ending comment: xx

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1.500 64.000 1720 50, 0000 0. 1620 86. 5000 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001

12. 348 0. 000 0. 000 0. 000 0. 000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 0. 244 0. 000 0. 000

60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 1060148

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

1754. 835 +-301. 431 +-0. 000 +-

Results

Singles: Doubles: Triples: Scaler 1: Scaler 2: 1767. 602 +-299. 833 +-0. 000 +-15. 393 +-0. 000 +-2. 943 1. 724 0. 000 0. 089 0. 000

Passive cycle raw data

Scaler2 0 0 0 0 0 0 0 0 0 R+A 30194 29956 30200 29991 30183 29456 Cycle Singles Scaler1 QC Tests 911 951 891 924 107063 105335 106471 106086 12014 12085 11956 11780 106095 105606 106213 105196 930 936 925 907 29691 29630 30355 30129 9 10 106398 105936

Passive cycle rate data

Mass QC Tests
0.000 Pass
0.000 Pass Doubles 306. 039 307. 120 301. 452 294. 563 303. 002 289. 693 295. 764 297. 680 303. 203 299. 816 Singles 1784.659 Cvcle Triples 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 1755. 850 1774. 789 234567 1774. 789 1768. 370 1768. 520 1760. 368 1770. 488 1753. 532 1773. 572 1765. 870

(2)

202411_AFAS-P Collar(UDL-1).txt

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant:
Cf252 expected doubles rate:
Cf252 measured doubles rate:
Doubles rate expected/measured:
New normalization constant:
Normalization test—
Measured percent precision:
Required percent precision:
Repeat measurement for at least: 0. 000 0. 449 1. 250 0. 004 0. 000 295. 756 301. 431 0. 981 1. 000 1.000 +- 0.000 data quality is inadequate. 0.415 0.300 0.000

Cycle raw data

Count time (sec): 60.000

R+A 30282 29915 30192 30029 30352 30098 Singles 106107 106288 Scaler2 QC 0 Pass 0 Pass QC Tests Cvcle A 11744 11977 Scaler1 11915 12202 12168 11852 11591 11966 106243 106031 Pass Pass O Pass 106396 105956 105418 106148 29635 29740 106000 105561 29975 29757 12089 11723 10

Cycle DTC rate data

Triples QC Tests 0.000 Pass Singles 1756. 372 1759. 390 1758. 639 1755. 105 1761. 190 1753. 855 1744. 885 Doubles 309, 156 299, 150 304, 803 297, 298 303, 253 304, 286 300, 916 296, 415 298, 282 300, 750 Cvcle 1757. 056 1754. 588 1747. 269 10

152. 360 +-1. 967 +-0. 000 +-109. 275 +-0. 000 +-

Scaler1

6667 6477

6674 6523

6591 6579

Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000

0. 338 0. 093 0. 000 0. 376 0. 000

Mass QC Tests
0.000 Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles 1. 967 1. 800

. 800 . 350 . 817 . 100

1. 817 2. 100 1. 850 2. 200 2. 233 2. 400

1. 950

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                          24. 11. 05 11:32:03
4B5L3203. VER
                                                                                          PWR BF
                                                                                          Pu
0.000
Verification
Review disk file
                                                                                         On
Sample method
Measured
                                                                                       Isotopics id:
    Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                            0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.11.05
0.0000 +-
24.11.05
                                                                                                                                          0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                              0.0000
0.0000
0.0000
0.0000
0.0000
                                                       Pu date:
Am241:
Am date:
                                                                                                                                           0.0000
                                                                                                                                                                                                                               0.0000
                                                                                              1. 50
64. 00
64. 00
1720
50. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                           Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                        1.0000 +-
   Normalization constant:
                                                                                                                                                      0.0000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                       0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
Passive error messages
No passive calibration curve calibration
No known alpha calibration
```

(1)

202411_AFAS-P Bottom Fork(UDL-1).txt

```
INCC 6.23.2.9 HM32 Not Validated
```

```
6. 23. 2. 9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24.11.05 11:59:27
Results file name: 485L5927.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment:
                                                     Ending comment: xx
Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                                                                           1.500
64.000
                                                                                                                                                                        1720
50, 0000
                                                                                                                                                                             0. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
```

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 312 0. 000 0. 000 0. 000 0. 000 0. 029 0. 000 0. 000

60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:

Summed multiplicity distributions

R+A sums A sums

Results

Singles:	107. 915	+-	0. 215
Doubles:	1.052	+-	0.042
Triples:	0.000	+-	0.000
Scaler 1:	0.000	+-	0.000
Scaler 2:	0.000	+-	0.000

202411_AFAS-P Bottom Fork(UDL-1).txt

(2)

Normalization results for reference source: H4-694

Current normalization constant: 1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (
1.000 +- (Cf252 expected singles rate:
Cf252 measured singles rate:
Cf252 measured singles rate:
Singles rate expected/measured:
New normalization constant:
Normalization test — 0. 176 0. 215 0. 040 0. 000

Cycle raw data Count time (sec):

Results

Cycle

9 10

Cycle

234567

Passive cycle raw data

Singles

Passive cycle rate data

Singles 152. 767 153. 167 153. 450 153. 167 151. 767 152. 733 152. 350 149. 700 152. 333 152. 167

oount	LTING (300).	00.000			
Cycle	Singles	R+A	Α	Scaler1	Scaler2 QC Tests
1	6506	98	46	0	0 Pass
2	6541	104	43	0	0 Pass
3	6615	127	51	0	0 Pass
4	6593	124	59	0	0 Pass
5	6608	115	56	0	0 Pass
6	6553	111	45	0	0 Pass
7	6494	93	39	0	0 Pass
8	6534	120	47	0	0 Pass
9	6548	112	44	0	0 Pass
10	6544	99	42	ŏ	0 Pass

Cycle DTC rate data

Cycle	Singles	Doubles	Triples	QC Tests
1	107. 122	0. 867	0.000	Pass
2	107. 705	1. 017	0.000	Pass
3	108. 938	1. 267	0.000	Pass
4	108. 572	1. 083	0.000	Pass
5	108. 822	0. 983	0.000	Pass
6	107. 905	1. 100	0.000	Pass
7	106. 922	0. 900	0.000	Pass
8	107. 588	1. 217	0.000	Pass
9	107. 822	1. 133	0.000	Pass
10	107. 755	0. 950	0.000	Pass

Passive cycle raw data

Singles

3767 3760

Passive cycle rate data

Singles 62.783 62.667 64.767

62. 783 62. 667 64. 767 64. 933 64. 600 63. 550 63. 000 63. 983 63. 167 66. 433

Cycle

9 10

Cycle

8 9 10

```
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Isspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                   24. 12. 10 10:53:57
4CAK5357. VER
                                                                                                                   BWR TF
                                                                                                                   Pu
0.000
Verification
Review disk file
                                                                                                                   On
Sample method
Measured
                                                                                                                Isotopics id:
      Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu date:
Am date:
                                                                                                                                                                                                                              0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.12.10
0.0000 +-
24.12.10
                                                                                                                                                                                0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                                            0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                  0.0000
                                                                                                                                                                                                                                                                                              0.0000
                                                                                                                        1. 50
64. 00
64. 00
1720
50. 0000
0. 0080
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                                 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
 Normalization constant:
Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                                    1. 0000
0. 000 =
0. 000 =
0. 000 =
0. 000
                                                                                                                                                                                                 0.0000
0.000
0.000
0.000
```

Number passive cycles: Count time (sec): 10 60 Passive error messages

Known alpha analysis error Passive results

(1)

202412_AFAS-B_Collar.txt

INCC 5.1.2

```
Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Litem id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                              Facility:
                                                                                                               JM2G
AFAS
                                                                                                               24. 12. 10 11:11:02
4CAL1102. VER
                                                                                                             BWR COLLAR
XXXX
Pu
0.000
Verification
Review disk file
On
                                                                                                               Sample method
Measured
                                                                                                            Isotopics id:
    Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                                                                                                           0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                         0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.12.10
                                                                                                                                                                                                                                                                                     0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                                              0.0000
                                                                                                                                                                                                                                                                                        0.0000
                                                                                                                                                                                                                         0.0000
Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction
                                                                                                                    1. 50
64. 00
64. 00
1720
50. 0000
0. 1970
                                                                                                                        0. 1970
160. 0000
0. 6419
0. 1030
0. 0000
0. 0001
0. 0001
 Normalization constant:
Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                                1.0000 +-
0.000 +-
0.000 +-
0.000 +-
0.000
0.000
                                                                                                                                                                                            0.0000
0.000
0.000
0.000
     Number passive cycles:
Count time (sec):
```

Passive error messages Known alpha analysis error (2)

63. 988 +-0. 275 +-0. 000 +-111. 127 +-0. 000 +-

Scaler1

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests

QC Tests

Pass Pass

Pass

Pass Pass Pass Pass Pass Pass Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles

0. 250 0. 417 0. 133

0. 333 0. 217 0. 300 0. 250 0. 317 0. 350 0. 183

			20241	2_AFAS-B_Coll	ar.txt	
		I So	Singles: Doubles: Triples: caler 1: caler 2:	2108. 130 449. 059 0. 000 7. 808 0. 000	+- 1. +- 0. +- 0.	. 666 . 066 . 000 . 185 . 000
Passive	cycle raw da	ita				
Cycle 1 2 3 4 5 6 7 8 9 10	Singles 126573 126335 126193 126618 126540 125915 126381 126130 126867 126898	R+A 43868 43953 43849 43944 43814 43744 43554 44071 44295 44506	A 17213 16928 16879 16840 17036 16905 16958 17024 17466 17278	Scaler1 465 496 394 490 438 444 475 499 511 473	Scaler2 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Passive	cycle rate o	lata				
Cyc e 1 2 3 4 5 6 7 8 9	Singles 2110, 265 2106, 295 2103, 3977 2111, 015 2109, 714 2099, 290 2107, 062 2102, 876 2115, 168 2115, 685	Doub 444. 451. 450. 452. 446. 447. 443. 451. 454.	852 026 107 346 905 920 867 392 758	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Pass Pass Pass Pass Pass Pass Pass

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles
1. 433
1. 067
1. 367
1. 000
1. 133
1. 183
0. 967
1. 367
1. 350
1. 083

114. 703 +-1. 195 +-0. 000 +-3. 843 +-0. 000 +-

Scaler1

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

0. 448 0. 054 0. 000 0. 115 0. 000

QC Tests

QC Tests

Pass Pass

Pass

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

Results

Cycle

8 9 10

Cycle

8 9 10

Passive cycle raw data

Passive cycle rate data

Singles 115.500 113.817 114.900 115.600 112.533 115.883 117.017 113.867 112.833 115.083

INCC 5.1.2

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Isspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                           24. 12. 10 11:27:04
4CAL2704. VER
                                                                                           BWR BF
                                                                                           Pu
0.000
Verification
Review disk file
                                                                                           On
Sample method
Measured
                                                                                         Isotopics id:
     Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu date:
Am date:
                                                                                                                                                                               0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.12.10
0.0000 +-
24.12.10
                                                                                                                                           0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                             0.0000
                                                                                                                                                                                                                                  0.0000
                                                                                               1. 50
64. 00
64. 00
1720
50. 0000
0. 0060
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                           Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                          1.0000 +-
   Normalization constant:
                                                                                                                                                       0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                        0.000 +
0.000 +
0.000 +
0.000 +
0.000
                                                                                                                                                        0.000
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                     10
60
Passive error messages
No known alpha calibration
```

(1)

202412_AFAS-P_Top Fork.txt

INCC 5.1.2

```
Facility:
     Material balance area:
Detector type:
Detector id:
Electronics id:
                                                                                                   JM2G
AFAS
Detector Id.
Electronics id.
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id.
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                   24. 12. 10 13:29:37
4CAN2937. VER
                                                                                                  PWR TF
XXXX
Pu
0.000
Verification
Review disk file
On
                                                                                                   Sample method
Measured
                                                                                                Isotopics id:
    Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                 0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.12.10
                                                                                                                                                                                                                                                       0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                           0.0000
                                                                                                                                                                                                 0.0000
                                                                                                                                                                                                                                                         0.0000
Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction
                                                                                                       1. 50
64. 00
64. 00
1720
50. 0000
0. 0126
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
  Normalization constant:
                                                                                                                    1 0000 +-
                                                                                                                                                                       0.0000
                                                                                                                  0.000 +-
0.000 +-
0.000 +-
0.000
0.000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                                                                       0.000
0.000
0.000
```

10 60 Passive error messages

No known alpha calibration

(1)

(2)

202412_AFAS-P_Top Fork.txt Results Singles:

0. 252 +-0. 000 +-0. 000 +-99. 417 +-0. 000 +-0. 252 0. 000 0. 000 0. 382 0. 000 Doubles: Triples: Scaler 1: Scaler 2:

Scaler1 5847 5875 5940 5974 5929 6083 6021 5953 6040 5988 Scaler2 0 0 0 0 0 0 0 0 0 0 Singles 151 Cvcle R+A QC Tests Α 00000000000 Pass Pass 0000000000 Pass Pass Pass Pass Pass Pass 9 10 Pass Pass

Passive cycle rate data

Passive cycle raw data

Cvcle	Singles	Doubles	Triples	Mass	QC Tests
1	2, 517	0.000	0,000	0.000	Pass
2	0.000	0, 000	0,000	0.000	Pass
3	0.000	0, 000	0,000	0.000	Pass
4	0.000	0.000	0.000	0.000	Pass
5	0.000	0.000	0.000	0.000	Pass
6	0.000	0.000	0.000	0.000	Pass
7	0.000	0.000	0.000	0.000	Pass
8	0.000	0.000	0.000	0.000	Pass
9	0.000	0.000	0.000	0.000	Pass
10	0.000	0.000	0.000	0.000	Pass

14:00:24

202412_AFAS-P Top Fork(UDL-1).txt

Normalization results for reference source: H4-694

Cycle raw data

Cycle	Singles	R+A	Α	Scaler1	Scaler2 QC Tests
1	5896	84	37	0	0 Pass
2	5903	95	40	0	0 Pass
3	6035	92	39	0	0 Pass
4	5950	103	50	0	0 Pass
5	6050	87	38	0	0 Pass
6	5996	81	51	0	0 Pass
7	6027	80	38	0	0 Pass
8	5996	108	39	0	0 Pass
9	6014	97	42	0	0 Pass
10	6046	94	28	0	0 Pass

Cycle DTC rate data

Cycle	Singles	Doubles 5 8 1	Triples	QC Tests
1	96, 468	0. 778	0,000	Pass
2	96. 585	0. 912	0.000	Pass
3	98, 785	0. 878	0,000	Pass
4	97. 368	0. 878	0.000	Pass
5	99. 035	0.812	0.000	Pass
6	98, 135	0. 495	0,000	Pass
7	98, 652	0, 695	0,000	Pass
8	98, 135	1, 145	0,000	Pass
9	98, 435	0.912	0,000	Pass
10	98. 968	1. 095	0.000	Pass

Summed raw data

Number of good cycles: Total count time: Shift register singles sum: Shift register accidentals sum: Shift register accidentals sum: Shift register 1st scaler sum: Shift register 1st scaler sum: 10 600 59913 921 402

Summed multiplicity distributions

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec):

INCC 6.23.2.9 HM32 Not Validated

6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 24.12.10 14:
Results file name: 4CA00024.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx

1. 500 64. 000 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

1. 798 0. 005 0. 000 0. 000 0. 000 10 60. 000

A sums R+A sums

Messages

Normalization test -- data quality is inadequate

Results

Singles: Doubles: Triples: 98.057 + 0.860 +-0.000 +-

(1)

202412_AFAS-P_Collar.txt

Results

1725. 632 +-295. 043 +-0. 000 +-15. 183 +-0. 000 +-Singles: 1. 123 1. 131 0. 000 0. 204 0. 000 Doubles: Triples: Scaler 1: Scaler 2:

(2)

202412_AFAS-P_Collar.txt

Passive cycle raw data

Cycle	Singles	R+A	Α	Scaler1	Scaler2	QC Tests
1	103274	29211	11305	942	0	Pass
2	103543	29270	11434	867	0	Pass
3	103353	28783	11298	862	0	Pass
4	103576	28899	11208	981	0	Pass
5	103392	28751	11523	878	0	Pass
6	103183	28837	11172	901	0	Pass
7	103626	28914	11320	884	0	Pass
8	103714	29197	11416	936	0	Pass
9	103795	29250	11453	926	0	Pass
10	103769	29408	11471	933	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
- 1	1721. 490	298. 611	0.000	0.000	Pass
2	1725. 974	297. 444	0.000	0.000	Pass
3	1722. 807	291. 590	0.000	0.000	Pass
4	1726, 524	295, 026	0.000	0.000	Pass
5	1723. 457	287. 305	0.000	0.000	Pass
6	1719. 972	294. 592	0.000	0.000	Pass
7	1727. 358	293. 409	0.000	0.000	Pass
8	1728. 825	296. 527	0.000	0.000	Pass
9	1730. 175	296. 794	0.000	0.000	Pass
10	1729, 742	299, 129	0.000	0.000	Pass

INCC 5.1.2

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: JM2G AFAS 24. 12. 10 13:45:43 4CAN4543. VER PWR COLLAR XXXX Pu 0.000 Verification Review disk file On Sample matter Sample method Measured 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-24.12.10 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 1 0000 +-Normalization constant: 0 0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

Passive error messages

No passive calibration curve calibration No known alpha calibration

```
202412_AFAS-P Collar(UDL-1).txt
 INCC 6.23.2.9 HM32 Not Validated
                   6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 24.12.10 14:
Results file name: 4CA01938.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
                  Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                    1. 500
64. 000
1720
50. 0000
                                                                                                                    0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
                    Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
Number of cycles:
Count time (sec):
                                                                                                                        11. 807
0. 005
0. 000
0. 000
0. 000
                                                                                                                        10
60. 000
Summed raw data
Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
                                                                                                                                                        1035928
291310
114507
Summed multiplicity distributions
                                  R+A sums
                                                                           A sums
Messages
Normalization test -- data quality is inadequate
```

Results

1714. 998 +-294. 843 +-0. 000 +-Singles: Doubles: Triples: 1. 458 1. 099 0. 000

(1)

202412_AFAS-P_Bottom Fork.txt

INCC 5.1.2

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                                    JM2G
AFAS
                                                                                                                                    24. 12. 10 14:02:48
4CA00248. VER
                                                                                                                                  PWR BF
XXXX
Pu
0.000
Verification
Review disk file
On
                                                                                                                                    Sample method
Measured
                                                                                                                                Default

0D

0.0000 +-

0.0000 +-

100.0000 +-

0.0000 +-

0.0000 +-

00.01.01

0.0000 +-

00.01.01
                                                       Isotopics id:
        Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                             0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
24.12.10
                                                                                                                                                                                                                                                                                                                                    0. 0000
0. 0000
0. 0000
0. 0000
0. 0000
                                                                                                                                                                                                           0.0000
                                                                                                                                                                                                                                                              0.0000
                                                                                                                                                                                                                                                                                                                                       0.0000
    Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Triples gate fraction:
                                                                                                                                         1. 50
64. 00
64. 00
1720
50. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
```

1 0000 +-

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

Passive error messages

Normalization constant:

No passive calibration curve calibration No known alpha calibration

Normalization results for reference source: H4-694

Current normalization constant:
 Cf252 expected doubles rate:
 Cf252 measured doubles rate:
 Doubles rate expected/measured:
 New normalization constant:
 Normalization test - Measured percent precision:
 Required percent precision:
Repeat measurement for at least: 1.000 +- 0.000 288.417 -- 0.438 294.843 +- 1.099 0.978 -- 0.004 1.000 -- 0.000 data quality is inadequate. 0.373 0.300 0.000

Cycle raw data

Count time (sec): 60.000 Cycle Singles
1 103293
2 103539
3 103434
4 103944 Scaler2 QC Tests O Pass O Pass R+A 28844 28854 Scaler1 11260 11506 11531 11260 11619 29097 29175 Pass Pass O Pass O Pass O Pass O Pass O Pass O Pass 29568 29136 29172 29044 29204 29216 103907 103201 103754 103304 103852 103700 11361 11731 11434 8 9 10

Cycle DTC rate data

(2)

202412_AFAS-P_Bottom Fork.txt

Results

149. 938 +-1. 998 +-0. 000 +-107. 208 +-0. 000 +-0. 682 0. 054 0. 000 0. 344 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	150, 567	1, 733	0.000	0.000	Pass
2	149, 600	2, 117	0.000	0.000	Pass
3	145, 683	2. 083	0.000	0.000	Pass
4	151, 633	2. 117	0.000	0.000	Pass
5	148. 183	1. 700	0.000	0.000	Pass
6	153, 567	1. 950	0.000	0.000	Pass
7	151. 317	2. 167	0.000	0.000	Pass
8	148. 917	2. 183	0.000	0.000	Pass
9	149. 200	1. 950	0.000	0.000	Pass
10	150 717	1 093	0.000	0.000	Dacc

0 0000

0.000 0.000 0.000

202412_AFAS-P Bottom Fork(UDL-1).txt 202412_AFAS-P Bottom Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6. 23. 2. 9 HM32 NOT VAIIDATED

Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 24. 12. 10 14:
Results file name: 4CA03327. NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Gf252 expected singles rate: Gf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 104. 272 106. 102 0. 983 1. 000 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 6519 Scaler2 0 0 QC Tests Cycle Α Scaler1 40 52 40 36 49 38 51 51 49 42 Pass 6331 6408 Pass Pass Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 6402 6405 6443 6418 6470 6580 6440 118 105 103 132 8 9 10 108 126 77 Cycle DTC rate data Singles 107, 392 104, 258 105, 542 105, 442 105, 492 106, 125 105, 708 106, 575 108, 408 106, 075 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Doubles
1, 295
1, 112
1, 162
1, 362
0, 928
1, 078
1, 345
0, 945
1, 278
0, 578 QC Tests Pass Pass Cvcle Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 258 0. 005 0. 000 0. 000 0. 000 234567 Pass Pass Pass Pass Pass 60.000 8 9 10 Pass Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 64416 1116 448 Summed multiplicity distributions R+A sums A sums Messages Normalization test -- data quality is inadequate Results 106. 102 +-1. 108 +-0. 000 +-Singles: Doubles: Triples: 0.368 0.077 0.000 (1) (2) 202501_AFAS-B_Top Fork.txt 202501_AFAS-B_Top Fork.txt INCC 5.1.2 Results Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory charge code: 60. 807 +-0. 267 +-0. 000 +-107. 845 +-0. 000 +-0. 406 0. 037 0. 000 0. 443 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2: JM2G AFAS Detector Id.
Electronics id.
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id.
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Passive cycle raw data 25. 01. 17 10:08:07 51HK0807. VER Singles 3734 3587 3739 3641 3522 3637 3695 3606 3744 3579 Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 6422 6395 6509 6412 6660 6461 6555 6403 6473 6417 Cvcle QC Tests 15 14 19 17 12 10 18 12 21 Pass Pass BWR TF XXXX Pu 0.000 Verification Review disk file On 20 28 32 29 28 35 37 22 38 39 Sample method Measured 9 10 Passive cycle rate data Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Singles 62. 233 59. 783 62. 317 60. 683 58. 700 60. 617 61. 583 60. 100 62. 400 59. 650 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Doubles 0. 083 0. 233 0. 217 QC Tests Pass Pass Isotopics id: Cvcle Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0. 217 0. 200 0. 267 0. 417 0. 317 0. 167 0. 283 0. 483 Pass Pass Pass Pass Pass Pass Pass 25 01 0.0000 0.0000 0.0000 25.01.17 10 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0080 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001

Number passive cycles: 10 Count time (sec): 60 Passive error messages

Normalization constant:

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

Known alpha analysis error

(1)

0 0000

0.000 0.000 0.000

1 0000 +-

0.000 +-0.000 +-0.000 +-0.000 0.000

```
INCC 5.1.2
                                                                                           Results
                                                                                           0..........
```

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 25. 01. 17 10:27:12 51HK2712. VER BWR COLLAR Pu 0.000 Verification Review disk file On Sample method Measured

Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 +-0.0000 +-100.0000 +-0.0000 +-25.01.17 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 25.01.17

1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 160. 0000 0. 6419 0. 1030 0. 0000 0. 0001 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:

1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000 0.000 0.000 0.000

Number passive cycles: Count time (sec): 10 60

Passive error messages Known alpha analysis error

202501_AFAS-B_Bottom Fork.txt

(1)

INCC 5.1.2

JM2G AFAS

INCC 5.1.2

Facility:

Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error cal culation:
Accidentals method:
Inspector name:
Passive comment: 25. 01. 17 10:44:20 51HK4420. VER BWR BF XXXX Pu 0.000 Verification Review disk file On Sample method

Sample method Measured

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-25.01.17 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 25.01.17 0.0000

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0.0000 0.000 0.000 0.000

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

10 60

Passive error messages

No known alpha calibration

2050. 470 +-438. 952 +-0. 000 +-7. 262 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4 5	Singles 123109 123244 123229 122622 123161 122781	R+A 42113 42211 42680 42223 42758 42105	A 16346 16205 16159 15815 16235 15772	Scaler1 435 406 396 387 478 437	Scaler2 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass
3	123229	42680	16159	396		
					0	
					0	
6					0	
/	122183	42174	16032	478	0	Pass
8	123236	42856	16284	439	0	Pass
9	123297	42483	16022	457	0	Pass
10	123297	42256	15964	444	0	Pass

Passive cycle rate data

Cycle 1 2 3 4	Singles 2052, 493 2054, 744 2054, 494 2044, 371	Doubles 430. 016 434. 005 442. 600 440. 711	Triples 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000	QC Tests Pass Pass Pass Pass
5	2053, 360	442, 633	0.000	0.000	Pass
6	2047. 022	439. 460	0.000	0.000	Pass
7	2037. 049	436. 270	0.000	0.000	Pass
8	2054. 611	443. 451	0.000	0.000	Pass
9	2055. 628	441. 599	0.000	0.000	Pass
10	2050. 925	438. 777	0. 000	0.000	Pass

(2)

202501_AFAS-B_Bottom Fork.txt

Results

Singles:	111.972 +-	0.520
Doubles:	1. 232 +-	0.080
Triples:	0.000 +-	0.000
Scaler 1:	3. 358 +-	0.080
Scaler 2:	0.000 +-	0.000

Passive cycle raw data

Cycle	Singles 6606	R+A	A 47	Scaler1 207	Scaler2	QC Tests
- 1		98	41		0	Pass
2	6604	113	39	204	0	Pass
3	6758	158	51	214	0	Pass
4	6926	133	47	200	0	Pass
5	6650	105	39	197	0	Pass
6	6748	116	43	204	0	Pass
7	6669	123	40	231	0	Pass
8	6743	129	60	191	0	Pass
9	6677	112	47	173	0	Pass
10	6802	118	53	194	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	110. 100	0. 850	0.000	0.000	Pass
2	110, 067	1. 233	0.000	0.000	Pass
3	112, 633	1. 783	0.000	0.000	Pass
4	115. 433	1. 433	0.000	0.000	Pass
5	110. 833	1. 100	0.000	0.000	Pass
6	112. 467	1. 217	0.000	0.000	Pass
7	111, 150	1. 383	0.000	0.000	Pass
8	112, 383	1, 150	0.000	0.000	Pass
9	111, 283	1. 083	0.000	0.000	Pass
10	113, 367	1. 083	0.000	0.000	Pass

Results

Cycle

8 9 10

Cycle

567

8 9 10

Passive cycle raw data

Singles

Passive cycle rate data

Singles 2.200 1.917

1. 917 1. 783 2. 250 2. 067 1. 850 1. 767 2. 150 2. 133 2. 017

132 115 107

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles

0. 000 0. 000 0. 000 0. 000 -0. 017 0. 000 0. 000 0. 000 0. 000 0. 000 2. 013 +--0. 002 +-0. 000 +-96. 497 +-0. 000 +-

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

> Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests

QC Tests

Pass Pass

Pass Pass Pass Pass Pass Pass

Pass Pass

```
202501_AFAS-P_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                         25. 01. 17 10:59:24
51HK5924. VER
                                                                                         PWR TF
                                                                                         Pu
0.000
Verification
Review disk file
                                                                                        On
Sample method
Measured
                                                                                       Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
                                                                                                                                                                            0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
25.01.17
                                                                                                                                        0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                           0.0000
0.0000
0.0000
0.0000
0.0000
                                                      Pu date:
Am241:
Am date:
                                                                                                                                          0.0000
                                                                                                                                                                                                                             0.0000
                                                                                                                                                                              0.0000
25.01.17
                                                                                             1. 50
64. 00
64. 00
1720
50. 0000
0. 0126
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                          Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                       1.0000 +-
   Normalization constant:
                                                                                                                                                    0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                      0.000 +
0.000 +
0.000 +
0.000 +
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                   10
60
Passive error messages
No known alpha calibration
                                                                                                                                                         (1)
                                                                                                  202501_AFAS-P Top Fork(UDL-1).txt
INCC 6.23.2.9 HM32 Not Validated
                 Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 25.01.17 11:
Results file name: 51HL3122.NOR
Inspection number:
Measurement option: Normalization
Data source: IAFA DataZ file
Detector configuration: Passive
0 C tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: XX
                                                                Facility: PPFF
                                                                                                                                                     11:31:22
                                               Ending comment: xx
                 Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient B deadtime:
*Coefficient C deadtime:
Triples gate fraction:
                                                                     Predelay:
                                                                                                                     1.500
64.000
                                                                                                                  1720
50, 0000
                                                                                                                     0. 0000
0. 0126
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
                    Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                        1. 423
0. 010
0. 000
0. 000
0. 000
                                                                                                                                                                             0. 039
0. 009
0. 000
Messages
```

Normalization test -- data quality is inadequate

910 378

95. 225 0. 877 0. 000

(1)

10 60. 000

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register lst scaler sum:
Shift register lst scaler sum:
Shift register 2nd scaler sum:

R+A sums A sums Number of cycles: Count time (sec):

Summed multiplicity distributions

Summed raw data

Results

```
(2)
                                                                                202501_AFAS-P Top Fork(UDL-1).txt
                                                                                          Scaler 1:
Scaler 2:
Normalization results for reference source: H4-694
                            Current normalization constant:

Cf252 expected singles rate:

Cf252 measured singles rate:

Singles rate expected/measured:

New normalization constant:

Normalization test —-
                                                                                                                                1.000
94.684
95.225
0.994
1.000
                                                                                                                                                                             0. 000
0. 212
0. 314
0. 040
0. 000
                                                                                                                    data quality is inadequate.
Cycle raw data
Count time (sec):
                                                                60,000
                     Singles
5773
5827
5845
5751
5818
                                                                                                                                                Scaler2 QC O Pass O Pass O Pass O Pass O Pass O Pass
                                                                                                                                                                          QC Tests
Cvcle
                                                                                         Α
                                                          R+A
                                                                                                                   Scaler1
                                                                 85
72
121
90
89
94
94
92
79
94
                                                                                                  34
50
29
50
30
45
35
37
39
                                                                                                                                                                 O Pass
O Pass
O Pass
O Pass
O Pass
O Pass
                               5865
5770
5779
5683
5878
     10
Cycle DTC rate data
                              Singles
94, 793
95, 693
95, 993
94, 427
95, 543
96, 327
94, 743
94, 893
93, 293
96, 543
                                                                     Doubles
0. 840
0. 357
1. 523
1. 007
0. 640
1. 057
0. 807
0. 940
0. 690
0. 907
                                                                                                             Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
                                                                                                                                    QC Tests
Pass
Pass
Cycle
                                                                                                                                      Pass
Pass
Pass
Pass
Pass
Pass
      10
```

添付-49

202501_AFAS-P_Collar.txt 202501_AFAS-P_Collar.txt

Results

INCC 5.1.2 INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Singles: Doubles: Triples: Scaler 1: Scaler 2: 1675. 991 +-286. 923 +-0. 000 +-14. 532 +-0. 000 +-2. 040 1. 414 0. 000 1. 205 0. 000 Passive cycle raw data 25. 01. 17 11:18:26 51HL1826. VER Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 874 864 828 Cycle R+A 27924 27859 27577 QC Tests Singles 10786 10706 10742 10743 10598 10981 10727 11036 PWR COLLAR 99937 100299 100234 100293 100359 101043 100762 101027 101274 28283 28030 28069 28229 27807 27998 28370 Pu 0.000 Verification Review disk file 881 892 891 843 On Theoretical method Measured 822 908 916 8 9 10 Passive cycle rate data Default C0 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Singles 1670. 591 1665. 857 1671. 892 1670. 808 1671. 792 1672. 892 1684. 295 1679. 611 1684. 028 1688. 146 Doubles 285. 798 286. 048 280. 746 292. 502 290. 701 284. 965 291. 870 279. 679 283. 732 293. 188 Isotopics id: Cycle QC Tests Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242: Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 Pass Pass 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-25.01.17 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 Pass Pass Pass Pass Pass Pass 567 Pu date: Am241: Am date: 0.0000 0.0000 8 9 10 0.0000 25.01.17 Pass Pass 1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction: 0. 0001 1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000 Number passive cycles: Count time (sec): Passive error messages No known alpha calibration (1) (2) 202501_AFAS-P Collar(UDL-1).txt 202501_AFAS-P Collar(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated Scaler 1: Scaler 2: Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 25.01.17 11:48:40
Results file name: 51HL4840.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
QC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Facility: PPFF Normalization results for reference source: H4-694 Current normalization constant:
Cf252 expected doubles rate:
Cf252 measured doubles rate:
Doubles rate expected/measured:
New normalization constant:
Normalization test—
Measured percent precision:
Required percent precision:
Repeat measurement for at least: 0.000 0.426 1.137 0.004 0.000 280. 681 288. 047 0. 974 1. 000 data quality is inadequate. 0.395 0.300 0.000 Cycle raw data Ending comment: xx Count time (sec): 60.000 Singles 100918 99724 Scaler2 0 0 QC Tests Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: Cvcle R+A 28165 27713 27788 28143 27835 27571 28192 28173 27948 28376 A 10866 10668 Scaler1 1.500 64.000 Pass Pass 10722 10872 10675 10569 10635 10590 1720 50, 0000 100369 100434 Pass Pass O Pass 99856 100108 100178 100266 0. 1620 86. 5000 0. 3458 0. 0299 0. 0299 0. 0000 0. 0001 0. 0001 100210 101249 10712 10861 10 Cycle DTC rate data Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 9. 065 0. 010 0. 000 0. 000 0. 000 0. 101 0. 009 0. 000 Singles 1673. 146 1653. 241 1663. 994 1665. 077 1655. 441 1659. 642 1660. 809 Triples QC Tests 0.000 Pass Doubles 288. 474 284. 237 284. 588 288. 007 286. 155 283. 520 292. 776 293. 209 287. 423 292. 077 Cvcle 60.000 Summed raw data 1662. 276 1661. 343 1678. 665 Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 1003312 279904 107170 10 Summed multiplicity distributions R+A sums A sums Messages Normalization test -- data quality is inadequate

Results

1663. 363 288. 047 0. 000

(1)

144. 167 +-1. 772 +-0. 000 +-103. 407 +-0. 000 +-

Scaler1

6281 6085

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

0. 782 0. 079 0. 000 0. 561 0. 000

QC Tests

QC Tests

Pass

Pass

Pass Pass Pass Pass Pass Pass Pass

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
GC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                                                                                                                                                                                                                                         Results
                                                                                                                                                                                                                                                                                                                                                                                                       Singles:
Doubles:
Triples:
Scaler 1:
Scaler 2:
                                                                               25. 01. 17 11:34:30
51HL3430. VER
                                                                                                                                                                                                                                                                                                                         Passive cycle raw data
                                                                                                                                                                                                                                                                                                                                              Singles
                                                                                                                                                                                                                                                                                                                         Cycle
                                                                               PWR BF
                                                                                                                                                                                                                                                                                                                                                            8782
8658
                                                                               Pu
0.000
Verification
Review disk file
                                                                                                                                                                                                                                                                                                                                                             8498
8661
                                                                                                                                                                                                                                                                                                                                                                                                  189
175
                                                                                                                                                                                                                                                                                                                                                                                                 203
186
178
187
                                                                                                                                                                                                                                                                                                                                                            8683
8499
                                                                                                                                                                                                                                                                                                                                                            8828
8435
                                                                               On
Sample method
Measured
                                                                                                                                                                                                                                                                                                                                9
10
                                                                                                                                                                                                                                                                                                                         Passive cycle rate data
                                                                             Isotopics id:
                                                                                                                                                                                                                                                                                                                                                          Singles
146. 367
144. 300
141. 633
144. 350
144. 717
141. 650
147. 133
140. 583
143. 017
147. 917
    Isotopics iu.
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
                                                                                                                                                                                                                                                                                                                                                                                                       Doubles
1.833
1.617
2.033
                                                                                                                                                                                                                                                                                                                         Cvcle
                                                                                                                                                         0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
25.01.17
                                                                                                                         0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                   0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                                                                                                                   234567
                                                                                                                                                                                                                                                                                                                                                                                                             1. 650
2. 000
1. 717
1. 817
1. 817
2. 033
1. 200
                                                Pu date:
Am241:
Am date:
                                                                                                                          0.0000
                                                                                                                                                                                                     0.0000
                                                                                                                                                           0.0000
25.01.17
                                                                                   1. 50
64. 00
64. 00
1720
50. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
                        Predelay:
Gate length:
2nd gate length:
High voltage:
                             Die away time
Efficiency
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                             0. 0001
                                                                                            1.0000 +-
   Normalization constant:
                                                                                                                                    0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                           0.000 +-
0.000 +-
0.000 +-
0.000
0.000
    Number passive cycles:
Count time (sec):
                                                                                                      10
60
Passive error messages
No passive calibration curve calibration
No known alpha calibration
                                                                                                                                       (1)
```

202501_AFAS-P Bottom Fork(UDL-1).txt

INCC 6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 25.01.17 12:
Results file name: 51HM0407.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Facility: PPFF 12:04:07 Ending comment: xx Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1.500 64.000 1720 50, 0000

0. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 0. 940 0. 010 0. 000 0. 000 0. 000 0. 098 0. 009 0. 000 60.000

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 62108 1040 431

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

102. 573 +-1. 005 +-0. 000 +-

(1)

202501_AFAS-P Bottom Fork(UDL-1).txt

(2)

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test —-0.000 0.167 0.506 0.040 0.000 101. 475 102. 573 0. 989 1. 000 data quality is inadequate.

Cycle raw data

Count time (sec): 60,000 Scaler2 QC O Pass O Pass O Pass O Pass O Pass O Pass Singles 6157 QC Tests Cvcle Α Scaler1 103 104 51 35 39 47 31 36 50 48 42 52 6258 6330 6269 6240 92 105 102 O Pass O Pass O Pass O Pass O Pass O Pass 6101 6272 6092 6308 6081 10

Cycle DTC rate data

Cycle	Singles	Doubles	Triples	QC Tests
1	101. 677	0. 857	0.000	Pass
2	103. 360	1. 140	0.000	Pass
3	104. 560	0. 873	0.000	Pass
4	103. 543	0. 957	0.000	Pass
5	103.060	1. 173	0.000	Pass
6	100. 743	1. 157	0.000	Pass
7	103. 593	1. 007	0.000	Pass
8	100. 593	0. 923	0.000	Pass
9	104. 193	1. 207	0.000	Pass
10	100. 410	0. 757	0.000	Pass

```
INCC 5.1.2
```

INCC 5.1.2

Facility:

Material balance area:
 Detector type:
 Detector id:
 Electronics id:
 Inventory change code:
 I/O code:
 Measurement date:
 Results file name:
 Isnection number:
 Item id:
 Stratum id:
 Material type:
 Original declared mass:
 Measurement option:
 Data source:
 GC tests:
 Error calculation:
 Accidentals method:
 Inspector name:
 Passive comment: 25. 02. 12 10:13:41 52CK1341. VER BWR TF Pu 0.000 Verification Review disk file On Sample method Measured Isotopics id:

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics source code: Pu238: Pu239: Pu240: Pu241: Pu date: Am date: 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +25.02.12
0.0000 +25.02.12 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000

1. 50 64. 00 64. 00 1720 50. 0000 0. 0080 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:

1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 + 0.000 + 0.000 + 0.000 + 0.000 0.000 0.000 0.000

Number passive cycles: Count time (sec): 10 60

Passive error messages

Known alpha analysis error

(1)

202502_AFAS-B_Collar.txt

INCC 5.1.2

Facility: JM2G AFAS

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Litem id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 25. 02. 12 10:26:44 52CK2644. VER BWR COLLAR XXXX Pu 0.000 Verification Review disk file On Sample method

Sample method Measured

Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-25.02.12 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000

(1)

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction 1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 0. 1970 160. 0000 0. 6419 0. 1030 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0.0000 0.000 0.000 0.000

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

10 60

Passive error messages

Known alpha analysis error

Results

59. 478 +-0. 333 +-0. 000 +-105. 415 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Scaler2 0 0 0 0 0 0 0 0 0 Cycle Singles Scaler1 27 29 32 36 30 34 35 33 37 17 14 12 12 12 15 14 14 9 8 6372 6370 6248 6324 6384 3666 3618 3610 3411 3547 3570 3542 3504 3542 3677 6372 6267 8 9 10 6278 6371 6263

Passive cycle rate data

Cycle 1 2 3 4 5	Singles 61.100 60.300 60.167 56.850 59.117	Doubles 0. 167 0. 250 0. 333 0. 400 0. 300	Triples 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000	QC Tests Pass Pass Pass Pass Pass
					Pass
5	59. 117	0. 300	0.000	0.000	Pass
6	59. 500	0. 317	0.000	0.000	Pass
7	59. 033	0. 350	0.000	0.000	Pass
8	58, 400	0. 317	0.000	0.000	Pass
9	59. 033	0. 467	0.000	0.000	Pass
10	61. 283	0. 433	0.000	0.000	Pass

(2)

202502_AFAS-B_Collar.txt

Results

2013. 651 +-429. 578 +-0. 000 +-7. 273 +-0. 000 +-2. 053 1. 401 0. 000 0. 156 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle	Singles	R+A	Α	Scaler1	Scaler2	QC Tests
1	120839	41472	15670	450	0	Pass
2	120590	41329	15760	408	0	Pass
3	120668	41220	15469	472	0	Pass
4	119870	40646	15548	477	0	Pass
5	120779	41283	15329	413	0	Pass
6	121191	41588	15875	429	0	Pass
7	120610	41238	15569	396	0	Pass
8	121016	41348	15408	432	0	Pass
9	121104	41540	15600	472	0	Pass
10	121133	41523	15545	415	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	2014. 635	430. 590	0.000	0.000	Pass
2	2010, 482	426, 700	0.000	0.000	Pass
3	2011, 783	429, 738	0.000	0.000	Pass
4	1998, 474	418, 837	0.000	0.000	Pass
5	2013, 634	433, 126	0.000	0.000	Pass
6	2020, 505	429, 106	0.000	0.000	Pass
7	2010, 815	428. 369	0.000	0.000	Pass
8	2017, 586	432, 894	0.000	0.000	Pass
9	2019, 054	432, 894	0.000	0.000	Pass
10	2019. 538	433, 528	0.000	0.000	Pass

Scaler1

200 206 211

Triples

0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Doubles
1. 333
0. 817
0. 883
1. 267
1. 250
0. 833
1. 517
1. 067
0. 783
1. 100

109. 983 +-1. 085 +-0. 000 +-3. 225 +-0. 000 +-

Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000

QC Tests

QC Tests Pass Pass

Pass

INCC 5.1.2

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                            25. 02. 12 10:45:47
52CK4547. VER
                                                                                            BWR BF
                                                                                            Pu
0.000
Verification
Review disk file
                                                                                            On
Sample method
Measured
                                                                                          Isotopics id:
     Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu date:
Am date:
                                                                                                                                                                                0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
25.02.12
0.0000 +-
25.02.12
                                                                                                                                            0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                 0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                              0.0000
                                                                                                                                                                                                                                   0.0000
                                                                                               1. 50
64. 00
64. 00
1720
50. 0000
0. 0060
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                           Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                          1.0000 +-
   Normalization constant:
                                                                                                                                                        0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                         0.000 +
0.000 +
0.000 +
0.000 +
0.000
                                                                                                                                                         0.000
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                      10
60
Passive error messages
No known alpha calibration
```

202502_AFAS-P_Top Fork.txt

(1)

Facility:

INCC 5.1.2

JM2G AFAS

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
I tem id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 25. 02. 12 11:02:52 52CL0252. VER PWR TF XXXX Pu 0.000 Verification Review disk file On Sample method

Sample method Measured

Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-25.02.12 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0.0000 0.0000

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction 1. 50 64. 00 64. 00 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0.0000 0.000 0.000 0.000

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd:

10 60

Passive error messages

No known alpha calibration

202502_AFAS-P_Top Fork.txt

(2)

Results

Results

Cycle

8 9 10

Cycle

567

8 9 10

Passive cycle raw data

6698 Passive cycle rate data

Singles 110.833 110.517

110. 517 109. 850 110. 500 110. 667 109. 683 110. 383 108. 000 107. 767 111. 633

2. 322 +-0. 000 +-0. 000 +-94. 895 +-0. 000 +-0.065 0.000 0.000 0.280 0.000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle	Singles	R+A	Α	Scaler1	Scaler2	QC Tests
1	128	0	0	5692	0	Pass
2	134	0	0	5714	0	Pass
3	135	0	0	5642	0	Pass
4	133	0	0	5730	0	Pass
5	126	0	0	5802	0	Pass
6	142	0	0	5723	0	Pass
7	133	0	0	5673	0	Pass
8	167	0	0	5687	0	Pass
9	150	0	0	5664	0	Pass
10	145	0	0	5610	0	Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	2. 133	0.000	0.000	0.000	Pass
2	2, 233	0.000	0.000	0.000	Pass
3	2, 250	0.000	0.000	0.000	Pass
4	2, 217	0.000	0.000	0.000	Pass
5	2, 100	0.000	0.000	0.000	Pass
6	2, 367	0.000	0.000	0.000	Pass
7	2, 217	0.000	0.000	0.000	Pass
8	2, 783	0.000	0.000	0.000	Pass
9	2, 500	0.000	0.000	0.000	Pass
10	2. 417	0.000	0.000	0.000	Pass

202502_AFAS-P Top Fork(UDL-1).txt 202502_AFAS-P Top Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 25.02.12 11:
Results file name: 52CL0531.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Gf252 expected singles rate: Gf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 92. 935 93. 687 0. 992 1. 000 11:05:31 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 5721 5728 5735 5585 5786 5737 5750 5663 5717 5626 Scaler2 0 0 QC Tests Pass Pass Pass Cycle Scaler1 Α 39 36 26 39 34 40 31 38 39 83 77 77 78 Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 89 103 96 87 73 8 9 10 Cycle DTC rate data Singles 93. 957 94. 073 94. 190 91. 690 95. 040 94. 223 94. 440 92. 990 93. 890 92. 373 Doubles 0. 762 0. 828 0. 678 0. 845 0. 645 0. 912 1. 045 1. 078 0. 812 0. 562 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 QC Tests Pass Pass Cvcle Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 393 0. 005 0. 000 0. 000 0. 000 234567 Pass Pass Pass Pass 10 60. 000 Pass 8 9 10 Pass Pass Pass Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 57048 848 355 Summed multiplicity distributions R+A sums A sums Messages Normalization test -- data quality is inadequate Results 93. 687 +-0. 817 +-0. 000 +-Singles: Doubles: Triples: 0. 329 0. 053 0. 000 (1) (2) 202502_AFAS-P_Collar.txt 202502_AFAS-P_Collar.txt JM2G AFAS 25. 02. 13 14:44:27 52D04427. VER PWR COLLAR XXXXX Pu 0.000 Verification Review disk file On Theoretical method

INCC	5.	1.	2

INCC 5.1.2

Facility:

Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error cal culation:
Accidentals method:
Inspector name:
Passive comment: Default CO 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-25.02.13

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 1620 86. 5000 0. 3458 0. 0299 0. 0000 0. 0001 0. 0001 Normalization constant: Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 1.0000 +-0.000 +-0.000 +-0.000 +-0.000 0.000

Number passive cycles: Count time (sec):

Passive error messages

No known alpha calibration

Passive	cvcle raw da	Do T Sca	ingles: oubles: riples: aler 1: aler 2:	1646. 504 280. 236 0. 000 14. 405 0. 000	i +- 1 i +- 0 i +- 1	. 020 . 392 . 000 . 200
Cycle 1 2 3 4 5 6 7 8 9	Singles 98841 98702 98821 98955 98218 99177 98262 98789 98866 99131	R+A 27305 26975 27180 27575 27024 27401 26698 27356 27556 27502	A 10441 10439 10352 10369 10417 10501 10336 10467 10571 10633	Scaler1 892 822 924 892 871 825 815 864 847 891	Scaler 2 0 0 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
Passive Cycle 1 2 3 4 5 6 7 8 9 10	Singles 1647, 585 1647, 585 1645, 267 1647, 251 1649, 485 1637, 198 1653, 186 1637, 932 1646, 718 1648, 001 1652, 419	Doub 281. 275. 280. 286. 276. 281. 272. 281. 283. 281.	227 757 626 930 940 828 855 644 245	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Pass

0.0000 0.000 0.000 0.000

0.0000

0.0000 25.02.13

0. 0000 0. 0000 0. 0000 0. 0000 0. 0000

0.0000

```
202502_AFAS-P Collar(UDL-1).txt
 INCC 6.23.2.9 HM32 Not Validated
                   6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 25.02.13 14:
Results file name: 52005038.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
0C tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
                                                                                                                                               14:50:38
                  Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                 1.500
64.000
1720
50.0000
                                                                                                                 0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
                    Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                                        8. 890
0. 000
0. 000
0. 000
0. 000
                                                                                                                                                                           0. 126
0. 004
0. 000
Messages
Normalization test -- data quality is inadequate
Summed raw data
Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 2nd scaler sum:
Summed multiplicity distributions
                                  R+A sums A sums
Number of cycles:
Count time (sec):
                                                                                                                     10
60. 000
Results
                                                                                                                  Singles:
Doubles:
Triples:
                                                                                                                                                       1635. 045
281. 350
0. 000
                                                                                                                                                                                                                    1. 883
1. 154
0. 000
```

202502_AFAS-P_Bottom Fork.txt

0. 0000 0. 0000 0. 0000 0. 0000 0. 0000

0.0000

(1)

INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: INCC 5.1.2 JM2G AFAS 25. 02. 13 15:04:33 52DP0433. VER PWR BF XXXX Pu 0.000 Verification Review disk file On Sample method Measured Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-25.02.13 0.0000 0.0000 25.02.13 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0 0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

10 60

Passive error messages

No passive calibration curve calibration No known alpha calibration

Normalization results for reference source: H4-694

Current normalization constant:
 Cf252 expected doubles rate:
 Cf252 measured doubles rate:
 Doubles rate expected/measured:
 New normalization constant:
 Normalization test - Measured percent precision:
 Required percent precision:
Repeat measurement for at least: 1. 000 275. 271 281. 350 0. 978 1. 000 data quality is inadequate. 0.410 0.300 0.000

Cycle raw data

Count time (sec): 60,000 R+A 27257 27314 27311 27058 27227 27441 27760 Cycle Singles 1 98446 2 98662 Scaler2 QC Tests O Pass O Pass Scaler1 A 10386 10308 10205 10439 10174 10437 10632 10110 10434 10440 Pass Pass 99026 98320 98720 Pass Pass 99290 98007 98519 98540 Pass 26866 27036 27009 0 Pass 0 Pass 0 Pass 8 9 10

Cycle DTC rate data

Doubles 281. 343 283. 595 285. 262 277. 141 284. 378 283. 561 285. 630 279. 424 276. 857 276. 307 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Singles QC Tests Pass Cvcle Singles 1632, 109 1635, 710 1636, 194 1641, 779 1630, 009 1636, 677 1646, 180 1624, 791 1633, 326 1633, 677 Pass 234567 Pass Pass Pass Pass Pass 8 9 10 Pass Pass Pass

(2)

202502_AFAS-P_Bottom Fork.txt

Results

141. 260 +-1. 983 +-0. 000 +-101. 353 +-0. 000 +-Singles: Doubles: Triples: 0. 494 0. 081 0. 000 0. 527 0. 000 Scaler 1: Scaler 2:

Passive cycle raw data

Scaler2 0 0 0 0 0 0 0 0 0 Singles 8417 Cvcle R+A 188 192 204 171 201 176 188 223 186 181 Scaler1 QC Tests 67 74 78 79 57 67 63 86 72 77 6120 6043 6215 6108 5880 6199 5967 6089 6078 6113 8566 8570 8304 8505 8453 8368 8594 8518 8461 8 9 10

Passive cycle rate data

Mass OC Tests
0.000 Pass
0.000 Pass Doubles 2. 017 1. 967 2. 100 1. 533 2. 400 1. 817 2. 083 2. 283 1. 900 1. 733 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 Cycle Singles 140, 283 142, 767 142, 833 138, 400 141, 750 140, 883 139, 467 143, 233 141, 967 141, 017

202502_AFAS-P Bottom Fork(UDL-1).txt 202502_AFAS-P Bottom Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated 6. 23. 2. 9 HM32 NOT VAIIDATED

Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 25. 02. 13 15:
Results file name: 52PP0634. NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Normalization results for reference source: H4-694 Current normalization constant: Gf252 expected singles rate: Gf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test — 1. 000 99. 519 100. 340 0. 992 1. 000 data quality is inadequate Cycle raw data Count time (sec): 60.000 Singles 6134 6133 5992 6207 6117 6027 6035 6090 6042 6017 Scaler2 0 0 R+A 105 QC Tests Cycle Α Scaler1 36 52 36 46 37 29 36 47 37 32 Pass 88 100 Pass Pass Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1. 500 64. 000 1720 50. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001 O Pass 108 105 84 98 112 107 8 9 10 109 Cycle DTC rate data Singles 101, 250 101, 233 98, 883 102, 467 100, 967 99, 467 99, 600 100, 517 99, 717 99, 300 Triples
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000
0.000 QC Tests Pass Pass Cvcle Doubles 1.150 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0. 983 0. 000 0. 000 0. 000 0. 000 1. 150 0. 600 1. 067 1. 033 1. 133 0. 917 1. 033 1. 083 1. 167 1. 283 0. 053 0. 004 0. 000 234567 Pass Pass Pass Pass Pass 8 9 10 Messages Pass Pass Pass Normalization test -- data quality is inadequate Summed raw data Number of good cycles: Total count time: Shift register singles sum: Shift register accidentals sum: Shift register accidentals sum: Shift register 1st scaler sum: Shift register 2nd scaler sum: 60794 1016 388 Summed multiplicity distributions R+A sums A sums Number of cycles: Count time (sec): 10 60. 000 Results Singles: Doubles: Triples: 100. 340 1. 047 0. 000 0. 361 0. 059 0. 000 (1) (2) 202503_AFAS-B_Top Fork.txt 202503_AFAS-B_Top Fork.txt INCC 5.1.2 Results Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory charge code: 59. 478 +-0. 282 +-0. 000 +-104. 185 +-0. 000 +-0. 211 0. 038 0. 000 0. 241 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2: JM2G AFAS Detector Id.
Electronics id.
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id.
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: Passive cycle raw data 25. 03. 07 11:03:22 537L0322. VER Singles 3547 3601 3635 3577 3497 3578 3604 3527 3565 Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 6273 6311 6219 6249 6264 6237 6147 6271 6243 6297 Cvcle QC Tests Pass Pass BWR TF XXXX Pu 0.000 Verification Review disk file On 28 32 26 20 26 37 28 41 32 18 17 14 10 10 6 11 16 9 Sample method Measured 9 10 3565 3556 Passive cycle rate data Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Singles 59. 117 60. 017 60. 583 59. 617 58. 283 59. 633 60. 067 58. 783 59. 417 59. 267 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Doubles 0. 167 0. 167 0. 250 0. 200 0. 167 0. 267 0. 517 0. 283 0. 417 QC Tests Pass Pass Isotopics id: Cvcle Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 Pass Pass Pass Pass Pass Pass Pass 25 03 07 0.0000 0.0000 0.0000 25.03.07 10 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Triples gate fraction: 1. 50 64. 00 64. 00 1720 50. 0000 0. 0080 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 Normalization constant: 1 0000 +-0 0000 0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

Passive error messages

Known alpha analysis error

10 60

```
INCC 5.1.2
                                                                                                          Results
```

INCC 5.1.2

Facility:

Material balance area:
 Detector type:
 Detector id:
 Electronics id:
 Inventory change code:
 I/O code:
 Measurement date:
 Results file name:
 Isnection number:
 Item id:
 Stratum id:
 Material type:
 Original declared mass:
 Measurement option:
 Data source:
 GC tests:
 Error calculation:
 Accidentals method:
 Inspector name:
 Passive comment: 25. 03. 07 09:43:59 537J4359. VER BWR COLLAR

Pu 0.000 Verification Review disk file On Sample method Measured

Isotopics id:

Default 0D 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-00.01.01 0.0000 +-00.01.01 Isotopics id: Isotopics source code: Pu238: Pu239: Pu240: Pu241: Pu date: Am date: 0.0000 +0.0000 +100.0000 +0.0000 +0.0000 +25.03.07 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 25.03.07

1. 50 64. 00 64. 00 1720 50. 0000 0. 1970 160. 0000 0. 6419 0. 1030 0. 0000 0. 0001 0. 0001 Predelay: Gate length: 2nd gate length: High voltage: Die away time: Efficiency: Efficiency: Multiplicity deadtime: Coefficient A deadtime: Coefficient B deadtime: Coefficient C deadtime: Doubles gate fraction: Triples gate fraction:

1.0000 +-Normalization constant: 0.0000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 +-0.000 +-0.000 +-0.000 0.000 0.000 0.000 0.000

Number passive cycles: Count time (sec): 10 60

Passive error messages

Known alpha analysis error

(1)

202503_AFAS-B_Bottom Fork.txt

Facility:

INCC 5.1.2

JM2G AFAS

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
I tem id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
OC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment: 25. 03. 07 09:57:49 537J5749. VER BWR BF XXXX Pu 0.000 Verification Review disk file On Sample method

Sample method Measured

Isotopics id: Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date: 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 +-0.0000 +-100.0000 +-0.0000 +-0.0000 +-25.03.07 0. 0000 0. 0000 0. 0000 0. 0000 0. 0000 0.0000 0. 0000 25. 03. 07 0.0000

(1)

Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction 1. 50 64. 00 64. 00 1720 50. 0000 0. 0060 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001 0. 0001

Normalization constant: 1 0000 +-0.0000

0.000 +-0.000 +-0.000 +-0.000 0.000 Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 0.000 0.000 0.000

10 60

Passive error messages

No known alpha calibration

1981. 498 +-422. 082 +-0. 000 +-7. 063 +-0. 000 +-Singles: Doubles: Triples: Scaler 1: Scaler 2: 2. 770 1. 837 0. 000 0. 102 0. 000

Passive cycle raw data

Scaler2 0 0 0 0 0 0 0 0 0 Singles 119042 118168 117853 Cycle Scaler1 R+A 40687 40170 39479 40687 40336 40613 40152 40533 40612 40670 15047 14558 14972 15123 14929 15173 15203 15336 15392 15279 426 407 429 404 403 451 448 418 405 447 119191 119223 118958 119239 119346 119140 8 9 10

Passive cycle rate data

es Mass QC Tests
00 0.000 Pass

(2)

202503_AFAS-B_Bottom Fork.txt

Results

108. 053 +-1. 265 +-0. 000 +-3. 090 +-0. 000 +-0. 346 0. 107 0. 000 0. 101 0. 000 Singles: Doubles: Triples: Scaler 1: Scaler 2:

Passive cycle raw data

Cycle 1 2 3 4	Singles 6466 6441 6477 6518	R+A 111 150 102 116	A 48 53 38 50	Scaler1 198 185 161 165	Scaler2 0 0 0 0	QC Tests Pass Pass Pass Pass
3	6477	102	38	161	0	Pass
5	6420	126	40	163	0	Pass Pass
6	6482 6446	133 95	37 39	180 194	0	Pass Pass
8	6475	116	53	221	Ŏ	Pass
9 10	6654 6453	118 138	63 25	202 185	0	Pass Pass

Passive cycle rate data

Cycle	Singles	Doubles	Triples	Mass	QC Tests
1	107. 767	1. 050	0.000	0.000	Pass
2	107, 350	1, 617	0.000	0.000	Pass
3	107, 950	1. 067	0.000	0.000	Pass
4	108, 633	1, 100	0.000	0.000	Pass
5	107, 000	1, 433	0.000	0.000	Pass
6	108, 033	1, 600	0.000	0.000	Pass
7	107, 433	0. 933	0.000	0.000	Pass
8	107, 917	1. 050	0.000	0.000	Pass
9	110, 900	0.917	0.000	0.000	Pass
10	107. 550	1. 883	0.000	0.000	Pass

Results

```
202503_AFAS-P_Top Fork.txt
INCC 5.1.2
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                     25. 03. 07 10:12:54
537K1254. VER
                                                                                     BWR TF
                                                                                     Pu
0.000
Verification
Review disk file
                                                                                    On
Sample method
Measured
                                                                                   Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu445:
                                                                                                                                                                     0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
25.03.07
                                                                                                                                  0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                  0.0000
0.0000
0.0000
0.0000
0.0000
                                                    Pu date:
Am241:
Am date:
                                                                                                                                   0.0000
                                                                                                                                                                                                                   0.0000
                                                                                                                                                                      0.0000
25.03.07
                                                                                         1. 50
64. 00
64. 00
1720
50. 0000
0. 0126
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                         Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                  1.0000 +-
   Normalization constant:
                                                                                                                                              0.0000
  Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                  0.000 +
0.000 +
0.000 +
0.000 +
0.000
     Number passive cycles:
Count time (sec):
Passive error messages
No known alpha calibration
```

Passive cycle raw data Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 5641 5588 5560 5762 5645 5640 5635 5625 5674 5593 Cycle Singles 109 127 147 107 127 125 140 145 122 145 8 9 10 Passive cycle rate data

2. 157 +-0. 000 +-0. 000 +-93. 938 +-0. 000 +-

Singles: Doubles: Triples: Scaler 1: Scaler 2:

Singles 1.817 2.117 2.450 Cycle Doubles Triples QC Tests Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 Pass Pass 450 783 117 Pass Pass 2. 117 2. 083 2. 333 2. 417 2. 033 2. 417 Pass Pass Pass Pass 567 8 9 10 Pass Pass

202503_AFAS-P Top Fork(UDL-1).txt

(1)

```
INCC 6.23.2.9 HM32 Not Validated
```

Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P TOP
Electronics id: UDL-1
Measurement date: 25, 03, 07 10:
Results file name: 537K1924.NOR
Inspection number:
Measurement option: Normalization
Data source: IAFA DataZ file
Detector configuration: Passive
0 C tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: XX Facility: PPFF 10:19:24 Ending comment: xx

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicy deadtime:
Coefficient B deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction: 1.500 64.000 1720 50, 0000 0. 0000 0. 0126 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: Number of cycles: Count time (sec): 1. 415 0. 007 0. 000 0. 000 0. 000 0. 032 0. 006 0. 000

60.000

Summed raw data Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 56103 837 296

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

92. 090 0. 895 0. 000

(1)

202503_AFAS-P Top Fork(UDL-1).txt

(2)

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test —-0. 000 0. 204 0. 443 0. 040 0. 000 91. 416 92. 090 0. 993 1. 000 data quality is inadequate.

Cycle raw data

Count time (sec): 60,000 Scaler2 QC O Pass O Pass O Pass O Pass O Pass O Pass Singles 5659 QC Tests Cvcle R+A Α Scaler1 86 101 68 89 100 28 30 28 25 29 40 41 26 27 22 5722 5538 5648 5732 5590 5569 5546 5631 5468 O Pass O Pass O Pass O Pass O Pass O Pass 67 73 80 88 85 10

Cycle DTC rate data

Singles 92. 902 93. 952 90. 885 92. 718 94. 118 91. 752 91. 402 91. 018 92. 435 89. 718 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Doubles 0.960 1.177 0.660 QC Tests Pass Pass Cycle 1. 060 1. 177 0. 443 0. 527 0. 893 1. 010 1. 043 Pass Pass Pass Pass Pass Pass 10

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Electronics id:
Inventory change code:
1/0 code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                             25. 03. 07 10:31:25
537K3125. VER
                                                                                             PWR COLLAR
                                                                                             Pu
0.000
Verification
Review disk file
                                                                                             On
Sample method
Measured
     Isotopics id:
Isotopics source code:
Pu238:
Pu240:
Pu241:
Pu242:
Pu date:
Am241:
Am date:
                                                                                          0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
25.03.07
                                                                                                                                              0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                    0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                   0.0000 +-
25.03.07
                                                                                                                                               0.0000
                                                                                                                                                                                                                                      0.0000
 Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
Coefficient B deadtime:
Coefficient Geadtime:
Toubles gate fraction:
Triples gate fraction:
                                                                                                1. 50
64. 00
64. 00
1720
50. 0000
0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
                                                                                                           1.0000 +-
   Normalization constant:
                                                                                                                                                         0.0000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                          0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                       10
60
Passive error messages
No passive calibration curve calibration
No known alpha calibration
```

202503_AFAS-P Collar(UDL-1).txt

(1)

```
INCC 6.23.2.9 HM32 Not Validated
```

```
Facility: PPFF
Material balance area: XXXX
Detector type: PASS
Detector id: AFAS-P COLLAR
Electronics id: UDL-1
Measurement date: 25.03.07 10:31:59
Results file name: 537K3159.NOR
Inspection number: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
Detector configuration: Passive
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx
 Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient C deadtime:
Doubles gate fraction:
Triples gate fraction:
                                                                                                                                                                                             1.500
64.000
                                                                                                                                                                                        1720
50, 0000
```

```
50. 0000
0. 1620
86. 5000
0. 3458
0. 0299
0. 0000
0. 0001
0. 0001
Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
Number of cycles:
Count time (sec):
                                                                                                                                                              9. 265
0. 007
0. 000
0. 000
0. 000
                                                                                                                                                                                                                                                  0. 124
0. 006
0. 000
                                                                                                                                                           60.000
```

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum:
Shift register 1st scaler sum: 600 971337 265973 101267

Summed multiplicity distributions

R+A sums A sums

Messages

Normalization test -- data quality is inadequate

Results

1609. 857 +-274. 657 +-0. 000 +-

Results

Singles: Doubles: Triples: Scaler 1: Scaler 2: 1619. 382 +-274. 212 +-0. 000 +-13. 957 +-0. 000 +-2. 238 1. 283 0. 000 0. 128 0. 000

Passive cycle raw data

Cycle 1 2 3 4 5 6 7 8 9	Singles 96712 97248 97358 96798 97786 96971 97845 96598 97192 96985	R+A 26363 26789 26282 26062 26651 26315 27130 26358 26620 26683	A 10092 10170 10083 9894 10168 10013 10166 9903 10275 10054	Scaler1 852 807 825 823 851 812 815 860 880 849	Scaler2 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	QC Tests Pass Pass Pass Pass Pass Pass Pass P
-------------------------	---	---	---	---	---	---

Passive cycle rate data

Cycle 1 2 3 4 5 6 7 8	Singles 1612.091 1621.027 1622.861 1613.525 1629.996 1616.409 1630.980 1610.191 1620.094	Doubles 271. 335 277. 139 270. 135 269. 617 274. 872 271. 852 282. 893 274. 403 272. 569	Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000	Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000	QC Test: Pass Pass Pass Pass Pass Pass Pass Pas

202503_AFAS-P Collar(UDL-1).txt

(2)

Normalization results for reference source: H4-694

Current normalization constant:
Cf252 expected doubles rate:
Cf252 measured doubles rate:
Doubles rate expected/measured:
New normalization constant:
Normalization test—
Measured percent precision:
Required percent precision:
Repeat measurement for at least: 0. 000 0. 411 1. 452 0. 005 0. 000 270. 995 274. 657 0. 987 1. 000 1.000 +- 0.000 data quality is inadequate. 0.529 0.300 0.000

Cycle raw data

Count time (sec): 60.000

Cycle	Singles	R+A	Α	Scaler1	Scaler2 QC Tests
1	97162	26884	10063	0	0 Pass
2	96738	26421	10127	0	0 Pass
3	96657	26644	9918	0	0 Pass
4	97355	26599	10352	0	0 Pass
5	97278	26452	10121	0	0 Pass
6	97248	26426	9946	0	0 Pass
7	97602	26371	10268	0	0 Pass
8	97125	26946	10140	0	0 Pass
9	97161	26660	9942	0	0 Pass
10	97011	26570	10390	0	0 Pass

Cycle DTC rate data

-				
Cycle	Singles	Doubles	Triples	QC Tests
1	1610.328	280.500	0.000	Pass
2	1603.260	271.711	0.000	Pass
3	1601.909	278.915	0.000	Pass
4	1613. 546	270. 929	0. 000	Pass
5	1612. 262	272. 329	0. 000	Pass
6	1611. 762	274. 814	0. 000	Pass
7	1617. 664	268. 528	0. 000	Pass
8	1609. 712	280. 250	0. 000	Pass
9	1610. 312	278. 783	0. 000	Pass
10	1607. 811	269. 811	0. 000	Pass

(2)

```
INCC 5.1.2
```

```
INCC 5.1.2

Facility:
Material balance area:
Detector type:
Detector id:
Inventory change code:
I/O code:
Measurement date:
Results file name:
Inspection number:
Item id:
Stratum id:
Material type:
Original declared mass:
Measurement option:
Data source:
QC tests:
Error calculation:
Accidentals method:
Inspector name:
Passive comment:
                                                                                                25. 03. 07 11:21:38
537L2138. VER
                                                                                                PWR BF
                                                                                                Pu
0.000
Verification
Review disk file
                                                                                               On
Sample method
Measured
                                                                                             Default

0D

0.0000 +-

0.0000 +-

100.0000 +-

0.0000 +-

0.0000 +-

00.01.01

0.0000 +-

00.01.01
      Isotopics id:
Isotopics source code:
Pu238:
Pu239:
Pu240:
Pu241:
Pu242:
Pu44tc
                                                                                                                                                                                          0.0000 +-
0.0000 +-
100.0000 +-
0.0000 +-
0.0000 +-
25.03.07
                                                                                                                                                   0.0000
0.0000
0.0000
0.0000
0.0000
                                                                                                                                                                                                                                             0.0000
0.0000
0.0000
0.0000
0.0000
                                                          Pu date:
Am241:
Am date:
                                                                                                                                                    0.0000
                                                                                                                                                                                                                                               0.0000
                                                                                                                                                                                            0.0000
25.03.07
                                                                                                    1. 50
64. 00
64. 00
1720
50. 0000
0. 0127
0. 0000
0. 0000
0. 0000
0. 0000
0. 0001
0. 0001
                            Predelay:
Gate length:
2nd gate length:
High voltage:
Die away time:
Efficiency:
  Efficiency:

Multiplicity deadtime:

Coefficient A deadtime:

Coefficient B deadtime:

Coefficient C deadtime:

Doubles gate fraction:

Triples gate fraction:
                                                                                                               1.0000 +-
   Normalization constant:
                                                                                                                                                                0.0000
 Passive singles bkgrnd:
Passive doubles bkgrnd:
Passive triples bkgrnd:
Passive scaler1 bkgrnd:
Passive scaler2 bkgrnd:
                                                                                                              0.000 +-
0.000 +-
0.000 +-
0.000
0.000
     Number passive cycles:
Count time (sec):
                                                                                                                            10
60
Passive error messages
No passive calibration curve calibration
No known alpha calibration
```

(1)

202503_AFAS-P Bottom Fork(UDL-1).txt INCC 6.23.2.9 HM32 Not Validated

Facility: PPFF
Material balance area: XXXX
Detector type: UDL-1
Detector id: AFAS-P BOTTOM
Electronics id: PASS
Measurement date: 25.03.07 11:
Results file name: 537L2416.NOR
Inspection number:
Measurement option: Normalization
Data source: IAEA DataZ file
Detector configuration: Passive
OC tests: On
Accidentals method: Measured
Inspector name:
Comment:
Ending comment: xx Facility: PPFF 11:24:16

Ending comment: xx

Predelay:
Gate length:
High voltage:
Die away time:
Efficiency:
Multiplicity deadtime:
Coefficient A deadtime:
*Coefficient B deadtime:
*Coefficient G deadtime:
Doubles gate fraction:
Triples gate fraction: Predelay: 1.500 64.000 1720 50, 0000 0. 0000 0. 0127 0. 0000 0. 0000 0. 0000 0. 0000 0. 0001

Passive singles bkgrnd: Passive doubles bkgrnd: Passive triples bkgrnd: Passive scaler1 bkgrnd: Passive scaler2 bkgrnd: 1. 082 0. 007 0. 000 0. 000 0. 000 0. 026 0. 006 0. 000

Messages

Normalization test -- data quality is inadequate

Summed raw data

Number of good cycles:
Total count time:
Shift register singles sum:
Shift register accidentals sum:
Shift register lst scaler sum:
Shift register lst scaler sum:
Shift register 2nd scaler sum: 600 60043 949 354

Summed multiplicity distributions

R+A sums A sums Number of cycles: Count time (sec): 10 60. 000

Results

98. 990 0. 985 0. 000

(1)

Results

Singles: Doubles: Triples: Scaler 1: Scaler 2: 139. 468 +-1. 797 +-0. 000 +-100. 162 +-0. 000 +-0. 440 0. 084 0. 000 0. 388 0. 000

Passive cycle raw data

Scaler2 0 0 0 0 0 0 0 0 0 Scaler1 6100 5978 6078 5929 Singles Cycle QC Tests 68 76 78 95 76 72 74 74 8443 8424 8347 8335 8406 8373 184 171 5985 5945 205 204 5938 6104 9 10 168 196

Passive cycle rate data

Singles 137. 117 141. 617 140. 717 140. 400 139. 117 140. 100 139. 550 139. 667 137. 483 Cvcle Doubles 1. 700 Triples QC Tests Mass 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 0. 000 1. 833 1. 567 1. 817 1. 483 1. 583 2. 217 2. 167 1. 567 2. 033 Pass 234567 Pass Pass Pass Pass Pass Pass Pass Pass 8 9 10

(2)

202503_AFAS-P Bottom Fork(UDL-1).txt

Scaler 1: Scaler 2:

Normalization results for reference source: H4-694

Current normalization constant: Cf252 expected singles rate: Cf252 measured singles rate: Singles rate expected/measured: New normalization constant: Normalization test —-0. 000 0. 161 0. 493 0. 040 0. 000 97. 971 98. 990 0. 990 1. 000 data quality is inadequate.

Cycle raw data

Count time (sec): 60,000 Scaler2 QC O Pass O Pass O Pass O Pass O Pass O Pass Singles 6068 QC Tests Cvcle R+A Α Scaler1 89 91 95 106 86 42 47 26 36 30 34 44 28 31 36 6071 5956 5963 5977 O Pass O Pass O Pass O Pass O Pass O Pass 5861 6143 5871 6072 6061 84 96 98 92 112 10

Cycle DTC rate data

Singles 100. 052 100. 102 98. 185 98. 302 98. 535 96. 602 101. 302 96. 768 100. 118 99. 935 Triples 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 0.000 Doubles 0. 777 0. 727 1. 143 1. 160 0. 927 0. 827 0. 860 1. 160 QC Tests Pass Pass Cycle Pass Pass Pass Pass Pass Pass 1. 010 1. 260 10

【IPCA2 性能確認試験】

- (1) 4.1 長期管理限界の妥当性確認
- (2) 4.2 ハードウェアの修理及び動作確認
- (3) 4.3 測定パラメータの評価

Annual IPCA2 Performance Report for JFY24 LA-CP-25-20218

 $Prepared \ for:$

Japan Atomic Energy Agency

Prepared by:

C.D. Rael, W. Geist, A. Lintereur Tresong

Safeguards Science and Technology Group (NEN-1)

Los Alamos, NM 87545 USA

March 2025

Controlled by: Carlos Rael, LANL/NEN-1, cdr@lanl.gov

This page is intentionally left blank.

Contents

Acr	ronyms	۷
	Overview	
	Plutonium Efficiency	
	2.1. Efficiency Monitoring	
	2.1. Efficiency Dependence on Environmental Conditions	
3.	AmLi Stability	9
4.	Curium Stability	10
5.	Power Supply Repair	14
6.	HPGe System Performance	15
7.	Load Cell Data	18
8.	Continuous Background Monitoring	18
9.	Operating Parameter Review	23
10.	Summary	25
11.	References	26

Acronyms

cps	counts per second
HPGe	High Purity Germanium

IAEA International Atomic Energy Agency ICS Integrated Cryocooling System

IPCA2 Improved Plutonium Canister Assay System 2

JAEA Japan Atomic Energy Agency JSGO Japan Safeguards Office

LANL Los Alamos National Laboratory

MIC Multi-Instrument Collect

NMCC Nuclear Material Control Center

1. Overview

This report summarizes the results of the monthly control measurements of the Improved Plutonium Canister Assay System 2 (IPCA2) performed over the period of November 2024 through March 2025 and presents the performance overview for JFY24. Monthly measurements were replaced by periodically measurement due to the abbreviated timeframe allowed since the contract approval. Measurements of plutonium neutron detection efficiency and curium stability were performed and analyzed. In addition to these measurements, the operating parameter were confirmed through a series of measurements, which are documented in this report. According to an earlier contractual agreement, the consensus was reached to discontinue HPGe control measurements starting in JFY21 as the HPGe gamma spectra showed deteriorated performance on all three gamma detectors [1].

All the JFY24 results are shown with respect to the control bounds established based on historical data collected May 2013 through March 2024; the use of AmLi was discontinued starting in JFY20. The Pu efficiency measurements demonstrated that the performance of the IPCA2 during this reporting period was stable within 2 σ level, apart from one measurement. The average Pu efficiency for the JFY24 period was 7.28% \pm 0.07. Measurements were compared to room temperature and humidity and no dependence was observed. In May 2022, IAEA provided the R7780 device (a Shift Register, Multiplicity and Time Recorder), colloquially known as the UDL1 for testing and evaluation. Following rigorous performance evaluation summarized in [2], the UDL1 was installed at IPCA2 for testing during regular IPCA2 control measurements.

Repairs were successfully made to the IPCA2 Power Supply Box. A new commercial off the shelf power supply was purchased and installed. The new power supply was a direct replacement, so no changes to the electrical drawings are necessary. The repair fixed the other suspected issues with JSR-15 and Combo Box. A detailed discussion on the repairs can be found in section 5.

2. Plutonium Efficiency

2.1. Efficiency Monitoring

Plutonium efficiency measurements were performed between November 2024 and March 2025. The JSR15, the standard shift register used on IPCA2, results are reported along with UDL1 results. The LANL plutonium standard, FZC158 (823.6 neutrons s⁻¹ emission rate), was used in all measurements. This source was placed in the IPCA2 for a duration of 3600 s during which 120 cycles of 30 s were used to calculate a Singles rate (in counts per second, cps). The Singles rate was divided by the source emission rate to determine an efficiency, as shown in Figure 1. Average efficiency corresponding to the JFY24 control period was calculated and corresponds to 7.28% \pm 0.07% for the JSR15 data. All JSR15 measurements were within the control chart 2σ bands, denoted with dotted lines, except for one as shown in Figure 1. The one measurement that is outside of 2σ occurred on March 4, 2025. At this time, the HPGe detectors had been removed, therefore changing the neutron scattering profile. For all the measurements the UDL1 and JSR15 results agreed within statistical uncertainties. The UDL1 measurements were acquired using multiplicity mode setting and data was analyzed using LANL INCC 6. The IPCA2 measurements on March 4, 2025, were performed during the JMOX meeting at LANL with JAEA and JSGO attendance.




Figure 1: Pu efficiency measurements for JFY24 acquired using standard IPCA2 data acquisition system (blue squares) and UDL1 Multiplicity mode acquisition (black squares).

These results in Figure 1 provide an overview of the overall IPCA2 performance during the JFY24 reporting period and confirm its reproducibility.

All the Pu efficiency measurements since May 2013 are summarized in Figure 2. Note that these measurements are plotted against the original control bounds established from data taken during the 2013 to 2017 time period [3].




Figure 2: Pu efficiency measurements for May 2013 – March 2024 acquired using standard IPCA2 data acquisition system (blue squares) and UDL1 list mode acquisition (black squares).

2.1. Efficiency Dependence on Environmental Conditions

Room temperature and humidity data has been collected alongside IPCA2 periodic measurements. JFY24 plutonium efficiency measurements exhibit no dependence on humidity, Figure 3, or room temperature, Figure 4. Since Los Alamos has experienced a dry winter season, the humidity has been very stable at 16% for all the measurements performed. No correlation in the efficiency as a function of temperature was observed.




Figure 3: JFY24 Pu efficiency measurements as a function of humidity.




Figure 4: JFY24 Pu efficiency measurements as a function of room temperature.

3. AmLi Stability

As reported in [1], AmLi stability measurements were discontinued during JFY21 reporting period. For completeness, an overview of all the historical decay corrected AmLi stability data from 2013 to 2021 is shown in Figure 5 against the original control bounds established in [3]. Note that the decay correction on the AmLi data is with respect to 01/12/2017 when the original control bounds were established.




Figure 5: Historical AmLi stability measurements from 2013 to 2021.

4. Curium Stability

In October 2018 LANL started performing monthly measurements with a curium source to evaluate its feasibility as a potential replacement for the AmLi stability measurements. The decay corrected (with respect to the first measurement on 10/15/2018) results of the curium measurements performed over the JFY24 reporting period are summarized in Figure 6 and correspond to an average count rate of 983.9 ± 4.4 cps. Standard JSR15 data are compared with UDL1 measurements acquired using the multiplicity mode setting and analyzed using INCC6 software. The average count rate for the UDL1 was 983.1 ± 5.1 cps. There are four measurements on or outside of the 3 σ line. During these measurements the HPGe detectors had been removed for IPCA2 detector realignment repair. Removing these detectors changed the neutron scattering profile, resulting in lower count rates. The scattering effect is more pronounced when compared to the Plutonium because the Curium source is more like a point source. The IPCA2 measurements in March 2025, were performed during the JMOX meeting. Note that the JSR15 data (green diamonds) may be hidden behind the UDL-1 data (black squares).

An overview of all decay corrected Cm stability data from October 2018 is shown in Figure 7. Curium stability was also evaluated as a function of humidity and temperature as shown in Figure

8 and Figure 9, respectively. No dependence on humidity and room temperature has been observed over the reporting period.




Figure 6: Curium count rates for JFY24 acquired using standard IPCA2 data acquisition system (green diamonds) and UDL1 list mode acquisition (black squares). Note that error bars are smaller than the size of the symbols.




Figure 7: Curium stability measurements from October 2018 – March 2025.




Figure 8: JFY24 Curium count rates as a function of humidity.




Figure 9: JFY24 Curium count rates as a function of room temperature.

5. Power Supply Repair

As noted in last year's annual report [5], we noted a degradation in the data in July 2023. Upon troubleshooting the issue, it was determined that the UPS had failed causing inadequate AC voltage to the power supply box causing it eventually to fail. In January 2025 a new power supply was installed in the Power Supply box and was connect to facility AC power. This repair fixed the degradation in data discovered during JFY23 control measurements[5]. Pictures are shown below of the internals of the power supply box and the failed power supply. The dark coloring on the circuit board indicates that the power supply failed on the AC side of the circuit confirming the cause of failure was an inadequate AC power source from the UPS. The UPS will not be replaced because IPCA2 will be shipped to Japan in early JFY25.




Figure 10: Power supply internal view and failed power supply

6. HPGe System Performance

No work has been performed on the HPGe detector. For transparency and continuity of knowledge the following is an excerpt from JFY22's annual report.

Based on the updated contract, HPGe control measurements were discontinued during JFY21 as the gamma spectra showed deteriorated performance on all three gamma detectors [1]. For completeness, the present status of HPGe system hardware components is shown in Figure 10.




Figure 11: Configuration of IPCA2 HPGe system during JFY22.

The cause of the deterioration of the gamma resolution is likely associated with cooling interruptions due to the XCoolers failures observed throughout the years. Significant efforts were dedicated to resurrection of HPGe cooling system during 2018 and were reported in [4]. One of the potential reasons for deteriorating resolution could be buildup of surface dirt inside the detector housing, which could potentially be treated by a thermal cycling when a detector is allowed to warm up to a room temperature for sufficient amount of time (2-3 days) to allow for all the detector components to assume room temperature. This warm-up stage is then followed by normal cooling of the detector system. The thermal cycling process can help remove accumulated contamination and restore the energy resolution. Results of these trials are available in the JFY20 report [1]. During JFY22 another review of HPGe system was performed, molecular sieve and o-ring were closely inspected, cleaned and vacuum grease was reapplied as shown in Figure 11. The system is currently under further evaluation at LANL. To fully resolve the observed issues it was recommended to consider sending one of the HPGe detectors back to the manufacturer for evaluation and repair or plan for discarding the system due to planned replacement by ICS.










Figure 12: IPCA2 HPGe detectors review and cleaning; (top left-to-right) cooling head with molecular sieve in place, molecular sieve inspection, cooling head with molecular sieve removed; (bottom) HPGe detector junction box before cleaning.

LANL has received the IAEA ICS system for testing with IPCA2 during 2021. During the initial tests to evaluate the ICS overall status and performance, there was no measurable signal from the instrument with the gamma source located in close proximity and power and HV on. Following an appropriate cooling period, as soon as HV was applied, only low level noise was observed on the output signal coming from the detector. Following a consultation with IAEA, the ICS was sent for

repair to ORTEC. ORTEC is currently in the final stages of fabrication and testing of a replacement unit, which is planned to be installed at IPCA2 once shipped back to LANL.

7. Load Cell Data

Regular load cell measurements were performed during January 2025 to March 2025. The short time period of load cell measurements is the result of the delay in contract approval and the LANL facility crane was out of service. Each of these measurements resulted in a consistent weight of 22.69 kg.

8. Continuous Background Monitoring

As part of the contractual agreement, continuous neutron system background was acquired for IPCA2. The measurements were performed using MIC3 software and analyzed with Omniscient, LANL developed software. IPCA2 was continuously running from April 2024 to January 2025. Measurements were stopped beginning in February to allow time for detector repairs and initial preparations for shipment in JFY 25.

The continuous background results include the period between April 2024 and January 2025. Singles count rates over the reporting period are shown for each month in Figures 13 through 23 for JSR15 data. Overall, the Singles background exhibits regular variation between approximately 25 – 32 cps, which can be attributed to variation in cosmic ray background. The irregular high count rate spikes and intervals seen in Singles background correspond to various measurements that are occasionally performed in the High Bay area, where IPCA2 is located. Note that the High Bay area is used as a testing location for other LANL developed instrumentation and experiments are routinely performed throughout the year. Additional features are further detailed in the figure captions.




Figure 13: IPCA2 neutron background Singles during April 2024. Spikes correspond to use of sources in the IPCA2 location.




Figure 14: IPCA2 neutron background Singles during May 2024.




Figure 15: IPCA2 neutron background Singles during June 2024.




Figure 16: IPCA2 neutron background Singles during July 2024.




Figure 17: IPCA2 neutron background Singles during August 2024.




Figure 18: IPCA2 neutron background Singles during September 2024.




Figure 19: IPCA2 neutron background Singles during October 2024.




Figure 20: IPCA2 neutron background Singles during November 2024.




Figure 21: IPCA2 neutron background Singles during December 2024.




Figure 22: IPCA2 neutron background Singles during January 2024. There is a pause in data at the on about January 16for NMCC Training activities.

9. Operating Parameter Review

A series of measurement were performed to validate IPCA2's performance and provide confirmation of its operating parameters. These measurements included confirmation of optimum HV setting, efficiency, gate setting and die-away time. Table 1 shows the current operating parameters used in IPCA2 measurements.

Table 1, Current IPCA2 operating parameters.

Operating voltage [V]	1720
Predelay [μs]	3
Gate [µs]	24
Deadtime Parameters	
A	0.2065
В	0.0426
С	0
Multiplicity Deadtime	0
Efficiency (²⁵² Cf) [%]	7.19

Other measurements made to validate IPCA2's performance were:

• HV plateau




Figure 23: HV plateau for each of the signals, the Operating Voltage is confirmed to be 1720V.

Optimum pre-delay and gate

Table 2, Die-away calculated using the gate width pair of 20 and 40 μ s results in a gate of 23.0 μ s, therefore the optimum gate is 24 μ s.

Gate	Doubles	Die-away
(μS)	(cps)	(μS)
10	74.079	18.94
20	117.776	22.98
40	167.105	19.59
80	188.803	25.84
160	197.343	

In conclusion, the operating parameter measurements resulted in values in close agreement with the standard IPCA2 operating settings and therefore provide a valuable confirmation of IPCA2 robust and unchanged performance.

10. Summary

Table 3 provides and overview of all the control measurements performed over the reporting period (April 2024 – March 2025).

Table 3: The number of measurements taken monthly organized by type.

Month	Pu Eff	Cm	Load Cell
April 2024	NA	NA	NA
May 2024	NA	NA	NA
June 2024	NA	NA	NA
July 2024	NA	NA	NA
August 2024	NA	NA	NA
September 2024	NA	NA	NA
October 2024	NA	NA	NA
November 2024	1	1	NA
December 2024	2	2	NA
January 2025	3	3	2
February 2025	2	2	1
March 2025	2	2	2
Total	9	9	5

The results of the control measurements confirmed stability of plutonium efficiency, which stayed within 2σ , apart from the one discussed in section 2, of the overall average value of 7.33 % established from all the historical data (2013-2020). No dependence on environmental conditions (temperature, humidity) was observed. The average efficiency of the JFY24 measurements (performed between November 2024 and March 2025) corresponds to $7.28 \pm 0.04\%$.

Curium source measurements were performed over the reporting period and showed stable performance, those measurements outside of 3σ are due to the change in the neutron scattering profile because of the removal of the HPGe detectors. The average count rate (decay corrected with respect to the first measurement on 10/15/2018) for measurements performed between November 2024 and March 2025 corresponds to 983.9 ± 4.4 cps. No dependence on environmental conditions (temperature, humidity) was observed.

According to contractual agreement, the HPGe control measurements were discontinued starting in JFY21 as the HPGe gamma spectra showed deteriorated performance on all three gamma detectors [1].

UDL1 performance was evaluated as part of IPCA2 set-up and acquired data demonstrated statistical agreement between UDL1 data acquisition (both list mode as well as dataz) with JRS15/INCC/MIC system for Pu, ²⁴⁴Cm. In summary, the neutron system performance exhibits expected trends and no other measurements will be performed, as we will focus on preparing the system for shipment to Rokkasho in JFY 25.

11. References

- [1] D. Henzlova, J. Archuleta, A. Favalli, C.D. Rael, M.T. Swinhoe, "Annual IPCA2 Performance Report for JFY20", Los Alamos National Laboratory Technical Report, LA-UR-21-22978 (2021).
- [2] D. Henzlova, A. Sagadevan, "UDL1 Performance Test", Los Alamos National Laboratory Technical Report, TBD (2023).
- [3] M.T. Andrews, M.T. Swinhoe, J. Archuleta, D. Henzlova, A. Favalli, J.B. Marlow, "IPCA 2 Data Analysis and Updated Control Charts", Los Alamos National Laboratory Technical Report, LA-CP-20366 (2017).
- [4] D. Henzlova, J. Archuleta, M.T. Andrews, A. Favalli, J.B. Marlow, C.D. Rael, M.T. Swinhoe, "Annual IPCA2 Performance Report for JFY18", Los Alamos National Laboratory Technical Report, LA-UR-19-24828 (2019).
- [5] C.D. Rael, D. Henzlova, A. Sagadevan, D. Leyba, M.T. Swinhoe, "Annual IPCA2 Performance Report for JFY23", Los Alamos National Laboratory Technical Report, LA-UR-24-22781 (2024).